

Supplementary Table S4. *Statistical comparison between droplet digital PCR and real-time quantitative PCR in estimated vector copy number mean of cloned induced pluripotent stem cells*

<i>VCN comparison</i>	<i>ddPCR (1) vs. ddPCR (2)</i>	<i>qPCR (1) vs. qPCR (2)</i>	<i>ddPCR (1) vs. qPCR (1)</i>	<i>ddPCR (1) vs. qPCR (2)</i>	<i>ddPCR (1) vs. qPCR (2)</i>	<i>ddPCR (2) vs. qPCR (2)</i>
Low vs. low	NS	NS	NS	NS	NS	NS
Mid vs. mid	NS	NS	NS	NS	NS	NS
High vs. high	NS	NS	****a	****a	****a	****a

ddPCR, droplet digital PCR; NS, not significant; qPCR, real-time quantitative PCR; VCN, vector copy number.

****a $p < 0.0001$.