

Supplemental Table 1: Cut-off values according to the manufacturers' recommendations

	IgG (IU/mL)			IgM		
	Negative	Equivocal	Positive	Negative	Equivocal	Positive
Advia						
Centaur ^{®2}	<6.4	6.4-10	≥10	<0.9	0.9-1	≥1
Architect ^{®2}	<1.5	1.6-3	≥3	<0.5	0.5-0.6	≥0.6
AxSYM ^{®1}	<2	2-3	≥3	<0.5	0.55-0.6	≥0.6
Elecsys ^{®2}	<1	1-30	≥30	<0.8	0.8-1	≥1
Enzygnost [®]	<4	4-10	≥10	*	*	*
Liaison ^{®2}	<7.2	7.2-8.8	≥8.8	<6	6-8	≥8
Platelia ^{®2}	<6	6-9	≥9	<0.8	0.8-1	≥1
VIDAS ^{®1}	<4	4-8	≥8	<0.55	0.55-0.65	≥0.65
VIDIA ^{®1}	<2	3-5	≥5	<0.68	0.68-1	≥1

¹: Assessment 1

²: Assessment 2

* In each series, the cut-off value was calculated by adding 0.200 to the mean of the absorbance values of Toxoplasmosis/IgM negative control serum. After application of a correction factor (calculated from the Toxoplasmosis/IgM positive control serum) on the absorbance values, the resulting value "A" was interpreted as follows: negative if A<cut-off, positive if A>cut-off+0.050, equivocal if cut-off≤A≤cut-off+0.050