



Table of Onclarity by LA for HPV 16				
	Linear Array			
Onclarity	POS	NEG	.	Total
POS	600	23	1646	623
NEG	40	1885	6651	1925
.	13	36	200	.
Total	640	1908	.	2548
Frequency Missing = 8546				

Table of Onclarity by LA for HPV 18				
	Linear Array			
Onclarity	POS	NEG	.	Total
POS	173	14	507	187
NEG	23	2338	7790	2361
.	5	44	200	.
Total	196	2352	.	2548
Frequency Missing = 8546				

Table of Onclarity by LA for HPV 31				
	Linear Array			
Onclarity	POS	NEG	.	Total
POS	283	38	879	321
NEG	21	2206	7418	2227
.	7	42	200	.
Total	304	2244	.	2548
Frequency Missing = 8546				

Table of Onclarity by LA for HPV 45				
	Linear Array			
Onclarity	POS	NEG	.	Total
POS	116	7	373	123
NEG	28	2397	7924	2425
.	3	46	200	.
Total	144	2404	.	2548
Frequency Missing = 8546				

Table of Onclarity by LA for HPV 52				
	Linear Array			

Onclarity	POS	NEG	.	Total
POS	303	48	1029	351
NEG	38	2159	7268	2197
.	3	46	200	.
Total	341	2207	.	2548
Frequency Missing = 8546				

Table of Onclarity by LA for HPV 51				
	Linear Array			
Onclarity	POS	NEG	.	Total
POS	186	17	733	203
NEG	28	2317	7564	2345
.	4	45	200	.
Total	214	2334	.	2548
Frequency Missing = 8546				

Table of Onclarity by LA for HPV 3358				
	Linear Array			
Onclarity	POS	NEG	.	Total
POS	210	18	730	228
NEG	21	2299	7567	2320
.	1	48	200	.
Total	231	2317	.	2548
Frequency Missing = 8546				

Table of Onclarity by LA for HPV 396835				
	Linear Array			
Onclarity	POS	NEG	.	Total
POS	365	101	1560	466
NEG	63	2019	6737	2082
.	9	40	200	.
Total	428	2120	.	2548
Frequency Missing = 8546				

Table of Onclarity by LA for HPV 595666				
	Linear Array			
Onclarity	POS	NEG	.	Total
POS	375	52	1615	427
NEG	57	2064	6682	2121
.	8	41	200	.

Total	432	2116.	2548
Frequency Missing = 8546			

McNemar's Test	
Statistic (S)	4.5873
DF	1
Pr > S	0.0322
Simple Kappa Coefficient	
Kappa	0.9337
ASE	0.0082
95% Lower Conf Limit	0.9175
95% Upper Conf Limit	0.9498

McNemar's Test	
Statistic (S)	2.1892
DF	1
Pr > S	0.139
Simple Kappa Coefficient	
Kappa	0.8955
ASE	0.017
95% Lower Conf Limit	0.8623
95% Upper Conf Limit	0.9288

McNemar's Test	
Statistic (S)	4.8983
DF	1
Pr > S	0.0269
Simple Kappa Coefficient	
Kappa	0.8924
ASE	0.0138
95% Lower Conf Limit	0.8654
95% Upper Conf Limit	0.9195

McNemar's Test	
Statistic (S)	12.6
DF	1
Pr > S	0.0004
Simple Kappa Coefficient	
Kappa	0.8617
ASE	0.023
95% Lower Conf Limit	0.8166
95% Upper Conf Limit	0.9068

McNemar's Test	
Statistic (S)	1.1628

DF	1
Pr > S	0.2809
Simple Kappa Coefficient	
Kappa	0.8562
ASE	0.0152
95% Lower Conf Limit	0.8265
95% Upper Conf Limit	0.8859

McNemar's Test	
Statistic (S)	2.6889
DF	1
Pr > S	0.1011
Simple Kappa Coefficient	
Kappa	0.8825
ASE	0.0173
95% Lower Conf Limit	0.8486
95% Upper Conf Limit	0.9163

McNemar's Test	
Statistic (S)	0.2308
DF	1
Pr > S	0.631
Simple Kappa Coefficient	
Kappa	0.9066
ASE	0.0148
95% Lower Conf Limit	0.8776
95% Upper Conf Limit	0.9356

McNemar's Test	
Statistic (S)	8.8049
DF	1
Pr > S	0.003
Simple Kappa Coefficient	
Kappa	0.7776
ASE	0.0166
95% Lower Conf Limit	0.7451
95% Upper Conf Limit	0.8101

McNemar's Test	
Statistic (S)	0.2294
DF	1
Pr > S	0.632
Simple Kappa Coefficient	

Kappa	0.8474
ASE	0.0142
95% Lower Conf Limit	0.8195
95% Upper Conf Limit	0.8753