Submitting Institution: XXXX Patient Identifier: 0000 Report prepared on: 6/13/2016 12:00:00 AM

Dosimetric Analysis Summary:

Structure	Dose criteria	KBP feedback statement			
PTV 4500	All metrics meet protocol specifications	Dose metrics are unlikely to be improved further			
Bone Marrow	Some acceptable deviations	Some dose metrics could likely be improved			
Bowel	All metrics meet protocol specifications	Some dose metrics could likely be improved			
Rectum	Some acceptable deviations	Plan could be improved to meet protocol specifications			
Bladder	Some acceptable deviations	Some dose metrics could likely be improved			

Please review the subsequent pages of this report for individual structure DVH analysis and comparison to the knowledge-based planning reference for this patient.

The information contained in this report is designed to guide radiotherapy plan quality decision-making on the part of the submitting site. It is at the discretion of the submitting site whether or not to re-plan the case according to the patient-specific feedback in this report. Your site will need to review and determine if you will re-optimize the original plan using the KBP information provided or not re-optimize the original plan that was submitted and approved.

Submitting Institution: XXXX Patient Identifier: 0000 Report prepared on: 6/13/2016 12:00:00 AM

PTV 4500 Report



Submitted Plan from Institution XXXX for Patient 0000

Structure: ptv4500							
Docimetrie	Sh:44 a d	protocol compliance criteria			Knowledge_based plan quality check		
Parameter	Plan	Per	Variation	Deviation	KBP Plan	KBP Feedback	
1 al allieter	1 1411	Protocol	Acceptable	Unacceptable			
D95%(Gy)	45.00	>=45	>=43.65	<43.65	45.00	Plan meets per protocol	
						specifications and is	
						unlikely to be improved	
						further	
D97%(Gy)	44.67	>=43.65	>=40.5	<40.5	44.60	Plan meets per protocol	
						specifications and is	
						unlikely to be improved	
						further	
D99%(Gy)	43.84	>=40.5	>=39.6	<39.6	43.52	Plan meets per protocol	
						specifications and is	
						unlikely to be improved	
						further	
DMax(Gy)	49.46	<=51.75	<=54	>54	50.96	Plan meets per protocol	
						specifications and is	
						unlikely to be improved	
						further	

Submitting Institution: XXXX Patient Identifier: 0000 Report prepared on: 6/13/2016 12:00:00 AM

Bone Marrow Report



Submitted Plan from Institution XXXX for Patient 0000

Structure: PBM							
Destination	S-h	protocol compliance criteria			Knowledge based plan quality check		
Dosimetric	Blop	Per	Variation	Deviation	KBP Plan	KBP Feedback	
rarameter	Flan	Protocol	Acceptable	Unacceptable			
DMeanGy)	27.98	<=27	<=29	>29 (or	27.11	Plan exhibits acceptable	
				KBP		variation but this dose	
				prediction)		metric could likely be	
						improved	
V10(%)	84.45	<=85.5%	<=90%	>90% (or	80.70	Plan meets per protocol	
				KBP		specifications, but this	
				prediction)		dose metric could likely	
						be improved further	
V20(%)	71.44	<=66%	<=7 5 %	>75% (or	68.50	Plan exhibits acceptable	
				KBP		variation but this dose	
				prediction)		metric could likely be	
						improved	

Submitting Institution: XXXX Patient Identifier: 0000 Report prepared on: 6/13/2016 12:00:00 AM

Bowel Report



Submitted Plan from Institution XXXX for Patient 0000

Structure: Bowel							
Dosimetric	Submitted Plan	protocol compliance criteria			Knowledge_based plan quality check		
		Per	Variation	Deviation	KBP Plan	KBP Feedback	
rarameter		Protocol	Acceptable	Unacceptable			
D30%(Gy)	28.48	<=40	<=50	>50	28.59	Plan meets per protocol	
						specifications and is	
						unlikely to be improved	
						further	
DMax(Gy)	48.49	<=59.4	<=62.1	>62.1	47.36	Plan meets per protocol	
						specifications, but this	
						dose metric could likely	
						be improved further	
V45(cc)	48.52	<=200cc	<=250cc	>250cc(or	47.90	Plan meets per protocol	
				KBP		specifications, but this	
				prediction)		dose metric could likely	
						be improved further	

Submitting Institution: XXXX Patient Identifier: 0000 Report prepared on: 6/13/2016 12:00:00 AM

Rectum Report



Submitted Plan from Institution XXXX for Patient 0000

Structure: rectum							
Dosimetric	Submitted Plan	protocol compliance criteria			Knowledge based plan quality check		
		Per	Variation	Deviation	KBP Plan	KBP Feedback	
rarameter		Protocol	Acceptable	Unacceptable			
D50%(Gy)	36.97	<=45	<=54	>54	31.57	Plan meets per protocol	
						specifications, but this	
						dose metric could likely	
						be improved further	
D60%(Gy)	33.56	<=30	< =47.6	>47.6	25.78	Plan exhibits acceptable	
						variation but this dose	
						metric could likely be	
						improved to meet per	
						protocol specifications	
DMax(Gy)	48.43	<=50	<=55	>55	48.79	Plan meets per protocol	
						specifications and is	
						unlikely to be improved	
						further	

Submitting Institution: XXXX Patient Identifier: 0000 Report prepared on: 6/13/2016 12:00:00 AM

Bladder Report



Submitted Plan from Institution XXXX for Patient 0000

Structure: bladder							
Dosimetric	Submitted Plan	protocol compliance criteria			Knowledge_based plan quality check		
		Per	Variation	Deviation	KBP Plan	KBP Feedback	
1 al ameter		Protocol	Acceptable	Unacceptable			
D50%(Gy)	45.82	<=45	<=55	>55	45.63	Plan exhibits acceptable	
						variation but this dose	
						metric could likely be	
						improved	
DMax(Gy)	47.07	<=50	<=57.5	>57.5	47.21	Plan meets per protocol	
						specifications and is	
						unlikely to be improved	
						further	