

**Évaluation du Programme Pilote avec les TDR-e**  
**CHECKLIST POUR VISITE DE LABORATOIRE**

Date: \_\_\_\_\_

Nom du laboratoire : \_\_\_\_\_

ITEMS	OUI	NON
1. POS ou aides mémoire sont disponibles, visibles et à jour		
2. Thermomètre dans le site de stockage pour les TDR-e		
3. Les kits sont stockés à la bonne température (2-30 C)		
4. Suivi des stocks comprend le numéro de lot et la date d'expiration		
5. Thermomètre dans le site ou les TDR-e sont effectués		
6. Le site pour faire le test à la bonne température (15-40 C)		
7. Log de températures de labo est rempli chaque jour		
8. Chronomètre au site de test (cellulaire okay)		
9. Frigo (de 2-8 C) au laboratoire		
10. Contrôles au laboratoire		
11. Registre de contrôles faites au laboratoire		
12. L'écriture est lisible sur les registres du labo		
13. Les registres sont remplis au complet		
14. Poubelles pour les déchets dangereux et non dangereux dans le laboratoire		
15. EPIs (gants, blouse, protection pour le visage) sont disponibles		
16. Installation de lavage des mains est disponible		
17. Disponibilité d'un téléphone pour envoyer des données		

	# personnes	
	Titulaires	Stagiaires
Personnes qui ont rempli un questionnaire		
Total de personnes qui font les TDR-e		
Total de personnes qui travaillent dans ce labo		

Notes :

**Evaluation of Ebola RDT Pilot Program**  
**CHECKLIST FOR LABORATORY VISIT**

Date: \_\_\_\_\_

Laboratory name : \_\_\_\_\_

ITEMS	YES	NO
1. SOP or job aids are available, visible and up-to-date		
2. Thermometer in the storage site for the Ebola RDTs		
3. The kits are stored at the correct temperature (2-30 C)		
4. Inventory tracking includes the lot number and expiration date		
5. Thermometer in the site where Ebola RDTs are performed		
6. Site for performing the Ebola RDT is at the correct temperature (15-40 C)		
7. The laboratory temperature log is filled each day		
8. Timer at the testing site (cellphone okay)		
9. Refrigerator (at 2-8 C) in the laboratory		
10. Controls at the laboratory		
11. Register of control tests done at the laboratory		
12. Writing is legible on the laboratory registers		
13. The registers are completed in full		
14. Waste bins for hazardous and non-hazardous waste in the laboratory		
15. PPE (gloves, gown, face protection) are available		
16. Handwashing station is available		
17. A telephone for sending data is available		

	# persons	
	Functionaries	Trainees
Persons who completed a questionnaire		
Total persons who perform Ebola RDTs		
Total persons who work in this laboratory		

Notes :