

Table S4. Subject-specific and overall performance for a selection of eight features (a) to (h) and hypothesis (H4) relative SV via subject-specific calibration. The performance between ΔSV_{EIT} and ΔSV_{Ref} is evaluated in terms of angular error ϵ_α and angular concordance rate CR. The (†) indicates unrealistic solutions with calibrations coefficients *not* having identical sign for all subjects. Cell shadings indicate whether the acceptance criteria (see methods section) are met (green), not met (red), or met but with unrealistic calibration coefficients (yellow).

	(a) $\Delta\sigma_{\text{H}}$		(b) tStd _H		(c) $\Delta\sigma_{\text{L}}$		(d) tStd _L	
	ϵ_{Abs} (mL)	r (1)	ϵ_{Abs} (mL)	r (1)	ϵ_{Abs} (mL)	r (1)	ϵ_{Abs} (mL)	r (1)
S01	-6.2 ± 32.0	66.7	-24.8 ± 1.6	100.0	-8.7 ± 11.0	100.0	-2.6 ± 31.2	33.3
S02	-3.9 ± 24.9	81.8	-13.5 ± 15.7	100.0	-19.3 ± 8.2	100.0	-19.3 ± 8.2	100.0
S03	-14.5 ± 17.4	83.3	-22.3 ± 7.7	80.0	3.5 ± 18.7	85.7	3.7 ± 19.0	85.7
S04	1.7 ± 22.5	88.9	-0.2 ± 23.7	88.9	2.6 ± 17.5	100.0	-1.5 ± 20.8	100.0
S05	0.0 ± 23.4	85.7	1.9 ± 21.3	85.7	-21.0 ± 14.6	75.0	-9.6 ± 24.9	83.3
S06	4.0 ± 24.6	73.7	3.6 ± 26.2	66.7	10.3 ± 24.1	72.7	-2.8 ± 19.1	100.0
S08	7.0 ± 22.6	87.5	3.5 ± 8.0	100.0	6.8 ± 21.8	90.0	1.4 ± 21.4	100.0
S09	-4.5 ± 7.1	100.0	-3.0 ± 3.3	100.0	-21.4 ± 7.7	100.0	-23.9 ± 8.1	100.0
S10	-19.9 ± 16.1	66.7	-18.2 ± 16.2	66.7	-25.0 ± 10.6	66.7	-27.2 ± 4.0	66.7
All	-1.0 ± 23.0	80.9	(†) -3.9 ± 21.5	83.3	(†) -0.2 ± 22.5	84.6	(†) -5.8 ± 20.4	91.5

	(e) tStd _G		(f) $\Delta\sigma_{\text{H}}, \frac{\Delta\sigma_{\text{H}}}{\sigma_{\text{G}}}$		(g) $\Delta\sigma_{\text{L}}, \frac{\Delta\sigma_{\text{L}}}{\sigma_{\text{G}}}$		(h) V _T	
	ϵ_{Abs} (mL)	r (1)	ϵ_{Abs} (mL)	r (1)	ϵ_{Abs} (mL)	r (1)	ϵ_{Abs} (mL)	r (1)
S01	-22.6 ± 14.2	50.0	1.7 ± 8.9	100.0	1.5 ± 6.2	100.0	1.3 ± 4.8	100.0
S02	-3.4 ± 25.3	81.8	-1.0 ± 16.5	88.9	-6.3 ± 20.0	87.5	-2.5 ± 15.1	88.9
S03	-4.1 ± 7.2	100.0	-7.0 ± 18.7	83.3	-1.0 ± 12.8	100.0	-1.7 ± 25.7	85.7
S04	3.4 ± 24.5	81.8	0.4 ± 15.0	100.0	0.8 ± 15.0	100.0	-1.0 ± 16.0	100.0
S05	10.9 ± 29.2	60.0	-4.2 ± 10.3	100.0	-1.1 ± 6.0	100.0	-2.6 ± 6.4	100.0
S06	10.2 ± 23.3	75.0	5.4 ± 19.4	84.6	5.4 ± 20.8	83.3	1.8 ± 17.0	92.3
S08	-2.2 ± 23.1	60.0	10.7 ± 22.1	77.8	8.3 ± 18.9	90.0	4.6 ± 12.9	100.0
S09	-14.4 ± 9.5	100.0	-1.2 ± 2.4	100.0	-6.6 ± 15.5	100.0	-0.4 ± 9.0	100.0
S10	-26.2 ± 4.5	66.7	-14.7 ± 16.9	66.7	-1.7 ± 17.0	100.0	-8.4 ± 13.3	100.0
All	(†) 2.0 ± 24.2	74.4	1.0 ± 17.5	87.7	1.3 ± 16.7	93.0	-0.4 ± 15.3	94.7