

## Validation and comparison of two NGS assays for the detection of EGFR T790M resistance mutation in liquid biopsies of NSCLC patients

### SUPPLEMENTARY MATERIALS

**Supplementary Table 1: Somatic variants of the multiplex reference standard and expected allele frequencies**

Gene	Variant	Expected allelic frequency (%)		
		0.1% reference standard	1% reference standard	5% reference standard
<i>EGFR</i>	L858R	0.10	1.00	5.00
<i>EGFR</i>	$\Delta$ E746-A750	0.10	1.00	5.00
<i>EGFR</i>	T790M	0.10	1.00	5.00
<i>EGFR</i>	V769_D770insASV	0.10	1.00	5.00
<i>KRAS</i>	G12D	0.13	1.30	6.30
<i>NRAS</i>	Q61K	0.13	1.30	6.30
<i>NRAS</i>	A59T	0.13	1.30	6.30
<i>PIK3CA</i>	E545K	0.13	1.30	6.30

**Supplementary Table 2: Clinical non-small cell lung cancer (NSCLC) samples analyzed in this study.** See\_Supplementary\_Table 2

**Supplementary Table 3: Reference samples (synthetic plasma) analyzed in this study.** See\_Supplementary\_Table 3