

Supplementary Table 2. Data Set for Three-Dimensional Anteversion

No. of trials	Planning value (°)	With surgical navigation		Without surgical navigation	
		Measured value (°)	Error (°)	Measured value (°)	Error (°)
	A	B	B-A	C	C-A
1	4.07	2.80	1.27	2.92	1.15
2	4.07	3.71	0.36	3.81	0.26
3	4.07	3.50	0.57	10.99	6.92
4	4.07	2.25	1.82	4.61	0.54
5	4.07	3.19	0.88	1.44	2.63
6	4.07	1.36	2.71	1.70	2.37
7	4.07	1.67	2.40	0.72	3.35
8	4.07	1.15	2.92	8.26	4.19
9	4.07	4.22	0.15	13.83	9.76
10	4.07	3.04	1.03	2.33	1.74
11	4.07	2.84	1.23	5.42	1.35
12	4.07	3.00	1.07	0.29	3.78
13	4.07	5.15	1.08	7.17	3.10
14	4.07	4.85	0.78	7.44	3.37
15	4.07	4.15	0.08	7.24	3.17
16	4.07	4.09	0.02	4.80	0.73
17	4.07	3.59	0.48	4.83	0.76
18	4.07	3.59	0.48	15.62	11.55
19	4.07	5.79	1.72	7.88	3.81
20	4.07	5.57	1.50	4.04	0.03
21	4.07	1.89	2.18	5.62	1.55
22	4.07	2.39	1.68	9.20	5.13
23	4.07	2.23	1.84	3.59	0.48
24	4.07	5.20	1.13	13.90	9.83
25	4.07	3.95	0.12	0.54	3.53
26	4.07	5.45	1.38	7.62	3.55
27	4.07	3.68	0.39	1.89	2.18
28	4.07	1.33	2.74	5.48	1.41
29	4.07	4.09	0.02	4.81	0.74
30	4.07	3.46	0.61	10.24	6.17