



S1 Fig. Results of a pilot study to determine the NTAP time for decontamination experiments. Pilot experiment to determine the required NTAP treatment time for decontamination was carried out on PGLA sutures with *E. coli*. Number of surviving bacteria on sutures after NTAP treatment decreased with increasing treatment time. 3-minute NTAP treatment led to complete inactivation of bacteria.