

Supplemental Section

Figure S1. Analytical resolution of 26DM (quantitation ion peak), as compared between injections made on a DB-17MS (top chromatogram) and a DB-FFAP (bottom chromatogram) analytical column. In each chromatogram, 26DM analyte peak is labeled with an asterisk.




Table S1: Instrument operation parameters and settings.

Agilent 7890A, 7890B		Agilent 7000C	
Instrument operation parameters	Setting	Instrument operation parameters	Setting
Injection port	Pulsed splitless	MSD mode	Positive ion detection
Injection port temperature	250°C	Ionization mode	Electron impact
Injection volume	1 µL	Source temperature	280°C
Injection pulse pressure	30 psi until 0.35 min	Quad temperature	150°C
Septum purge flow	3 mL/min	Collision gas	N ₂ at 1.5 mL/min
Purge flow to split vent	60 mL/min at 0.35 min	Collision pressure	35 psi
Column 1	Constant flow 1.68 ± 0.05 mL/min (analytical run) Constant flow -2.20 ± 0.05 mL/min (post-run)	Quench gas	He at 2.25 mL/min
Column 2	Constant flow 1.85 ± 0.05 mL/min (analytical run) Constant flow 10.5 ± 0.05 mL/min (post-run)	Detector gain	20 (x 10 ⁵)
MSD transfer line temperature	280°C		
Oven program (analytical run)	80°C (hold 2 min) 30°C/min ramp to 180°C 15°C/min ramp to 240°C		
Oven program (post-run)	240°C (hold 5.5 min)		

Table S2: Compound-dependent MS parameters and ion transitions for AA analysis.

Analyte	Precursor ion	Product ion	Collision energy	Dwell time
	(m/z)	(m/z)	(V)	(ms)
OTOL	Quantitation ion	253	116	30
	Confirmation ion	253	106	25
	Internal standard, quantitation ion	259	122	30
	Internal standard, confirmation ion	259	112	25
26DM	Quantitation ion	267	148	10
	Confirmation ion	267	120	25
	Internal standard, quantitation ion	273	154	15
	Internal standard, confirmation ion	273	126	30
OANS	Quantitation ion	269	135	30
	Confirmation ion	269	190	15
	Internal standard, quantitation ion	276	139	35
	Internal standard, confirmation ion	276	157	15
1AMN	Quantitation ion	289	144	20
	Confirmation ion	289	127	30
	Internal standard, quantitation ion	295	150	20
	Internal standard, confirmation ion	296	151	20
2AMN	Quantitation ion	289	144	20
	Confirmation ion	289	127	30
	Internal standard, quantitation ion	296	151	20
	Internal standard, confirmation ion	295	150	20
4ABP	Quantitation ion	315.1	170	30
	Confirmation ion	315.1	141	40
	Internal standard, quantitation ion	324	179	20
	Internal standard, confirmation ion	324	150	45

Table S3: Sample processing and instrument parameters were varied from method conditions to test the ruggedness of assay. Three different column lots were tested to assess the ruggedness of the DB-FFAP column.

Analyte	Measured concentration (ng/L)															
	3 μ L PFPA, control	2 μ L PFPA (%-diff)	4 μ L PFPA (%-diff)	15 hr hydrolysis, control	14 hr hydrolysis (%-diff)	16 hr hydrolysis (%-diff)	250°C inlet temperature, control	240°C inlet temperature (%-diff)	260°C inlet temperature (%-diff)	30 psi pulsed pressure, control	25 psi pulsed pressure (%-diff)	35 psi pulsed pressure (%-diff)	Column lot #1	Column lot #2 (%-diff)	Column lot #1	Column lot #3 (%-diff)
OTOL	448.1	515.9 (15.12)	441.8 (-1.40)	448.1	460.4 (2.73)	513.1 (14.49)	214.7	219.0 (2.02)	203.4 (-5.27)	214.7	181.2 (-15.61)	215.1 (0.20)	345.9 600.0 842.4 431.4 750.1	341.9 (-1.15) 527.9 (12.78) 884.8 (4.92) 466.6 (7.84) 753.3 (0.42)	170.8 488.1 64.4 159.4 491.4	154.0 (9.83) 405.5 (0.49) 73.9 (14.65) 168.1 (5.62) 562.5 (14.46)
26DM	456.0	413.9 (-9.23)	419.6 (-7.98)	456.0	435.1 (-4.60)	411.6 (-9.74)	150.7	149.1 (-1.02)	154.7 (2.65)	150.7	156.2 (3.67)	152.3 (1.09)	139.2 281.2 192.7 66.2 221.6	146.6 (1.13) 319.4 (12.72) 212.7 (0.90) 66.7 (0.75) 216.8 (2.21)	108.6 445.1 17.7 112.8 454.6	113.4 (14.40) 405.4 (11.39) 19.5 (10.20) 120.3 (6.70) 474.4 (4.36)
OANS	403.4	391.9 (-2.86)	375.2 (-7.00)	403.4	398.6 (-1.20)	389.2 (3.52)	133.6	139.4 (4.33)	138.2 (3.49)	133.6	141.3 (5.81)	151.3 (13.27)	125.8 345.2 307.6 63.5 290.4	125.2 (0.45) 345.6 (0.13) 295.9 (3.89) 64.3 (1.36) 300.7 (3.49)	147.1 475.8 41.8 143.7 462.4	143.8 (2.24) 474.3 (0.35) 43.3 (3.50) 145.9 (1.51) 471.2 (1.90)
1AMN	314.9	295.4 (-6.21)	311.7 (-1.02)	314.9	319.3 (1.39)	287.7 (-6.64)	97.9	106.8 (9.15)	98.6 (0.71)	97.9	98.8 (0.96)	100.7 (2.87)	40.7 40.9 60.0 23.9 22.7	37.3 (4.81) 38.4 (-6.46) 58.5 (2.52) 21.8 (9.28) 21.9 (3.45)	102.5 433.6 13.0 99.9 419.9	97.1 (-5.27) 415.1 (4.26) 13.4 (3.01) 99.0 (-0.97) 378.3 (9.78)
2AMN	315.8	309.5 (-2.00)	308.2 (-2.43)	315.8	312.8 (-0.95)	311.3 (-1.44)	93.0	101.7 (9.37)	91.2 (-1.89)	93.0	92.0 (-1.08)	92.9 (-0.07)	10.4 317.7 497.5 16.8 396.1	103.1 (1.19) 309.5 (-2.52) 488.7 (-1.78) 16.5 (-1.54) 393.5 (-0.66)	108.3 442.9 15.0 102.6 436.5	102.6 (5.31) 434.7 (-1.97) 15.7 (4.60) 98.8 (-3.90) 424.4 (-2.69)
4ABP	337.0	335.9 (-0.32)	334.3 (-0.78)	337.0	344.6 (2.26)	335.7 (-0.38)	105.6	119.5 (13.20)	112.9 (6.91)	105.6	114.1 (8.10)	116.4 (12.17)	24.0 267.8 339.7 21.6 283.5	24.4 (-1.72) 271.8 (1.49) 332.4 (-2.17) 21.6 (0.24) 283.7 (0.08)	413.5 5.2 52.5 105.6 405.2	444.1 (1.42) 5.4 (3.96) 61.3 (16.75) 111.7 (5.75) 405.2 (0.14)

Table S4: Characterization statistics obtained for QC low and high pools, using 39 replicates from each pool.

Analyte	QC low pool, averaged concentration (ng/L)	CV	QC high pool, averaged concentration (ng/L)	CV
OTOL	150.7	8.76	485.9	6.44
26DM	121.29	7.80	502.6	4.22
OANS	153.4	2.74	505.8	4.33
1AMN	110.8	5.56	458.1	5.11
2AMN	111.7	5.27	479.8	5.01
4ABP	109.0	4.04	446.1	5.24