

Supplementary Data

Quantification of Fecal Short Chain Fatty Acids by Liquid Chromatography Tandem Mass Spectrometry— Investigation of Pre-Analytic Stability

Gerhard Liebisch ^{1,*}, Josef Ecker ^{2,3}, Sebastian Roth ¹, Sabine Schweizer ¹, Veronika Öttl ¹, Hans-Frieder Schött ¹, Hongsup Yoon ^{3,4}, Dirk Haller ^{3,4}, Ernst Holler ⁵, Ralph Burkhardt ¹ and Silke Matysik ¹

¹ Institute for Clinical Chemistry and Laboratory Medicine, University Hospital Regensburg, Franz-Josef-Strauss-Allee 11, 93053 Regensburg, Germany; sebastian.roth@ukr.de (S.R.); Schweizer.sabine@gmx.net (S.S.); Veronika.Oettl@stud.uni-regensburg.de (V.O.); hans-frieder.schoett@isas.de (H.-F.S.); Ralph.Burkhardt@ukr.de (R.B.), silke.matysik@ukr.de (S.M.)

² Nutritional Physiology, Technical University of Munich, 85354 Freising, Germany; josef.ecker@tum.de

³ ZIEL-Institute for Food & Health, Technical University of Munich, 85354 Freising, Germany; Hongsup.Yoon@tum.de (H.Y.); dirk.haller@tum.de (D.H.);

⁴ Chair of Nutrition and Immunology, Technical University of Munich, 85354 Freising, Germany

⁵ Department of Hematology and Oncology, Internal Medicine III, University Hospital Regensburg, 93053 Regensburg, Germany; Ernst.Holler@ukr.de

* Correspondence: gerhard.liebisch@ukr.de; Tel.: +49-941-944-6240

Suppl. Figure 1A-D: Product ion spectra of SCFA-3NPH derivatives

Product ion spectra were generated in the enhanced product ion mode at a collision energy of -20 V.

FA 2:0







Figure S1A

FA 3:0**Figure S1B**

FA 4:0**Figure S1C**

FA 3:0(2Me)**Figure S1D**

Suppl. Figure 2A-D: Calibration lines:

Typical calibration line with 6 matrix based levels. Analyte concentrations are displayed in $\mu\text{mol}/2\text{g DW}$.

FA 2:0

	Sample Name	Sample Type	Analyte Peak Area (counts)	Analyte Concentration	IS Peak Area (counts)	Calculated Concentration	Accuracy (%)
1	CAL 1_Eigen_07-18	Standard	1.21e+004	0.461	1.32e+006	0.387	84.0
2	CAL 2_Eigen_07-18	Standard	1.27e+005	27.3	1.31e+006	31.2	114.
3	CAL 3_Eigen_07-18	Standard	2.51e+005	60.1	1.37e+006	61.4	102.
4	CAL 4_Eigen_07-18	Standard	1.12e+006	296.	1.29e+006	301.	102.
5	CAL 5_Eigen_07-18	Standard	3.48e+006	1080.	1.17e+006	1040.	97.0
6	CAL 6_Eigen_07-18	Standard	6.00e+006	2040.	1.02e+006	2060.	101.




Figure S2A

FA 3:0

	Sample Name	Sample Type	Analyte Peak Area (counts)	Analyte Concentration	IS Peak Area (counts)	Calculated Concentration	Accuracy (%)
1	CAL 1_Eigen_07-18	Standard	4.96e+003	0.430	2.48e+006	0.412	95.7
2	CAL 2_Eigen_07-18	Standard	1.69e+005	9.43	2.41e+006	10.2	108.
3	CAL 3_Eigen_07-18	Standard	3.43e+005	19.9	2.61e+006	18.9	95.1
4	CAL 4_Eigen_07-18	Standard	1.55e+006	93.4	2.32e+006	95.9	103.
5	CAL 5_Eigen_07-18	Standard	6.66e+006	466.	2.08e+006	457.	98.1
6	CAL 6_Eigen_07-18	Standard	1.10e+007	893.	1.75e+006	899.	101.




Figure S2B

FA 4:0

	Sample Name	Sample Type	Analyte Peak Area (counts)	Analyte Concentration	IS Peak Area (counts)	Calculated Concentration	Accuracy (%)
1	CAL 1_Eigen_07-18	Standard	0.00e+000	0.0819	2.59e+006	No Peak	0.00
2	CAL 2_Eigen_07-18	Standard	9.25e+004	8.53	2.54e+006	8.56	100.
3	CAL 3_Eigen_07-18	Standard	1.92e+005	17.8	2.63e+006	17.0	95.6
4	CAL 4_Eigen_07-18	Standard	9.02e+005	84.8	2.53e+006	82.7	97.5
5	CAL 5_Eigen_07-18	Standard	4.64e+006	465.	2.35e+006	459.	98.8
6	CAL 6_Eigen_07-18	Standard	7.89e+006	879.	2.06e+006	887.	101.





Figure S2C

FA 3:0(2Me)

	Sample Name	Sample Type	Analyte Peak Area (counts)	Analyte Concentration	IS Peak Area (counts)	Calculated Concentration	Accuracy (%)
1	CAL 1_Eigen_07-18	Standard	0.00e+000	0.0761	2.59e+006	No Peak	0.00
2	CAL 2_Eigen_07-18	Standard	7.89e+004	5.99	2.54e+006	6.04	101.
3	CAL 3_Eigen_07-18	Standard	1.65e+005	12.9	2.63e+006	12.1	93.5
4	CAL 4_Eigen_07-18	Standard	7.56e+005	60.7	2.53e+006	57.2	94.2
5	CAL 5_Eigen_07-18	Standard	3.00e+006	256.	2.35e+006	245.	95.7
6	CAL 6_Eigen_07-18	Standard	5.24e+006	471.	2.06e+006	486.	103.

**Figure S2D**

Suppl. Table 1: LoD determination

	FA 2:0	FA 3:0
Mean (blank)	1.06	0.10
SD (blank)	0.23	0.03
LoB	1.44	0.15
SD (low conc.)	0.27	0.028
LoD	1.88	0.20

$$\text{LoB} = \text{mean}(\text{blank}) + 1.645(\text{SD blank})$$

$$\text{LoD} = \text{LoB} + 1.645(\text{SD low concentration sample})$$

Displayed are mean and standard deviation calculated from internal standard blanks of nine different batches. Variations at low concentrations were calculated from the analysis of five sample aliquots of an 8-fold dilution of low QC (see Table 2). Limit of blank (LoB) and limit of detection (LoD) are calculated using the displayed equations. Values are [$\mu\text{mol/g DW}$]