

SUPPLEMENTARY TABLE S1. PRECISION STUDY: EFFECT OF SYSTEM RATE OF CHANGE ON CGM SYSTEM AND REFERENCE AGREEMENT

<i>CGM system rate of change (mg/[dL·min])</i>	<i>No. of paired CGM system-YSI reference readings</i>	<i>Percent of CGM system readings within</i>			
		<i>15/15% of reference</i>	<i>20/20% of reference</i>	<i>30/30% of reference</i>	<i>40/40% of reference</i>
Less than -2	475	79	91	97	99
(-2, -1)	1137	84	92	98	99
(-1, 1)	10,928	87	94	98	100
(1, 2)	1013	82	91	98	99
More than 2	539	76	87	97	99