



Supporting Information

for *Small*, DOI: 10.1002/sml.201704363

Microfluidic Contact Lenses

*Nan Jiang, Yunuen Montelongo, Haider Butt, and Ali K. Yetisen**

SUPPORTING INFORMATION

Microfluidic Contact Lenses

Nan Jiang,¹ Yunuen Montelongo,² Haider Butt,³ Ali K. Yetisen^{4,5,6,}*

¹ School of Engineering and Applied Sciences, Harvard University, Cambridge,
Massachusetts 02138, USA

² Department of Chemistry, Imperial College London, South Kensington Campus,
London SW7 2AZ, UK

³ Nanotechnology Laboratory, School of Engineering, University of Birmingham,
Birmingham B15 2TT, UK

⁴ School of Chemical Engineering, the University of Birmingham, Edgbaston,
Birmingham B15 2TT, UK

⁵ Institute of Translational Medicine, Mindelsohn Way, Edgbaston, Birmingham, B15
2TH, UK

⁶ Harvard-MIT Division of Health Sciences and Technology, Harvard University and
Massachusetts Institute of Technology, Cambridge, Massachusetts 02139, USA

KEYWORDS: Microfluidics; contact lenses; laser ablation; tear film; diagnostics

e-mail: akyetisen@gmail.com




Figure S1. Evaluation of transparency of the fabricated microfluidic contact lens. (a) A photograph of the polymerized PEGDA hydrogel. Scale bar=1 cm. (b) Schematic of microfluidic contact lens cross-section. Red arrows show light propagation through the microfluidic contact lens. (c) Top view of microfluidic contact lens. Scale bar= 3 mm.