

D:\GCMS\GCMS METHOD\Organic\Extract.qgm

Method

[Comment]

===== Analytical Line 1 =====

[AOC-20i+s]

of Rinses with Presolvent :5
 # of Rinses with Solvent(post) :8
 # of Rinses with Sample :2
 Plunger Speed(Suction) :High
 Viscosity Comp. Time :0.2 sec
 Plunger Speed(Injection) :High
 Syringe Insertion Speed :High
 Injection Mode :Normal
 Pumping Times :5
 Inj. Port Dwell Time :0.0 sec
 Terminal Air Gap :No
 Plunger Washing Speed :High
 Washing Volume :6uL
 Syringe Suction Position :0.0 mm
 Syringe Injection Position :0.0 mm
 Solvent Selection :All A,B,C

[GC-2010]

Column Oven Temp. :60.0 °C
 Injection Temp. :260.00 °C
 Injection Mode :Split
 Flow Control Mode :Linear Velocity
 Pressure :73.2 kPa
 Total Flow :16.3 mL/min
 Column Flow :1.21 mL/min
 Linear Velocity :40.1 cm/sec
 Purge Flow :3.0 mL/min
 Split Ratio :10.0
 High Pressure Injection :OFF
 Carrier Gas Saver :OFF
 Splitter Hold :OFF
 Oven Temp. Program

Rate	Temperature(°C)	Hold Time(min)
-	60.0	2.00
7.00	250.0	0.00
15.00	280.0	29.00

< Ready Check Heat Unit >

Column Oven : Yes
 SPL1 : Yes
 MS : Yes

< Ready Check Detector(FTD) >

< Ready Check Baseline Drift >

< Ready Check Injection Flow >

SPL1 Carrier : Yes
 SPL1 Purge : Yes

< Ready Check APC Flow >

< Ready Check Detector APC Flow >

External Wait :No
 Equilibrium Time :1.0 min

[GC Program]

[GCMS-QP2010 Ultra]

IonSourceTemp :230.00 °C
 Interface Temp. :270.00 °C
 Solvent Cut Time :4.50 min
 Detector Gain Mode :Relative
 Detector Gain :+0.00 kV
 Threshold :1000

[MS Table]

--Group 1 - Event 1--
 Start Time :5.00min
 End Time :60.13min
 ACQ Mode :Scan
 Event Time :0.20sec
 Scan Speed :3333
 Start m/z :40.00
 End m/z :650.00

Sample Inlet Unit :GC

[MS Program]

Use MS Program :OFF