

## Supplemental Data

**Supplemental Table 1: Solid Tumor Focus Assay Gene list.** Targeted regions from 69 genes are assayed. The OncoPrint™ Focus Assay was the base assay and it was expanded in-house with 17 alterations in the genes highlighted in red. This assay is designed to detect known clinically important single nucleotide variants, small insertions and deletions, high copy number amplifications, and gene rearrangements. Not all exons of all genes are assessed nor are all exons evaluated in their entirety


Hotspots Genes				Amplification (CNV) Genes		Gene Fusions	
AKT1	<i>EP300</i>	IDH2	PDGFRA	ALK	KIT	ABL1	MET
ALK	ERBB2	JAK1	PIK3CA	AR	KRAS	ALK	NTRK1
<i>APC</i>	ERBB3	JAK2	<i>POLE</i>	<i>AXL</i>	MET	AKT3	NTRK2
AR	ERBB4	JAK3	<i>PTEN</i>	BRAF	MYC	AXL	NTRK3
<i>ARAF</i>	ESR1	<i>KEAP1</i>	<i>RAC1</i>	CCND1	MYCN	BRAF	PDGFRA
<i>B2M</i>	<i>FBXW7</i>	KIT	RAF1	CDK4	PDGFRA	EGFR	PPARG
BRAF	FGFR1	KRAS	RET	CDK6	PIK3CA	ERBB2	RAF1
CDK4	FGFR2	MAP2K1	ROS1	EGFR	<i>RICTOR</i>	ERG	RET
<i>CDKN2A</i>	FGFR3	MAP2K2	<i>SMAD4</i>	ERBB2		ETV1	ROS1
<i>CREBBP</i>	GNA11	MET	SMO	FGFR1		ETV4	
CTNNB1	GNAQ	MTOR	<i>TP53</i>	FGFR2		ETV	
DDR2	HRAS	<i>NFE2L2</i>	<i>U2AF1</i>	FGFR3		FGFR1	
EGFR	IDH1	NRAS		FGFR4		FGFR2	
						FGFR3	

DNA

RNA

Gene alterations in **RED** were added in expanded Solid Tumor Focus Assay

**Supplemental Table 2: Comparison of genetic alterations within patient population at UHCMC using different assays.** Alterations listed by rate within UHCMC cohort for **A)** In-house Focus Assay, **B)** FoundationOne®, and **C)** for in-house EGFR *Therascreen*® assay. Data for the in-house Focus Panel, which includes both the Oncomine™ Focus Assay and the expanded Solid Tumor Focus Assay, were collected from 2016-2018. FoundationOne® and EGFR *Therascreen*® data were collected from 2014-2016.



## Supplemental Data: Appendix 1

### Part 1- OncoPrint Focus Assay Validation Summary

#### Section I: Fusions (RNA) validation

Test Category	Material Used
Analytical Sensitivity (LOD)	H2228 cell line ( <i>EML4-ALK</i> and <i>ALK-PTPN3</i> fusions) RNA serially diluted (100% to 3%) into RNA from non-fusion cell line (H1975)
Accuracy and Specificity	22 FFPE tumor samples analyzed against orthogonal test method(s)
	Reference material (Seraseq FFPE Fusion RNA-I) containing 10 gene fusions
Precision and Reproducibility	6 H2228 cell line dilutions analyzed within and between runs
	5 FFPE tumor samples analyzed within and between runs
	Seraseq FFPE Fusion RNA-I (10 gene fusions) analyzed within and between runs

**Limit of Detection (LoD):** RNA from H2228 cell line (known to contain *EML4-ALK* and *ALK-PTPN3* fusions) was serially diluted in the RNA from fusion-negative cell line (NCI-H1975) resulting in dilution ranging from 3% to 100% H2228 tumor RNA. Both *EML4-ALK* and *ALK-PTPN3* fusions were detected in all dilutions, see Table I-1. For detection of fusions, these experiments confirm a limit of detection of 3% tumor content (nuclei) for fusion detection.

**Table I-1.** Detection of two different *ALK* fusions in different HCC827 dilution levels.

% HCC228 Tumor RNA	Run	Fusion reads (normalized fusion reads)	
		<i>EML4(6)-ALK(20)</i>	<i>ALK(11)-PTPN3(3)</i>
100	A	6253 (.046)	4884 (.036)
50	A	5984 (.043)	3438 (.025)
50	B	8083 (.036)	5781 (.026)
50	B	9680 (.041)	5973 (.026)
25	A	9374 (.037)	2758 (.011)
25	B	8960 (.034)	4035 (.015)
25	B	9666 (.038)	3223 (.013)
12.5	A	3502 (.025)	1336 (.01)
12.5	B	5965 (.028)	1249 (.006)
12.5	B	3575 (.031)	955 (.008)
12.5	C	3884 (.024)	2526 (.016)
12.5	C	3120 (.022)	2094 (.015)
6	A	4074 (.018)	1460 (.006)
6	C	2580 (.015)	1616 (.009)
6	C	2966 (.017)	2520 (.014)
3	C	2230 (.011)	1651 (.008)
3	C	2581 (.009)	1510 (.005)
0	A	ND	ND

Normalized fusion reads (fusion reads/total mapped RNA reads) is depicted in parentheses for each fusion. **ND:** non detected

**Limits of Quantification and Reportable Range:** Sensitivity and replication studies using H2228 cell line demonstrate a combined analytic sensitivity of a 100% [88%-100%] in detection of both *ALK* fusions (samples containing 3% to 100% H2228 RNA). Studies on FFPE clinical tumor samples (tumor nuclei ranging for 20% to 70%) demonstrate analytic sensitivity of 100% [83%-100%] in detection of fusions (5 different common fusions tested). The clinically reportable range for fusions will be alterations that meet 'call' thresholds as determined by studies with H2228 cell line (Table I-2), which have been demonstrated to empirically hold in FFPE clinical tumor samples containing 20% to 70% tumor nuclei (expected test samples for assay), see Table I-3.

**Table I-2. Sequencing Metrics: Minimum Criteria for Calling Fusion Positive**

Metric	Threshold	Minimum Observed	Maximum Observed
Total Mapped RNA Reads	10,000	113,876	285,797
Fusion Reads	500	955	9680
Normalized Fusion Reads (fusion reads/total mapped RNA reads)	.005	.005	.046

- The values are derived from sensitivity and reproducibility experiments with H2228 cell line.

**Accuracy:** We evaluated the accuracy of the assay in its ability to detect fusions in FFPE clinical samples and characterized reference material. Results from clinical FFPE samples were compared with expected results seen by alternative NGS methodology (Foundation One; ArcherDx FusionPlex) and FDA-approved ALK IHC and/or ALK FISH testing. The 9 confirmed detected (positive) cases are shown in Table I-3. There were 13 confirmed negative cases, zero false positive, and zero false negative cases. Given the infrequent occurrence of many of the fusion in our clinical sample set, the Seraseq FFPE Fusion RNA reference material was used to increase the number of different variants tested on the panel. All 10 key fusions were detected and pass sequencing metrics detailed above, see Table I-4.

**Positive percent agreement (PPA;accuracy):** Clinical + reference FFPE samples: 100% [79%-100%]

**Positive predictive value (PPV):** Clinical + reference FFPE samples: 100% [79%-100%]

**Specificity:** Clinical + reference FFPE samples: 100% [72%-100%]

**Table I-3. Gene fusions detected in FFPE clinical samples**

Sample	Tumor %	Gene Fusion Detected	Confirmation Method	Fusion Reads	Mapped RNA Reads	Normalized Fusion Reads
OFA_2	40	<i>MET(13)-(15)</i> (exon 14 skip alteration)	NGS(FO)	22103	224723	.098
OFA_7	30	<i>EML4(6)-ALK(20)</i> (variant 3a/b)	NGS(FO; ADx); IHC	583	112647	.005
OFA_11	20	<i>EML4(6)-ALK(20)</i> (variant 3a/b)	NGS(FO; ADx); IHC	7487	151875	.049
OFA_12	20	<i>EML4(13)-ALK(20)</i> (variant 1)	NGS(FO; ADx); IHC	10389	138278	.075
OFA_13	70	<i>EML4(13)-ALK(20)</i> (variant 1)	NGS(FO; ADx); IHC	15602	181964	.086
OFA_14	40	<i>EML4(6)-ALK(20)</i> (variant 3a/b)	NGS(FO; ADx); IHC	8806	188895	.047
OFA_17	30	<i>MET(13)-(15)</i> (exon 14 skip alteration)	NGS(FO)	21444	181627	.118
OFA_18	40	<i>EML4(18)-ALK(20)</i> fusion (variant 5)	NGS(FO; ADx); IHC	45606	267877	.170
OFA_19	50	<i>KIF5B(15)-RET(12)</i>	NGS(FO; ADx)	47531	317170	.150

FO: Foundation Medicine One Testing; ADx: ArcherDx Fusion Plex Testing; IHC: FDA-approved Ventana ALK IHC

**Table I-4. Gene fusions detected in Seraseq FFPE Fusion RNA reference material.**

RNA Fusion	Fusion Reads	Normalized Fusion Reads	COSMIC/NCBI ID	OFA Variant ID
<i>EML4(13) - ALK(20)</i>	10819 ± 2096	.044 ± .001	AB462411	EML4-ALK.E13A20.AB462411
<i>KIF5B(24) - RET(11)</i>	11729 ± 2307	.048 ± .001	COSF1262	KIF5B-RET.K24R11.COSF1262
<i>NCOA4(7) - RET(12)</i>	10311 ± 2592	.042 ± .004		NCOA4-RET.N7R12
<i>CD74(6) - ROS1(34)</i>	4843 ± 1017	.020 ± .002	COSF1200	CD74-ROS1.C6R34.COSF1200
<i>SLC34A2(4) - ROS1(34)</i>	5701 ± 1154	.023 ± .002	COSF1198	SLC34A2-ROS1.S4R34.COSF1198
<i>TPM3(7) - NTRK1(10)</i>	12106 ± 2955	.049 ± .004	COSF1318	TPM3-NTRK1.T7N10.COSF1318
<i>ETV6(5) - NTRK3(15)</i>	16151 ± 3271	.066 ± .001	COSF571	ETV6-NTRK3.E5N15.COSF571.1
<i>FGFR3(17) - TACC3(11)</i>	12592 ± 1832	.052 ± .006		FGFR3-TACC3.F17T11
<i>FGFR3(17) - BAIAP2L1(2)</i>	7657 ± 1671	.031 ± .001	COSF1347	FGFR3-BAIAP2L1.F17B2.COSF1347
<i>PAX8(9) - PPARG(2)</i>	8223 ± 1725	.034 ± .003	COSF1217	PAX8-PPARG.P9P2.COSF1217

### **Precision and Reproducibility:**

**Intra-run experiments:** (1) Six different H2228 cell line dilutions ranging from 3% to 50% tumor (*EML4-ALK* and *ALK-PTPN3* fusions) were run in duplicate on the same run. The two expected fusions were detected in all 6 dilutions. (2) Five different clinical FFPE tumor samples (with tumor content ranging from 20% to 60%) were each run in triplicate. The


expected fusion was correctly called (detected above threshold) in all samples tested. (3) Seraseq FFPE Fusion RNA-I reference material were processed in duplicate in the same batch and run. The 10 expected fusions were correctly detected above threshold in all samples tested (mean coefficient of variation (CV) of .016% for normalized fusion reads) with 100% specificity. (4) A negative control (normal lung RNA) was run in triplicate and no fusions were detected.

Inter-run experiments: (1) Three different H2228 cell line dilutions ranging from 6% to 25% tumor were run on two different runs (6% and 25% dilution) or three different runs (12% dilution), on different days, with different reagents. The two expected fusions were detected in all 3 dilutions. (2) Five different clinical FFPE tumor samples (with tumor content ranging from 20% to 60%) were each run on two different runs, on different days, with different reagents. The expected fusion was correctly called (detected above threshold) in all samples tested. (3) Seraseq FFPE Fusion RNA-I reference material was run on two different batches on different days. The 10 expected fusions were correctly detected above threshold in all samples tested (mean CV of 0.02% for normalized fusion reads) with 100% specificity.

Overall Intermediate Precision: 100% [90% - 100%]

Summary of precision experiments for H2228 cell line (Figure I-1A) and FFPE clinical tumor samples (Figure I-1B) are shown below.

**Figure I-1A/B. Precision studies:** Chart showing normalized fusion reads for in precision studies involving (A) H2228 dilution samples and (B) FFPE clinical NSCLC tumor samples. Each symbol represents finding in an individual sample. All 'called' samples reached 500 fusion reads threshold.



\*Dilutions were run over three different runs spanning 3 months.

\*Each sample was run in quadruplicate over two different runs (3 months apart).

**Results of Sequencing Metrics in FFPE Clinical Samples (RNA):**

**Table I-5.** Sequencing metrics observed based on testing in clinical FFPE samples. Thresholds established based on experiments with ALK fusion (H2228) cell line, see Figure III-2.

Metric	Threshold	Minimum Observed	Maximum Observed	Mean	Median
Total Mapped Fusion Reads	10,000	29,194	436,285	205,557	212,631
Fusion Reads	500	583	47531	24934	23055
Normalized Fusion Reads	0.005	0.005	0.216	0.107	0.102

## II. Single Nucleotide Variants (SNV) and InDel(s) (DNA) Validation

Test Category	Material Used
Analytical Sensitivity (LOD)	Characterized reference material (quantified by ddPCR) containing 33 variants ranging from 3-5% allele frequency were tested
Accuracy and Specificity	44 FFPE tumor samples analyzed against orthogonal test method(s) Characterized reference material (Horizon diagnostics, Oncology hotspot control, CAP proficiency samples) testing 200 SNVs (VAF: 3-35%) and 10 indels (VAF: 2-35%)
Precision and Reproducibility	Characterized reference material (QMRS, AOHC) were analyzed within and between runs


**Limit of Detection (LoD):** Reference material from Horizon Diagnostics in which the variants were quantified by qPCR were analyzed focusing on 33 variants with VAF ranging from 3-6%. All 33 variants were detected and observed in the expected 95% CI for each variant, see Table II-1. These results demonstrate an empirical limit of detection of VAF of 3%. To be conservative, a VAF of 5% will be the stated limit of detection.

**Table II-1.** Detection of low frequency clinically-relevant variants in quantified reference material.

Gene	Variant	Expected VAF	Observed VAF	95% CI - Observed VAF at 500X coverage	Reference Standard
ALK	F1174L	5.0%	5.5%	3.4% - 7.3%	Horizon Tru-Q2
BRAF	V600G	4.0%	4.8%	2.6% - 6.1%	Horizon Tru-Q2
BRAF	V600K	4.0%	3.1%	2.6% - 6.1%	Horizon Tru-Q1
BRAF	V600M	4.0%	3.8%	2.6% - 6.1%	Horizon Tru-Q3
BRAF	V600R	4.0%	4.1%	2.6% - 6.1%	Horizon Tru-Q4
EGFR	L858R	4.2%	3.0%	2.7% - 6.2%	Horizon Tru-Q2
EGFR	L858R	3.0%	4.0%	1.8% - 4.9%	Horizon QMRS
EGFR	L861Q	4.2%	4.3%	2.7% - 6.2%	Horizon Tru-Q4
EGFR	T790M	4.2%	3.6%	2.7% - 6.2%	Horizon Tru-Q1
GNA11	Q209L	5.0%	5.1%	3.4% - 7.3%	Horizon Tru-Q3
GNAQ	Q209L	5.0%	5.9%	3.4% - 7.3%	Horizon Tru-Q2
IDH1	R132C	5.0%	5.0%	3.4% - 7.3%	Horizon Tru-Q1
IDH1	R132H	5.0%	6.1%	3.4% - 7.3%	Horizon Tru-Q3
IDH2	R140Q	5.0%	4.5%	3.4% - 7.3%	Horizon Tru-Q2
IDH2	R172K	5.0%	4.1%	3.4% - 7.3%	Horizon Tru-Q4
JAK2	V617F	5.0%	5.0%	3.4% - 7.3%	Horizon Tru-Q1
KIT	D816V	5.0%	4.8%	3.4% - 7.3%	Horizon Tru-Q4
KRAS	G12A	5.0%	4.3%	3.4% - 7.3%	Horizon Tru-Q1
KRAS	G12C	5.0%	4.5%	3.4% - 7.3%	Horizon Tru-Q4
KRAS	G12D	5.0%	4.6%	3.4% - 7.3%	Horizon Tru-Q4
KRAS	G12R	5.0%	5.8%	3.4% - 7.3%	Horizon Tru-Q1
KRAS	G12S	5.0%	5.1%	3.4% - 7.3%	Horizon Tru-Q3
KRAS	G12V	3.8%	4.4%	2.4% - 5.7%	Horizon Tru-Q2
KRAS	Q61H	5.0%	4.9%	3.4% - 7.3%	Horizon Tru-Q4
KRAS	Q61L	5.0%	5.7%	3.4% - 7.3%	Horizon Tru-Q2
KRAS	A146T	5.0%	3.9%	3.4% - 7.3%	Horizon Tru-Q3
NRAS	Q61H	5.0%	5.6%	3.4% - 7.3%	Horizon Tru-Q3
NRAS	Q61K	5.0%	5.1%	3.4% - 7.3%	Horizon Tru-Q1
NRAS	Q61L	5.0%	7.1%	3.4% - 7.3%	Horizon Tru-Q2
NRAS	Q61R	5.0%	5.8%	3.4% - 7.3%	Horizon Tru-Q4
PDGFRA	D842V	5.0%	5.0%	3.4% - 7.3%	Horizon Tru-Q4
PIK3CA	E542K	5.0%	4.2%	3.4% - 7.3%	Horizon Tru-Q3
PIK3CA	E545K	5.0%	5.7%	3.4% - 7.3%	Horizon Tru-Q2

**Limits of Quantification and Reportable Range:** A commercial FFPE reference control (variants quantified by ddPCR) were analyzed four times in three different runs. We measured the range of observed VAF at known common pathogenic cancer hotspots within our targeted panel and compare them to their respective expected values (ranging from 1.0% to 24.5%). A linear reportable range is observed, see Figure II-2.

**Figure II-2.** Boxplot showing the range of observed VAF in FFPE Horizon QMRS control (n=4).



**Accuracy:** (1) One of the objectives was to evaluate the accuracy of the assay in its ability to detect relevant mutations in well-characterized reference standard materials (Horizon diagnostics quantified standards: QMRS and TRU-Q1-4 5%; AcroMetrix Oncology Hotspot Control™, and pool of CAP proficiency samples). All reference material have been reported to be validated by at least one alternative method. All 200 SNVs (VAF ranging from 3-35%) and 10 indels (VAF 2-35%) covered by this panel were detected. Given that the OFA assay focuses on clinically relevant mutations, we focused our accuracy studies on well-characterized reference material that better represented these targeted regions and emphasize clinically relevant mutations. (2) The second objective was to evaluate of the accuracy of the assay in its ability to detect mutations in FFPE clinical samples which were confirmed by an alternate method. 44 FFPE samples, including 39 advanced stage NSCLC samples were evaluated by the OFA assay. The samples had orthogonal testing by FDA-approved CDx for EGFR (Therascreen,  $n=30$ ) or validated CGP NGS (Foundation Medicine, FM;  $n=29$ ). For complete list of samples, validation method, expected results and called results please refer to Table 2-'Validation Sample Master List' in the appendix. All 22 SNVs and 10 insertion/deletions (<30bp) were detected.

	Reference Material	FFPE clinical samples	Reference + clinical samples
PPA (accuracy)	100% [98%-100%]	100% [82%-100%]	100% [98%-100%]
PPV	100% [66%-100%]	100% [66%-100%]	100% [80%-100%]


\*[brackets] represent 95% confidence interval

**Precision and Reproducibility:**

**Intra-run experiments:** Two different characterized reference standard materials (Horizon QMRS™ and AcroMetrix Oncology Hotspot Control™) containing 140 SNVs (VAF 3-35%) and 10 indels (VAF 2-35%) were then processed in duplicate in the same batch and run. One SNV (*IDH2* c.474A>G) expected at VAF of 5% was not detected in one of the duplicate runs. All other variants were detected on both runs, with a mean coefficient of variation (CV) of 0.06% for allele frequencies.

**Inter-run experiments:** Horizon QMRS™ and AcroMetrix Oncology Hotspot Control™ reference materials containing 140 SNVs (VAF 3-35%) and 10 indels (VAF 2-35%) were then processed on three different batches and on three different runs. One indel (*PDGFRA* c.1698\_1912del15) expected at VAF of 8% was not detected in one of the three triplicate runs. All other variants were detected on both runs, with a mean coefficient of variation (CV) of 0.1% for allele frequencies. Of note the missed indel is below what will be reported as the sensitivity of the assay for indels (VAF ≥10%).

**Figure II-4.** Plot shows observed VAF in FFPE Horizon QMRS control in reproducibility studies. (Samples labeled Run '1a' and '1b' were sequenced on the same run and batch.)



Overall Intermediate Precision:

SNVs: VAF 5-10%: 99.3% [95.8% - 99.9%]; VAF >10%: 100% [98.9% - 100%]

InDels: VAF ≥10%: 100% [80%-100%]



### III. High Copy Number Amplification (DNA) Validation

Test Category	Material Used
Analytical Sensitivity (LOD)	DNA of 5 tumor cell lines (NCI-H69, HCC827, BT474, SNU16, and HCC1143) bearing 9 gene amplifications ( $\geq 8$ copies, ploidy < 4; 7 genes) were pooled with HapMap DNA (YRI population) in ratios ranging from low to high tumor content (20% to 100%), creating a total set of 56 different gene amplifications.
Accuracy and Specificity	
Accuracy and Specificity	26 FFPE tumor samples analyzed against orthogonal test method (Foundation One)
Precision and Reproducibility	HCC827 (20% and 50% dilutions) and HCC1143 (20% dilution) analyzed within and between runs

**Limit of Detection (LoD):** DNA from 5 tumor cell lines bearing 9 high copy number ( $\geq 8$ ) gene amplifications (7 different genes) were pooled with HapMAP DNA (YRI population) in ratios creating a total set of 56 different gene amplification (25 of which from low tumor content (20-30%) samples). High performance was achieved for amplification detection at lower tumor purities (20-30%), with an overall sensitivity of 96% [78%-100%] (24/25), see Table III-1 and III-2. For high copy number amplifications ( $\geq 8$ ), these experiments confirm a limit of detection of 30% tumor content for gene amplification detection.

**Table III-1.** Summary of copy number amplification detection performance in cell line dilutions.

Cell Line	Ploidy*	Gene	Copy Number*	Percentage of tumor cell line in pool with amplified gene			
				100%	50%	30%	20%
NCI-H69	1.6	MYCN	14	Detected	Detected	Detected	Detected
HCC827	2.8	CDK4	12	Detected	Detected	Detected	Detected
HCC827	2.8	EGFR	12	Detected	Detected	Detected	Detected
HCC827	2.8	MYC	8	Detected	Detected	Detected	Detected
BT474	2.7	ERBB2	14	Detected	Detected	Detected	Detected
SNU16	3.7	FGFR2	14	Detected	Detected	Detected	Detected
SNU16	3.7	MYC	14	Detected	Detected	Detected	Detected
HCC1143	3.4	CCND1	12	Detected	Detected	Detected	Detected
HCC1143	3.4	MYC	8	Detected	Detected	Detected	Detected

Ploidy and copy numbers from Cosmic Cell Line Project.

**Table III-2.** Summary of copy number amplification detection performance in cell line pools.

Cell Line	Percent of tumor cell line in pool containing amplified gene				
	HCC827		BT474		NCI-H69
Gene (Copy Number)	CDK4 (12x)	EGFR (12x)	MYC (8x)	ERBB2 (14x)	MYCN (14x)
Cell Line Pool 1	33% -detected	33% -detected	33% -detected	33% -detected	33% -detected
Cell Line Pool 2	40% -detected	40% -detected	40% -detected	20% -detected	40% -detected
Cell Line Pool 3	40% -detected	40% -detected	40% -detected	40% -detected	20% -detected
Cell Line Pool 4	20% -detected	20% -detected	20% -not called*	20% -detected	60% -detected

**Limits of Quantification and Reportable Range:** The clinically reportable range for high copy amplifications will be in samples with tumor purity ranging from 20-100% as determined by cell line sensitivity studies. These have been demonstrated to empirically hold in FFPE clinical tumor samples containing  $\geq 30\%$  tumor nuclei, see Table III-3.

#### Accuracy and Specificity:

**Table III-3 – Copy number amplification detection performance.**

	Tumor Fraction (20-30%)					Tumor Fraction ( $\geq 30\%$ )					Confirmation Method
	Total # of amplifications	False Negatives	False Positives	PPA [95% CI]	PPV [95% CI]	Total # of amplifications	False Negatives	False Positives	PPA [95% CI]	PPV [95% CI]	
FFPE Clinical Tumor Samples	7	1*	0	88% [47-99%]	100% [56-100%]	27	1**	0	96% [80-100%]	100% [85-100%]	NGS (Foundation One)
Cell line DNA Pools/Dilutions	24	1***	1	96% [78-100%]	96% [78-100%]	34	0	0	100% [87-100%]	100% [87-100%]	COSMIC Cancer Cell Project

\* MET amplification was detected at calculated 3-fold gain which is below the 5-fold gain need to be called positive by Oncomine pipeline.

\*\*KRAS amplification was detected at calculated 4.6-fold gain is below the 5-fold gain need to be called positive by Oncomine pipeline.

\*\*\*MYC amplification was detected at calculated 3.5-fold gain is below the 5-fold gain need to be called positive by Oncomine pipeline.

#### Precision and Reproducibility:

**Intra-run experiments:** (1) Two different cell lines (HCC827 and HCC1143) were diluted to 20% with HapMap DNA, and were then processed in duplicate in the same batch and run. The 5 expected high copy gene amplifications were correctly detected above threshold in all samples tested (mean CV=2.4% for calculated copy number) with 100% specificity. (2) HCC827 cell line was diluted to 50%, and were then processed in duplicate in the same batch and run. The 3 expected high copy amplifications

were correctly detected above threshold in all samples tested (mean CV=7.8% for calculated copy number) with 100% specificity.

Inter-run experiments:

(1) Two different cell lines (HCC827 and HCC1143) were diluted to 20% with HapMap DNA, and were then processed on three different batches and on 3 different runs. The 3 expected high copy amplifications were correctly detected above threshold in all samples tested (mean CV=3.8% for calculated copy number) with 100% specificity. (2) HCC827 cell line was diluted to 50%, and were then processed on two different batches and on different runs. The 3 expected high copy amplifications were correctly detected above threshold in all samples tested (mean CV=4.6% for calculated copy number) with 100% specificity.

Overall Intermediate Precision:

20-30% tumor nuclei 100% [74.6% - 100%]  
 >30% tumor nuclei 100% [62.8% - 100%]

**Results of Sequencing Metrics in FFPE Clinical Samples (DNA):**

**Table III-4.** Sequencing metrics observed based on testing in FFPE clinical samples. Minimum and maximum range of metrics are based on confirmed detected variants. The following thresholds were used to determine which variants to include in the analysis.

Metric	Threshold	Minimum Observed	Maximum Observed	Mean	Median
Tumor %	>20%	20%	70%	43%	40%
DNA Mean Depth	>300X	304X	3304X	1585X	1524X
DNA Uniformity	>90%	90.6%	98.9%	95.7%	95.8%
% Aligned Bases	>95%	96.9%	99.2%	98.6%	98.7%
Mutation Allele Frequency	>0.05	0.06	0.85	0.33	0.29
Tumor Copy Number	>5X	5X	62X	17X	11X

## Part 2 - Focused Solid Tumor Assay Validation Summary

Test Category	Material Used
Analytical Sensitivity (LOD)	Characterized reference material (TruQ1 5%; quantified by ddPCR) containing 15 variants ranging from 3-5% allele frequency were tested
Accuracy and Specificity	21 FFPE tumor samples analyzed against orthogonal test method(s)
	Characterized reference material (AcroMetrix Oncology Hotspot Control, AOHC) testing 196 SNVs (VAF ranging from 5-35%) and 8 indels (VAF 5-35%)
	HapMAP pooled dilution studies ((NA18511, NA18867, NA18924, NA19108, NA19147, NA19240)
Precision and Reproducibility	Characterized reference material (AOHC) were analyzed within and between runs for 5 replicates

**Limit of Detection (LoD):** Reference material from Horizon Diagnostics (TruQ1 5%) in which the variants were quantified by qPCR were analyzed. All 15 variants were detected and observed in the expected 95% CI for each variant. These results demonstrate an empirical limit of detection of VAF of 5%.


### Accuracy:

(1) One of the objectives was to evaluate the accuracy of the assay in its ability to detect relevant mutations in well-characterized reference standard material (AcroMetrix Oncology Hotspot Control™). All reference material have been reported to be validated by at least one alternative method. All 196 SNVs (VAF ranging from 5-35%) and 8 indels (VAF 5-35%) covered by this panel were detected. This analysis was repeated five times (triplicate in one run) and two other replicates on different run each to look at inter-run and intra-run reproducibility. There was good concordance as only 2 out of possible 1020 variants were missed (99.8% accuracy) with mean coefficient of variation (CV) of 9% for allele frequencies.

(2) The second objective was to evaluate of the accuracy of the assay in its ability to detect mutations in FFPE clinical samples which were confirmed by an alternate method. 21 FFPE samples were evaluated. All 28 SNVs and 13/16 CNVs were detected. Three lower level amplification (MYC, KRAS, CCND1, <7x) by OFA assay in one sample that did not meet the threshold to be called (4.5-5.5x). Given the low tumor percentage (20%) for the sample and that therapeutic significance, which this assay designed to answer, is based on very high copy number changes, this variance between assays is acceptable. For high copy number amplifications ( $\geq 8$ ), these experiments confirm a limit of detection of 30% tumor content for gene amplification detection.

**HapMAP SNP analysis:** HapMAP samples (NA18511, NA18867, NA18924, NA19108, NA19147, NA19240) were pooled and sequenced five times on two different runs. The expected variant frequency in each HapMAP was determined by using HapMap3 database. The expected MAF for each test base substitution in pooled samples was calculated based on the number of alternate alleles present in mix constituents and on mixing ratios. There was great concordance between expected and detected variant frequencies, see Figure I-1.

**Figure I-1.** Allele frequencies measured in pooled samples (y axis) match the frequencies expected based on the mixing ratios of characterized HapMap samples.



Appendix - Table 2A. Validation Sample Master List (FFPE Clinical Samples)

SampleID	Format	Tumor %	Tissue Type	Tissue	Tumor Type	Tumor Subtype	Expected Results	Expected Results Methodology	Applicable Validation Element	Concordant	Replicates
OFA_1	FFPE	40	Surgical	Ovary	Breast	Carcinoma NOS	MYC amp, CCND1 amp	NGS (FM)	Accuracy	Yes	
OFA_2	FFPE	40	Surgical	Lung	NSCLC	Lung Carcinoma NOS	MET exon 14 alt splice (3028_3208+1GG>TT)	NGS (FM)	Accuracy; Fusion; Precision; RNA stability	Yes	4
OFA_3	FFPE	60	Surgical, bx	Liver	Breast	Carcinoma NOS	MYC amp, CCND1 amp, FGFR1 amp, ESR1 Y537S	NGS (FM)	Accuracy	Yes	
OFA_4	FFPE	50	Surgical, bx	Lung	Breast	Carcinoma NOS	MYC amp	NGS (FM)	Accuracy	Yes	
OFA_5	FFPE	60	Surgical, bx	Lung	NSCLC	Lung Adenocarcinoma	MYC amp	Therascreen (EGFR CDx); NGS (FM)	Accuracy	Yes	
OFA_6	FFPE	40	Surgical, bx	Liver	Breast	Carcinoma NOS	PIK3CA H1047R	NGS (FM)	Accuracy	Yes	
OFA_7	FFPE	30	Surgical, bx	Lymph Node	NSCLC	Lung Carcinoma NOS	EML4-ALK fusion (Variant 3a/b), CDK4 amp	Ventana ALK IHC (FDA-approved CDx); NGS (FM)(ArcherDx)	Accuracy	Yes	
OFA_8	FFPE	30	Surgical, bx	Lung	NSCLC	Lung Adenocarcinoma	EGFR E709A, G719A; MET amp	Therascreen (EGFR CDx); NGS (FM)	Accuracy	Yes*	
OFA_9	FFPE	60	Surgical, bx	Liver	NSCLC	Non-Small Cell Carcinoma	EGFR amp; MET amp	Therascreen (EGFR CDx); NGS (FM); ALK FISH (CDx)	Accuracy	Yes	
OFA_10	FFPE	30	Surgical, bx	Lung	NSCLC	Non-Small Cell Carcinoma	MET amp, PIK3CA R1023Q	Therascreen (EGFR CDx); NGS (FM)	Accuracy	Yes	
OFA_11	FFPE	20	Surgical, bx	Lymph Node	NSCLC	Non-Small Cell Carcinoma	EML4-ALK fusion (Variant 3a/b), CDK4 amp	Ventana ALK IHC (FDA-approved CDx); NGS (FM)(ArcherDx)	Accuracy; Fusion; Precision; RNA stability	Yes	
OFA_12	FFPE	20	Surgical, bx	Pericardium	NSCLC	Lung Adenocarcinoma	EML4-ALK fusion (Variant 1), MYC amp	Ventana ALK IHC (FDA-approved CDx); NGS (FM)(ArcherDx)	Accuracy	Yes	4
OFA_13	FFPE	70	Cytology, Fluid	Pericardial Fluid	NSCLC	Lung Adenocarcinoma	EML4-ALK fusion (Variant 1), MYC amp	Ventana ALK IHC (FDA-approved CDx); NGS (FM)(ArcherDx)	Accuracy	Yes	
OFA_14	FFPE	40	Cytology, Fluid	Pericardial Fluid	NSCLC	Lung Adenocarcinoma	EML4-ALK fusion (Variant 3a/b)	Ventana ALK IHC (FDA-approved CDx); NGS (FM)(ArcherDx)	Accuracy	Yes	
OFA_15	FFPE	40	Surgical, bx	Lung	NSCLC	Lung Adenocarcinoma	MET amp; FGFR1 amp; MYC amp (equ)	NGS (FM)	Accuracy	Yes	
OFA_16	FFPE	40	Cytology, FNA	Lymph Node	NSCLC	Lung Adenocarcinoma	EGFR A289V, L858R; MET amp, MYC amp	NGS (FM)	Accuracy	Yes	
OFA_17	FFPE	30	Cytology, Fluid	Pericardial Fluid	NSCLC	Lung Adenocarcinoma	MET exon 14 alt splice	Therascreen (EGFR CDx); NGS (FM)	Accuracy; Fusion; Precision; RNA stability	Yes	4
OFA_18	FFPE	40	Surgical, bx	Lymph Node	NSCLC	Lung Adenocarcinoma	EML4-ALK fusion (Variant 5)	Ventana ALK IHC (FDA-approved CDx); NGS (FM)(ArcherDx)	Accuracy; Fusion; Precision; RNA stability	Yes	4
OFA_19	FFPE	50	Surgical	Pleura	NSCLC	Lung Adenocarcinoma	KIF5B-RET fusion, MYC amp	Therascreen (EGFR CDx); NGS (FM)(ArcherDx)	Accuracy; Fusion; Precision; RNA stability	Yes	4
OFA_20	FFPE	60	Surgical, bx	Trachea	NSCLC	Lung Carcinoma NOS	MET amp	Therascreen (EGFR CDx); NGS (FM)(ArcherDx)	Accuracy	Yes	
OFA_21	FFPE	40	Surgical	Heart	NSCLC	Lung Adenocarcinoma	BRAF G466V; CCND1 amp, CDK4 amp, KRAS amp	Therascreen (EGFR CDx); NGS (FM)	Accuracy	Yes	
OFA_22	FFPE	20	Surgical	Lung	NSCLC	Lung Adenocarcinoma	EGFR A767_V769dupASV	Therascreen (EGFR CDx); NGS (FM)	Accuracy	Yes	
OFA_23	FFPE	30	Surgical	Lung	NSCLC	Lung Adenocarcinoma	EGFR G719C, S768I; PIK3CA H1047L	Therascreen (EGFR CDx); NGS (FM)(Ampliseq)	Accuracy	Yes	
OFA_24	FFPE	50	Surgical	Lung	NSCLC	Lung Adenocarcinoma	BRAF G466V; CCND1 amp, CDK4 amp, KRAS amp	Therascreen (EGFR CDx); NGS (FM)	Accuracy	Yes	
OFA_25	FFPE	40	Surgical, bx	Lung	NSCLC	Lung Adenocarcinoma	KRAS G12A, amp; MYC amp	Therascreen (EGFR CDx); NGS (FM)	Accuracy	Yes*	
OFA_26	FFPE	60	Surgical, bx	Lung	NSCLC	Lung Adenocarcinoma	EGFR L858R	Therascreen (EGFR CDx)	Accuracy	Yes	
OFA_27	FFPE	30	Surgical, bx	Lung	NSCLC	Lung Adenocarcinoma	FGFR1 amp; CCND1 amp; MYC amp	Therascreen (EGFR CDx); NGS (FM)	Accuracy	Yes	
OFA_28	FFPE	50	Surgical, bx	Lung	NSCLC	Lung Adenocarcinoma	CCND1 amp; KRAS G12C	Therascreen (EGFR CDx); NGS (FM)	Accuracy	Yes	
OFA_29	FFPE	50	Surgical	Lung	NSCLC	Lung Adenocarcinoma	KRAS G12D	Therascreen (EGFR CDx); ALK FISH (ALK CDx)	Accuracy	Yes	
OFA_30	FFPE	60	Surgical, bx	Lung	Breast	Carcinoma NOS	ERBB2 amp; CCND1 amp	ERBB2 FISH; NGS (FM)	Accuracy	Yes	
OFA_31	FFPE	50	Surgical, bx	Lymph Node	NSCLC	Lung Adenocarcinoma	EGFR E746_A750del, T790M	Therascreen (EGFR CDx); NGS (Ampliseq)	Accuracy	Yes	
OFA_32	FFPE	50	Surgical, bx	Lung	NSCLC	Lung Adenocarcinoma	EGFR exon 19 deletion (EGFR E746_A750del)	Therascreen (EGFR CDx)	Accuracy	Yes	
OFA_33	FFPE	70	Cytology, Fluid	Pleural fluid	NSCLC	Lung Adenocarcinoma	EGFR exon 19 deletion (EGFR E746_A750del)	Therascreen (EGFR CDx); ALK FISH (CDx); NGS (Ampliseq)	Accuracy	Yes	
OFA_34	FFPE	60	Surgical, bx	Lymph Node	NSCLC	Lung Adenocarcinoma	EGFR exon 19 deletion	Therascreen (EGFR CDx); ALK FISH (CDx); NGS (Ampliseq)	Accuracy	Yes	
OFA_35	FFPE	60	Surgical	Lung	NSCLC	Lung Adenocarcinoma	EGFR exon 19 deletion	Therascreen (EGFR CDx); ALK FISH (CDx)	Accuracy	Yes	
OFA_36	FFPE	30	Surgical, bx	Sacrum	NSCLC	Lung Adenocarcinoma	EGFR exon 19 deletion (EGFR E746_A750del)	Therascreen (EGFR CDx); NGS (FM)	Accuracy	Yes	
OFA_37	FFPE	20	Cytology, FNA	Lymph Node	NSCLC	Lung Adenocarcinoma	negative	Therascreen (EGFR CDx)	Accuracy	Yes	
OFA_38	FFPE	40	Surgical, bx	Lung	NSCLC	Lung Adenocarcinoma	EGFR L858R	Therascreen (EGFR CDx)	Accuracy	Yes	
OFA_39	FFPE	30	Surgical, bx	Brain	NSCLC	Lung Adenocarcinoma	EGFR L858R	Therascreen (EGFR CDx); ALK FISH (ALK CDx)	Accuracy	Yes	
OFA_40	FFPE	40	Surgical, bx	Lymph Node	NSCLC	Lung Adenocarcinoma	EGFR exon 19 deletion	Therascreen (EGFR CDx); ALK FISH (ALK CDx)	Accuracy	Yes	
OFA_41	FFPE	50	Cytology, FNA	Lymph Node	NSCLC	Lung Adenocarcinoma	negative	Therascreen (EGFR CDx)	Accuracy	Yes	
OFA_42	FFPE	40	Surgical, bx	Pericardial Tissue	NSCLC	Lung Adenocarcinoma	negative	Therascreen (EGFR CDx); ALK FISH (ALK CDx)	Accuracy	Yes	
OFA_43	FFPE	20	Surgical, bx	Chest Wall	NSCLC	Lung Adenocarcinoma	EGFR E746_A750del	Therascreen (EGFR CDx)	Accuracy	Yes	
OFA_44	FFPE	40	Surgical, bx	Lymph Node	NSCLC	Lung Adenocarcinoma	KRAS G12C	Therascreen (EGFR CDx); ALK FISH (ALK CDx)	Accuracy	Yes	

Appendix - Table 2B. Validation Sample Master List (Cell Lines)

SampleID	Format	Cell Line(s)	Tumor Type	Tumor Subtype	Expected Results	Expected Results Methodology	Applicable Validation Element	Concordant	Replicates
HCC827_100	DNA	HCC827 cell line	NSCLC	Lung Adenocarcinoma	EGFR amp, CDK4 amp, MYC amp, EGFR 746_750del (VAF >10%)	Cell Lines (Cosmic Cell Line Project)	CNV: Accuracy	Yes	1
HCC827_50	DNA	HCC827/HapMap (50:50)	NSCLC	Lung Adenocarcinoma	EGFR amp, CDK4 amp, MYC amp, EGFR 746_750del (VAF >10%)	Cell Lines (Cosmic Cell Line Project)	CNV: Accuracy, LoD, Precision	Yes	3
HCC827_30	DNA	HCC827/HapMap (30:70)	NSCLC	Lung Adenocarcinoma	EGFR amp, CDK4 amp, MYC amp, EGFR 746_750del (VAF >10%)	Cell Lines (Cosmic Cell Line Project)	CNV: Accuracy, LoD	Yes	1
HCC827_20	DNA	HCC827/HapMap (20:80)	NSCLC	Lung Adenocarcinoma	EGFR amp, CDK4 amp, MYC amp, EGFR 746_750del (VAF >10%)	Cell Lines (Cosmic Cell Line Project)	CNV: Accuracy, LoD, Precision	Yes	4
BT474_100	DNA	BT-474 cell line	Breast Carcinoma	Breast Carcinoma	ERBB2 amp, PIK3CA K111N (VAF >10%)	Cell Lines (Cosmic Cell Line Project)	CNV: Accuracy	Yes	1
BT474_50	DNA	BT-474/HapMap (50:50)	Breast Carcinoma	Breast Carcinoma	ERBB2 amp, PIK3CA K111N (VAF >10%)	Cell Lines (Cosmic Cell Line Project)	CNV: Accuracy, LoD	Yes	1
BT474_30	DNA	BT-474/HapMap (30:70)	Breast Carcinoma	Breast Carcinoma	ERBB2 amp, PIK3CA K111N (VAF >10%)	Cell Lines (Cosmic Cell Line Project)	CNV: Accuracy, LoD	Yes	1
BT474_20	DNA	BT-474/HapMap (20:80)	Breast Carcinoma	Breast Carcinoma	ERBB2 amp, PIK3CA K111N (VAF >10%)	Cell Lines (Cosmic Cell Line Project)	CNV: Accuracy, LoD	Yes	1
HCC1143_100	DNA	HCC1143 cell line	Breast Carcinoma	Breast Carcinoma	CCND1 amp; MYC amp	Cell Lines (Cosmic Cell Line Project)	CNV: Accuracy	Yes	1
HCC1143_50	DNA	HCC1143/HapMap (50:50)	Breast Carcinoma	Breast Carcinoma	CCND1 amp; MYC amp	Cell Lines (Cosmic Cell Line Project)	CNV: Accuracy, LoD	Yes	1
HCC1143_30	DNA	HCC1143/HapMap (30:70)	Breast Carcinoma	Breast Carcinoma	CCND1 amp; MYC amp	Cell Lines (Cosmic Cell Line Project)	CNV: Accuracy, LoD	Yes	1
HCC1143_20	DNA	HCC1143/HapMap (20:80)	Breast Carcinoma	Breast Carcinoma	CCND1 amp; MYC amp	Cell Lines (Cosmic Cell Line Project)	CNV: Accuracy, LoD, Precision	Yes	3
H69_100	DNA	HTB-119D(NCI-H69) cell line	NSCLC	Lung Adenocarcinoma	MYCN amp, PIK3CA 106_108del (VAF >20%)	Cell Lines (Cosmic Cell Line Project)	CNV, InDel: Accuracy	Yes	1
H69_50	DNA	HTB-119D(NCI-H69)/HapMap (50:50)	NSCLC	Lung Adenocarcinoma	MYCN amp, PIK3CA 106_108del (VAF >20%)	Cell Lines (Cosmic Cell Line Project)	CNV, InDel: Accuracy, LoD	Yes	1
H69_30	DNA	HTB-119D(NCI-H69)/HapMap (30:70)	NSCLC	Lung Adenocarcinoma	MYCN amp, PIK3CA 106_108del (VAF >20%)	Cell Lines (Cosmic Cell Line Project)	CNV, InDel: Accuracy, LoD	Yes	1
H69_20	DNA	HTB-119D(NCI-H69)/HapMap (20:80)	NSCLC	Lung Adenocarcinoma	MYCN amp, PIK3CA 106_108del (VAF >20%)	Cell Lines (Cosmic Cell Line Project)	CNV, InDel: Accuracy, LoD	Yes	1
SNU16_100	DNA	CRL-5974D (SNU-16) cell line	Stomach Carcinoma	Stomach Carcinoma	FGFR2 amp; MYC amp	Cell Lines (Cosmic Cell Line Project)	CNV: Accuracy	Yes	1
SNU16_50	DNA	CRL-5974D (SNU-16)/HapMap (50:50)	Stomach Carcinoma	Stomach Carcinoma	FGFR2 amp; MYC amp	Cell Lines (Cosmic Cell Line Project)	CNV: Accuracy, LoD	Yes	1
SNU16_30	DNA	CRL-5974D (SNU-16)/HapMap (30:70)	Stomach Carcinoma	Stomach Carcinoma	FGFR2 amp; MYC amp	Cell Lines (Cosmic Cell Line Project)	CNV: Accuracy, LoD	Yes	1
SNU16_20	DNA	CRL-5974D (SNU-16)/HapMap (20:80)	Stomach Carcinoma	Stomach Carcinoma	FGFR2 amp; MYC amp	Cell Lines (Cosmic Cell Line Project)	CNV: Accuracy, LoD	Yes	1
CL Pool 1	DNA	HCC827/NCI-H69/BT-474 (1:1:1)	Cell Line Pool	Cell Line Pool	MYCN amp; ERBB2 amp; CDK4 amp; MYC amp; PIK3CA K111N (VAF >10%); EGFR amp, p.746_750del (VAF >10%)	Cell Lines (Cosmic Cell Line Project)	CNV, InDel: Accuracy, LoD	Yes	1
CL Pool 2	DNA	HCC827/NCI-H69/BT-474 (2:2:1)	Cell Line Pool	Cell Line Pool	MYCN amp; ERBB2 amp; CDK4 amp; MYC amp; PIK3CA K111N (VAF 5-10%); 106_108del (VAF <20%); EGFR amp, 746_750del (VAF >10%)	Cell Lines (Cosmic Cell Line Project)	CNV, InDel: Accuracy, LoD	Yes	1
CL Pool 3	DNA	HCC827/NCI-H69/BT-474 (2:1:2)	Cell Line Pool	Cell Line Pool	MYCN amp; ERBB2 amp; CDK4 amp; MYC amp; EGFR p.746_750del (VAF >20%); amp; PIK3CA K111N (VAF >10%)	Cell Lines (Cosmic Cell Line Project)	CNV, InDel: Accuracy, LoD	Yes	1
CL Pool 4	DNA	HCC827/NCI-H69/BT-474 (1:1:3)	Cell Line Pool	Cell Line Pool	MYCN amp; ERBB2 amp; CDK4 amp; MYC amp; PIK3CA K111N (VAF 5-10%); 106_108del (VAF >20%); EGFR amp, p.746_750del (VAF >20%)	Cell Lines (Cosmic Cell Line Project)	CNV, InDel: Accuracy, LoD	Yes	1
H1975 D-1	DNA	H1975/H2228 cell lines (100:0)	NSCLC	Lung Adenocarcinoma	EGFR L858R, T790M (VAFs >10%); MYC amp (12x)	Cell Lines (Cosmic Cell Line Project)	SNV: LoD	Yes	1
H1975 D-2	DNA	H1975/H2228 cell lines (50:50)	NSCLC	Lung Adenocarcinoma	EGFR L858R, T790M (VAFs >10%); MYC amp (6x)	Cell Lines (Cosmic Cell Line Project)	SNV: LoD	Yes	1
H1975 D-3	DNA	H1975/H2228 cell lines (25:75)	NSCLC	Lung Adenocarcinoma	EGFR L858R, T790M (VAFs 5-10%); MYC amp (4x)	Cell Lines (Cosmic Cell Line Project)	SNV: LoD	Yes	1
H1975 D-4	DNA	H1975/H2228 cell lines (12:88)	NSCLC	Lung Adenocarcinoma	EGFR L858R, T790M (VAFs <5%); MYC amp (2x)	Cell Lines (Cosmic Cell Line Project)	SNV: LoD	No (variants below threshold)	2
H1975 D-7	DNA	H1975/H2228 cell lines (0:100)	NSCLC	Lung Adenocarcinoma	negative	Cell Lines (Cosmic Cell Line Project)	SNV: LoD	Yes	1
RNA neg ctrl	RNA	Negative Control	NA	Normal Lung Tissue	negative	NGS (ArcherDx Lugn Panel)	Fusion Specificity	Yes	3
H2228 R_100	RNA	H2228/H1975 cell lines (100:0)	NSCLC	Lung Adenocarcinoma	EML4-ALK fusion Variant 3a/b, ALK-PTPN3 fusion	Cell Lines (Cosmic Cell Line Project)	Fusion: Accuracy	Yes	1
H2228 R_50	RNA	H2228/H1975 cell lines (50:50)	NSCLC	Lung Adenocarcinoma	EML4-ALK fusion Variant 3a/b, ALK-PTPN3 fusion	Cell Lines (Cosmic Cell Line Project)	Fusion: LoD, Precision	Yes	3
H2228 R_25	RNA	H2228/H1975 cell lines (25:75)	NSCLC	Lung Adenocarcinoma	EML4-ALK fusion Variant 3a/b, ALK-PTPN3 fusion	Cell Lines (Cosmic Cell Line Project)	Fusion: LoD, Precision	Yes	3
H2228 R_12	RNA	H2228/H1975 cell lines (12:88)	NSCLC	Lung Adenocarcinoma	EML4-ALK fusion Variant 3a/b, ALK-PTPN3 fusion	Cell Lines (Cosmic Cell Line Project)	Fusion: LoD, Precision	Yes	5
H2228 R_6	RNA	H2228/H1975 cell lines (6:94)	NSCLC	Lung Adenocarcinoma	EML4-ALK fusion Variant 3a/b, ALK-PTPN3 fusion	Cell Lines (Cosmic Cell Line Project)	Fusion: LoD, Precision	Yes	3
H2228 R_3	RNA	H2228/H1975 cell lines (3:97)	NSCLC	Lung Adenocarcinoma	EML4-ALK fusion Variant 3a/b, ALK-PTPN3 fusion	Cell Lines (Cosmic Cell Line Project)	Fusion: LoD, Precision	Yes	2
H2228 R_0	RNA	H2228/H1975 cell lines (0:100)	NSCLC	Lung Adenocarcinoma	negative	Cell Lines (Cosmic Cell Line Project)	Fusion Specificity	Yes	1

Appendix - Table 2C. Validation Sample Master List (Standard Reference Material)

SampleID	Format	Reference Standard	Vendor	Expected Results	Expected Results Methodology	Applicable Validation Element	Concordant	Replicates
QMRS_1	FFPE	Quantitative Multiplex Reference Standard	HdX (HD200)	6 SNVs (VAF 2-10%) and 4 SNVs (VAF>10%)	Reference material (confirmed by ddPCR)	SNV/InDel: Accuracy; precision	Yes	4
CAP Pool_1_1	DNA	Pool of 5 CAP proficiency samples	CAP (MTP-4.5.6; NGSST01.03)	10 SNVs (VAF 5-10%) and 4 SNVs (VAF>10%)	Reference material (confirmed by NGS Ampliseq)	SNV/InDel: Accuracy; precision	Yes	3
CAP Pool2_1a	DNA	Pool of 2 CAP proficiency samples	CAP (MTP-6; NGSST01)	7 SNVs (VAF>10%)	Reference material (confirmed by NGS Ampliseq)	SNV/InDel: Accuracy; precision	Yes	2
Tru-Q1	DNA	Tru-Q1 (5% Tier)-Reference Standard	HdX (catalog# HD728)	7 SNVs (VAF 4-8%) and 3 SNVs (VAF>10%)	Reference material (confirmed by ddPCR)	SNV/InDel: Accuracy/sensitivity	Yes	1
Tru-Q2	DNA	Tru-Q2 (5% Tier)-Reference Standard	HdX (catalog# HD729)	10 SNVs (VAF 4-8%) and 3 SNVs (VAF>10%)	Reference material (confirmed by ddPCR)	SNV/InDel: Accuracy/sensitivity	Yes	1
Tru-Q3	DNA	Tru-Q3 (5% Tier)-Reference Standard	HdX (catalog# HD730)	9 SNVs (VAF 4-8%) and 3 SNVs (VAF>10%)	Reference material (confirmed by ddPCR)	SNV/InDel: Accuracy/sensitivity	Yes	1
Tru-Q4	DNA	Tru-Q4 (5% Tier)-Reference Standard	HdX (catalog# HD731)	10 SNVs (VAF 4-8%) and 3 SNVs (VAF>10%)	Reference material (confirmed by ddPCR)	SNV/InDel: Accuracy/sensitivity	Yes	1
AOHC_1	DNA	AcroMatrix Oncology Hotspot Control	ThermoFisher (catalog #969056)	130 SNVs and 10 Indels (VAF 5-35%)	Reference material (confirmed by sanger)	SNV/InDel: Accuracy; precision	Yes	4
Seracare RNA_1	FFPE	FFPE Tumor Fusion RNA Reference Material v1	Seracare	10 different fusions	Reference material	Fusion: Accuracy; precision	Yes	3