








**Table S2**

Validation of the dPCR mutation analysis.

Diluted sample	#1	#2	#3	#4	#5	#6	#7
DNA of MIA Pa Ca-2 (ng)	10	1	0.1	0.01	0.001	0.0001	0
DNA of 293T (ng)	0	9	9.9	9.99	9.999	9.9999	10
Estimated <i>KRAS</i> mutation (%)	100	10	1	0.1	0.01	0.001	0
Results of dPCR							
Detected <i>KRAS</i> mutation (%)	100	9.73	0.98	0.09	0	0	0

The *KRAS* mutations were detected by dPCR. The MIA PaCa-2 DNA sample was serially diluted by adding that of 293T.

