

Direct mass spectrometry analysis of protein complexes and intact proteins up to >70 kDa from tissue

Rian L. Griffiths^{1#}, Albert Konijnenberg², Rosa Viner³ and Helen J. Cooper^{1*}

¹School of Biosciences, University of Birmingham, Edgbaston, Birmingham B15 2TT, UK.

²Thermo Fisher Scientific, Achtseweg Noord 5, 5651 GG Eindhoven, The Netherlands

³Thermo Fisher Scientific, River Oaks Parkway 355, San Jose, 95134, California, U.S.A

#Current address: School of Pharmacy, University of Nottingham, University Park, Nottingham NG7 2RD, UK.

*To whom correspondence should be addressed: h.j.cooper@bham.ac.uk

Supplementary Information:

Supplemental Figure 1. Comparison of deconvoluted native LESA mass spectra from **A)** two separate cerebellum locations sampled with 10 mM ammonium acetate, 5% MeOH; **B)** two separate cerebellum locations sampled with 10 mM ammonium acetate, 0.125% C8E4; **C)** two separate cerebellum locations one sampled with 10 mM ammonium acetate, 5% MeOH and one with 10 mM ammonium acetate, 0.125% C8E4; **D)** one cerebellum location and one cerebral cortex location sampled with 10 mM ammonium acetate, 0.125% C8E4; **E)** two separate medulla (kidney) locations one sampled with 10 mM ammonium acetate, 0.125% C8E4; **F)** one medulla (kidney) location and one cortex (kidney) location sampled with 10 mM ammonium acetate, 0.125% C8E4.

Supplemental Figure 2. Native LESA MS of rat brain tissue. A) Sampling location: cerebral cortex. LESA extraction/ionisation solvent: 10 mM ammonium acetate, 5% MeOH. B) Detail from main figure 1C. Sampling location cerebral cortex. LESA extraction/ionisation solvent: 10 mM ammonium acetate, 0.125% C8E4.

Supplemental Figure 3. Native LESA MS of rat kidney tissue. Sampling location: outer cortex. LESA extraction/ionisation solvent: 10 mM ammonium acetate, 5% MeOH.

Supplemental Figure 4. Native LESA HCD MS/MS of rat kidney tissue. Sampling location: outer cortex. LESA extraction/ionisation solvent: 10 mM ammonium acetate, 5% MeOH. HCD energy = 80%.

Supplemental File 1: Biopharma Finder search parameters and outputs for LESA mass spectrum shown in Fig 1A. (Cerebellum, 10 mM ammonium acetate, 5% MeOH).

Supplemental File 2: Biopharma Finder search parameters and outputs for LESA mass spectrum obtained from separate cerebellum location (Cerebellum, 10 mM ammonium acetate, 5% MeOH).

Supplemental File 3: Biopharma Finder search parameters and outputs for LESA mass spectrum shown in Fig 1B. (Cerebellum, 10 mM ammonium acetate, 0.125% C8E4).

Supplemental File 4: Biopharma Finder search parameters and outputs for LESA mass spectrum shown in Fig 1C. (Cerebral cortex, 10 mM ammonium acetate, 0.125% C8E4).

Supplemental File 5: Biopharma Finder search parameters and outputs for LESA mass spectrum shown in Fig 2A. (Kidney outer cortex, 10 mM ammonium acetate, 0.125% C8E4).


Supplemental File 6: Biopharma Finder search parameters and outputs for LESA mass spectrum shown in Suppl Fig 3. (Kidney outer cortex, 10 mM ammonium acetate, 5% MeOH).

Supplemental File 7: Biopharma Finder search parameters and outputs for LESA mass spectrum shown in Fig 2C. (Kidney medulla, 10 mM ammonium acetate, 0.125% C8E4).

Supplemental File 8: Biopharma Finder search parameters and outputs for LESA mass spectrum obtained from a separate medulla location. (Kidney medulla, 10 mM ammonium acetate, 0.125% C8E4).

A)


NativeLESA_method_glassslide_undosed_brain_5pMeOH_P4_run2.RAW ReSpect™ (Isotopically Unresolved) NL: 7.60E+006
 NativeLESA_method_glassslide_undosed_brain_5pMeOH_P4_run2.RAW



NativeLESA_method_glassslide_undosed_brain_5pMeOH_P7_run20.RAW ReSpect™ (Isotopically Unresolved) NL: 7.32E+006
 NativeLESA_method_glassslide_undosed_brain_5pMeOH_P7_run20.RAW

B)

NativeLESA_method_glassslide_rat_brain_5pMeOH_P4_run11.RAW ReSpect™ (Isotopically Unresolved) NL: 6.55E+006
 NativeLESA_method_glassslide_rat_brain_5pMeOH_P4_run11.RAW




NativeLESA_method_glassslide_undosed_brain_5pMeOH_P7_run19.RAW ReSpect™ (Isotopically Unresolved) NL: 7.11E+006
 NativeLESA_method_glassslide_undosed_brain_5pMeOH_P7_run19.RAW

Supplemental Figure 1. Comparison of deconvoluted native LESA mass spectra from **A)** two separate cerebellum locations sampled with 10 mM ammonium acetate, 5% MeOH and **B)** two separate cerebellum locations sampled with 10 mM ammonium acetate, 0.125% C8E4

C)


NativeLESA_method_glassslide_undosed_brain_5pMeOH_P4_run2.RAW ReSpect™ (Isotopically Unresolved) NL: 7.60E+006
 NativeLESA_method_glassslide_undosed_brain_5pMeOH_P4_run2.RAW



NativeLESA_method_glassslide_undosed_brain_5pMeOH_P7_run19.RAW ReSpect™ (Isotopically Unresolved) NL: 7.11E+006
 NativeLESA_method_glassslide_undosed_brain_5pMeOH_P7_run19.RAW

D)

NativeLESA_method_glassslide__rat_brain_5pMeOH_P4_run8.RAW ReSpect™ (Isotopically Unresolved) NL: 1.41E+007
 NativeLESA_method_glassslide__rat_brain_5pMeOH_P4_run8.RAW




NativeLESA_method_glassslide_undosed_brain_5pMeOH_P7_run19.RAW ReSpect™ (Isotopically Unresolved) NL: 7.11E+006
 NativeLESA_method_glassslide_undosed_brain_5pMeOH_P7_run19.RAW

Supplemental Figure 1 ctd. Comparison of deconvoluted native LESA mass spectra from **C)** two separate cerebellum locations one sampled with 10 mM ammonium acetate, 5% MeOH and one with 10 mM ammonium acetate, 0.125% C8E4 **D)** one cerebellum location and one cerebral cortex location sampled with 10 mM ammonium acetate, 0.125% C8E4 .

E)


NativeLESA_undosed_rat_kidney_run3.RAW ReSpect™ (Isotopically Unresolved) NL: 1.27E+007
 NativeLESA_undosed_rat_kidney_run3.RAW



NativeLESA_undosed_rat_kidney_run9.RAW ReSpect™ (Isotopically Unresolved) NL: 1.98E+007
 NativeLESA_undosed_rat_kidney_run9.RAW

F)

NativeLESA_undosed_rat_kidney_run9.RAW ReSpect™ (Isotopically Unresolved) NL: 1.98E+007
 NativeLESA_undosed_rat_kidney_run9.RAW




NativeLESA_undosed_rat_kidney_run5.RAW ReSpect™ (Isotopically Unresolved) NL: 2.95E+007
 NativeLESA_undosed_rat_kidney_run5.RAW



Cortex location 1;
10mM ammonium acetate, 0.125 % C8E4

Supplemental Figure 1 ctd. Comparison of deconvoluted native LESA mass spectra from **E)** two separate medulla (kidney) locations one sampled with 10 mM ammonium acetate, 0.125% C8E4 and **F)** one medulla (kidney) location and one cortex (kidney) location sampled with 10 mM ammonium acetate, 0.125% C8E4 .


A) Cerebral Cortex sampled without detergent




B) Cerebral Cortex sampled with detergent



Supplemental Figure 2. Native LESA MS of rat brain tissue. A) Sampling location: cerebral cortex. LESA extraction/ionisation solvent: 10 mM ammonium acetate, 5% MeOH. B) Detail from main figure 1C. Sampling location cerebral cortex. LESA extraction/ionisation solvent: 10 mM ammonium acetate, 0.125% C8E4.



Supplemental Figure 3. Native LESA MS of rat kidney tissue. Sampling location: outer cortex. LESA extraction/ionisation solvent: 10 mM ammonium acetate, 5% MeOH.



Supplemental Figure 4. Native LESA HCD MS/MS of rat kidney tissue. Sampling location: outer cortex. LESA extraction/ionisation solvent: 10 mM ammonium acetate, 5% MeOH. HCD energy = 80%.

Supplemental File 1: Biopharma Finder search parameters and outputs for LESA mass spectrum shown in Fig 1A. (Cerebellum, 10 mM ammonium acetate, 5% MeOH).

BioPharma Finder Report

Created: 08/04/2019 09:56:45


Sample Information

Raw File Name	C:\Xcalibur\Data\ms2LESA\NativeLESA_method_glassslide_undosed_brain_5pMeOH_P4_run2.RAW
Instrument Method	
Vial	
Injection Volume (µL)	0
Sample Weight	0
Sample Volume (µL)	0
ISTD Amount	0
Dil Factor	0

Chromatogram Parameters


Use Restricted Time	False
Time Limits	0.023 - 7.169 minutes
Scan Range	1 - 334
m/z Range	400 - 20000
Chromatogram Trace Type	TIC
Sensitivity	High
Rel. Intensity Threshold (%)	1

Chromatogram



Main Parameters (ReSpect™)	
Deconvolution Results Filter	
Output Mass Range	5000 - 100000
Deconvoluted Spectra Display Mode	Isotopic Profile (new)
Charge State Distribution	
Deconvolution Mass Tolerance	30 ppm
Choice of Peak Model	
Choice of Peak Model	Intact Protein
Resolution at 400 m/z	
Raw File Specific	6187
Generate XIC for Each Component	
Calculate XIC	False
Advanced Parameters (ReSpect™)	
Charge State Distribution	
Model Mass Range	10000 - 120000
Charge State Range	5 - 40
Minimum Adjacent Charges (low & high model mass)	3 - 3
Noise Parameters	
Rel. Abundance Threshold (%)	15
Deconvolution Quality	
Quality Score Threshold	0
Choice of Peak Model	
Target Mass	120000 Da
Peak Model Parameters	
Number of Peak Models	1
Left/Right Peak Shape	2:2
Peak Filter Parameters	
Peak Detection Minimum Significance Measure	1 Standard Deviations
Peak Detection Quality Measure	95%
Specialized Parameters	
Peak Model Width Factor	1
Intensity Threshold Scale	0.01
Deconvolution Parameters	
Noise Compensation	True
Charge Carrier	H
Negative Charge	False
Source Spectra Parameters	
Source Spectra Method	Sliding Windows
Sliding Windows Definition	
RT Range	0.500 - 3.000 minutes
Target Avg Spectrum Width	0.2 minutes
Target Avg Spectrum Offset Scan	1
Sliding Windows Merging Parameters	
Merge Tolerance	30 ppm
Max RT Gap	0.5 minutes
Min. Number of Detected Intervals	40

NativeLESA_method_glassslide_undosed_brain_5pMeOH_P4_run2



Sliding Windows ReSpect Masses Table

Row Number	Average Mass	Sum Intensity	Relative Abundance	Fractional Abundance	Score	Number of Charge States	Charge State Distribution	Number of Detected Intervals	Mass Std Dev	PPM Std Dev	Delta Mass	Scan Range	Start Time (min)	Stop Time (min)	Apex RT
1	51710.05	7604135.0 ₉	100.00	28.79	23.63	5	13 - 17	67	4.50	87.11	0.00	23 - 138	0.495	2.984	0.849
2	41712.63	4073971.5 ₂	53.58	15.43	21.32	4	12 - 15	102	2.57	61.59	-9997.42	24 - 135	0.516	2.920	1.171
3	71287.88	3610938.8 ₀	47.49	13.67	19.10	3	16 - 18	108	4.65	65.20	19577.84	23 - 138	0.495	2.984	0.999
4	41655.92	2885603.6 ₈	37.95	10.93	22.57	5	11 - 15	106	2.15	51.62	-10054.13	25 - 138	0.538	2.984	0.892
5	41779.80	1954143.2 ₉	25.70	7.40	24.03	5	11 - 15	88	2.96	70.86	-9930.25	37 - 136	0.795	2.941	1.343
6	42589.13	1479714.7 ₈	19.46	5.60	20.84	5	10 - 14	78	3.38	79.36	-9120.91	26 - 138	0.559	2.984	1.450
7	32040.57	1239853.2 ₃	16.30	4.69	16.68	3	8 - 10	51	2.15	67.16	-19669.48	37 - 100	0.795	2.169	1.600
8	65309.77	950060.43	12.49	3.60	21.82	4	17 - 20	58	5.48	83.87	13599.72	72 - 138	1.546	2.984	1.943
9	42656.79	761224.11	10.01	2.88	17.86	3	11 - 13	49	3.45	80.79	-9053.26	23 - 138	0.495	2.984	0.591
10	34900.39	690711.47	9.08	2.62	18.11	3	8 - 10	50	2.43	69.71	-16809.65	74 - 134	1.589	2.899	1.836
11	80462.07	669635.99	8.81	2.54	18.69	3	23 - 25	58	2.64	32.84	28752.02	58 - 131	1.246	2.834	2.094
12	67902.81	491095.13	6.46	1.86	15.81	3	21 - 23	41	3.69	54.36	16192.76	56 - 125	1.203	2.705	2.029

Supplemental File 2: Biopharma Finder search parameters and outputs for LESA mass spectrum obtained from separate cerebellum location (Cerebellum, 10 mM ammonium acetate, 5% MeOH).

BioPharma Finder Report

Created: 08/04/2019 10:20:13


Sample Information

Raw File Name	C:\Xcalibur\Data\ms2LESA\NativeLESA_method_glassslide_undosed_brain_5pMeOH_P7_run20.RAW
Instrument Method	
Vial	
Injection Volume (µL)	0
Sample Weight	0
Sample Volume (µL)	0
ISTD Amount	0
Dil Factor	0

Chromatogram Parameters

Use Restricted Time	False
Time Limits	0.023 - 7.169 minutes
Scan Range	1 - 334
m/z Range	400 - 20000
Chromatogram Trace Type	TIC
Sensitivity	High
Rel. Intensity Threshold (%)	1

Chromatogram



Main Parameters (ReSpect™)

Deconvolution Results Filter	
Output Mass Range	5000 - 100000
Deconvoluted Spectra Display Mode	Isotopic Profile (new)
Charge State Distribution	
Deconvolution Mass Tolerance	30 ppm
Choice of Peak Model	
Choice of Peak Model	Intact Protein
Resolution at 400 m/z	
Raw File Specific	6187
Generate XIC for Each Component	
Calculate XIC	False


Advanced Parameters (ReSpect™)

Charge State Distribution	
Model Mass Range	10000 - 120000
Charge State Range	5 - 40
Minimum Adjacent Charges (low & high model mass)	3 - 3
Noise Parameters	
Rel. Abundance Threshold (%)	15
Deconvolution Quality	
Quality Score Threshold	0
Choice of Peak Model	
Target Mass	120000 Da
Peak Model Parameters	
Number of Peak Models	1
Left/Right Peak Shape	2:2
Peak Filter Parameters	
Peak Detection Minimum Significance Measure	1 Standard Deviations
Peak Detection Quality Measure	95%
Specialized Parameters	
Peak Model Width Factor	1
Intensity Threshold Scale	0.01
Deconvolution Parameters	
Noise Compensation	True
Charge Carrier	H
Negative Charge	False

Source Spectra Parameters

Source Spectra Method	Sliding Windows
Sliding Windows Definition	
RT Range	2.000 - 9.000 minutes
Target Avg Spectrum Width	0.2 minutes
Target Avg Spectrum Offset Scan	1
Sliding Windows Merging Parameters	
Merge Tolerance	30 ppm
Max RT Gap	0.5 minutes
Min. Number of Detected Intervals	100

NativeLESA_method_glassslide_undosed_brain_5pMeOH_P7_run20



Sliding Windows ReSpect Masses Table

Row Number	Average Mass	Sum Intensity	Relative Abundance	Fractional Abundance	Score	Number of Charge States	Charge State Distribution	Number of Detected Intervals	Mass Std Dev	PPM Std Dev	Delta Mass	Scan Range	Start Time (min)	Stop Time (min)	Apex RT
1	42721.85	7320974.1 9	100.00	27.11	19.39	4	11 - 14	275	2.90	67.78	0.00	93 - 417	1.997	8.972	2.308
2	36392.08	4033161.7 6	55.09	14.93	26.19	6	7 - 12	287	2.70	74.18	-6329.77	93 - 416	1.997	8.951	2.995
3	42653.26	3874901.3 3	52.93	14.35	22.10	4	11 - 14	195	3.18	74.51	-68.59	93 - 345	1.997	7.427	2.802
4	51708.12	3572900.9 8	48.80	13.23	18.93	4	13 - 16	281	2.82	54.58	8986.27	129 - 417	2.770	8.972	3.038
5	42783.14	3141993.2 1	42.92	11.63	17.54	4	10 - 13	253	3.41	79.81	61.29	93 - 356	1.997	7.663	2.459
6	93214.90	1344648.8 9	18.37	4.98	27.12	6	24 - 29	128	7.84	84.10	50493.05	116 - 282	2.491	6.075	3.596
7	27814.45	1249522.3 5	17.07	4.63	18.36	3	8 - 10	195	0.97	34.93	-14907.40	124 - 328	2.662	7.062	3.360
8	42206.97	1115973.1 0	15.24	4.13	18.89	4	10 - 13	114	3.16	74.94	-514.88	131 - 255	2.813	5.495	3.746
9	70843.08	832450.64	11.37	3.08	19.79	4	17 - 20	110	6.70	94.63	28121.23	194 - 377	4.165	8.114	4.862
10	41966.52	519591.40	7.10	1.92	17.76	3	11 - 13	116	2.18	52.04	-755.33	164 - 297	3.521	6.397	3.789

Supplemental File 3: Biopharma Finder search parameters and outputs for LESA mass spectrum shown in Fig 1B. (Cerebellum, 10 mM ammonium acetate, 0.125% C8E4).

BioPharma Finder Report

Created: 08/04/2019 11:20:38


Sample Information

Raw File Name	C:\Xcalibur\Data\ms2LESA\NativeLESA_method_glassslide_undosed_brain_5pMeOH_P7_run19.RAW
Instrument Method	
Vial	
Injection Volume (µL)	0
Sample Weight	0
Sample Volume (µL)	0
ISTD Amount	0
Dil Factor	0

Chromatogram Parameters


Use Restricted Time	False
Time Limits	0.023 - 7.169 minutes
Scan Range	1 - 334
m/z Range	400 - 20000
Chromatogram Trace Type	TIC
Sensitivity	High
Rel. Intensity Threshold (%)	1

Chromatogram



Main Parameters (ReSpect™)	
Deconvolution Results Filter	
Output Mass Range	5000 - 100000
Deconvoluted Spectra Display Mode	Isotopic Profile (new)
Charge State Distribution	
Deconvolution Mass Tolerance	30 ppm
Choice of Peak Model	
Choice of Peak Model	Intact Protein
Resolution at 400 m/z	
Raw File Specific	6187
Generate XIC for Each Component	
Calculate XIC	False
Advanced Parameters (ReSpect™)	
Charge State Distribution	
Model Mass Range	10000 - 120000
Charge State Range	5 - 40
Minimum Adjacent Charges (low & high model mass)	3 - 3
Noise Parameters	
Rel. Abundance Threshold (%)	15
Deconvolution Quality	
Quality Score Threshold	0
Choice of Peak Model	
Target Mass	120000 Da
Peak Model Parameters	
Number of Peak Models	1
Left/Right Peak Shape	2:2
Peak Filter Parameters	
Peak Detection Minimum Significance Measure	1 Standard Deviations
Peak Detection Quality Measure	95%
Specialized Parameters	
Peak Model Width Factor	1
Intensity Threshold Scale	0.01
Deconvolution Parameters	
Noise Compensation	True
Charge Carrier	H
Negative Charge	False
Source Spectra Parameters	
Source Spectra Method	Sliding Windows
Sliding Windows Definition	
RT Range	2.000 - 4.500 minutes
Target Avg Spectrum Width	0.2 minutes
Target Avg Spectrum Offset Scan	1
Sliding Windows Merging Parameters	
Merge Tolerance	30 ppm
Max RT Gap	0.5 minutes
Min. Number of Detected Intervals	40

NativeLESA_method_glassslide_undosed_brain_5pMeOH_P7_run19



Sliding Windows ReSpect Masses Table

Row Number	Average Mass	Sum Intensity	Relative Abundance	Fractional Abundance	Score	Number of Charge States	Charge State Distribution	Number of Detected Intervals	Mass Std Dev	PPM Std Dev	Delta Mass	Scan Range	Start Time (min)	Stop Time (min)	Apex RT
1	20711.43	7111245.1 6	100.00	55.09	23.47	4	5 - 8	108	0.48	23.24	0.00	93 - 208	1.997	4.486	2.780
2	42718.68	2174219.2 9	30.57	16.84	17.62	4	9 - 12	90	2.27	53.17	22007.25	96 - 194	2.061	4.186	2.630
3	42654.86	1593266.1 7	22.40	12.34	16.77	3	9 - 11	91	2.43	57.01	21943.43	110 - 208	2.362	4.486	2.716
4	25204.98	1016270.9 2	14.29	7.87	17.29	3	6 - 8	62	1.73	68.49	4493.56	108 - 189	2.319	4.079	2.587
5	42782.08	1013706.9 7	14.25	7.85	18.77	4	9 - 12	40	2.91	68.07	22070.66	93 - 173	1.997	3.735	2.609

Supplemental File 4: Biopharma Finder search parameters and outputs for LESA mass spectrum shown in Fig 1C. (Cerebral cortex, 10 mM ammonium acetate, 0.125% C8E4).

BioPharma Finder Report

Created: 08/04/2019 12:25:49


Sample Information

Raw File Name	C:\Xcalibur\Data\ms2LESA\NativeLESA_method_glassslide__rat_brain_5pMeOH_P4_run8.RAW
Instrument Method	
Vial	
Injection Volume (µL)	0
Sample Weight	0
Sample Volume (µL)	0
ISTD Amount	0
Dil Factor	0

Chromatogram Parameters


Use Restricted Time	False
Time Limits	0.023 - 7.169 minutes
Scan Range	1 - 334
m/z Range	400 - 20000
Chromatogram Trace Type	TIC
Sensitivity	High
Rel. Intensity Threshold (%)	1

Chromatogram



Main Parameters (ReSpect™)	
Deconvolution Results Filter	
Output Mass Range	5000 - 100000
Deconvoluted Spectra Display Mode	Isotopic Profile (new)
Charge State Distribution	
Deconvolution Mass Tolerance	30 ppm
Choice of Peak Model	
Choice of Peak Model	Intact Protein
Resolution at 400 m/z	
Raw File Specific	6187
Generate XIC for Each Component	
Calculate XIC	False
Advanced Parameters (ReSpect™)	
Charge State Distribution	
Model Mass Range	10000 - 120000
Charge State Range	5 - 40
Minimum Adjacent Charges (low & high model mass)	3 - 3
Noise Parameters	
Rel. Abundance Threshold (%)	15
Deconvolution Quality	
Quality Score Threshold	0
Choice of Peak Model	
Target Mass	120000 Da
Peak Model Parameters	
Number of Peak Models	1
Left/Right Peak Shape	2:2
Peak Filter Parameters	
Peak Detection Minimum Significance Measure	1 Standard Deviations
Peak Detection Quality Measure	95%
Specialized Parameters	
Peak Model Width Factor	1
Intensity Threshold Scale	0.01
Deconvolution Parameters	
Noise Compensation	True
Charge Carrier	H
Negative Charge	False
Source Spectra Parameters	
Source Spectra Method	Sliding Windows
Sliding Windows Definition	
RT Range	0.430 - 3.300 minutes
Target Avg Spectrum Width	0.2 minutes
Target Avg Spectrum Offset Scan	1
Sliding Windows Merging Parameters	
Merge Tolerance	30 ppm
Max RT Gap	0.5 minutes
Min. Number of Detected Intervals	20

NativeLESA_method_glassslide_rat_brain_5pMeOH_P4_run8



Sliding Windows ReSpect Masses Table

Row Number	Average Mass	Sum Intensity	Relative Abundance	Fractional Abundance	Score	Number of Charge States	Charge State Distribution	Number of Detected Intervals	Mass Std Dev	PPM Std Dev	Delta Mass	Scan Range	Start Time (min)	Stop Time (min)	Apex RT
1	42655.42	14067523.01	100.00	36.74	30.11	9	8 - 16	118	2.21	51.85	0.00	27 - 152	0.581	3.285	1.600
2	42720.93	7994026.90	56.83	20.88	30.32	9	7 - 15	106	1.94	45.42	65.51	25 - 152	0.538	3.285	2.158
3	21144.61	4472423.83	31.79	11.68	20.62	4	5 - 8	70	1.00	47.18	-21510.81	20 - 114	0.430	2.469	0.527
4	27375.41	4450148.04	31.63	11.62	20.75	6	5 - 10	101	1.43	52.15	-15280.01	23 - 151	0.495	3.263	2.008
5	20711.88	1667804.98	11.86	4.36	21.42	4	5 - 8	35	1.09	52.44	-21943.53	24 - 112	0.516	2.426	0.785
6	20711.29	1260771.74	8.96	3.29	19.72	4	5 - 8	42	1.39	67.09	-21944.13	20 - 149	0.430	3.220	1.986
7	21121.87	862100.36	6.13	2.25	22.92	4	5 - 8	21	1.34	63.53	-21533.55	20 - 59	0.430	1.289	1.042
8	24798.65	828510.80	5.89	2.16	17.83	4	6 - 9	22	1.90	76.68	-17856.77	21 - 82	0.452	1.783	0.591
9	21162.61	779314.56	5.54	2.04	21.99	4	5 - 8	24	1.33	62.80	-21492.80	20 - 65	0.430	1.418	1.042
10	42960.13	727514.64	5.17	1.90	23.23	7	8 - 14	22	3.07	71.43	304.71	41 - 149	0.881	3.220	1.493
11	29735.45	648545.36	4.61	1.69	20.96	6	7 - 12	24	2.39	80.43	-12919.97	22 - 99	0.473	2.147	0.978
12	42784.83	534673.34	3.80	1.40	26.27	7	9 - 15	21	3.19	74.64	129.42	102 - 152	2.190	3.285	2.866

Supplemental File 5: Biopharma Finder search parameters and outputs for LESA mass spectrum shown in Fig 2A. (Kidney outer cortex, 10 mM ammonium acetate, 0.125% C8E4).

BioPharma Finder Report

Created: 08/04/2019 13:52:58


Sample Information

Raw File Name	C:\Xcalibur\Data\ms2LESA\NativeLESA_undosed_rat_kidney_run5.RAW
Instrument Method	
Vial	
Injection Volume (µL)	0
Sample Weight	0
Sample Volume (µL)	0
ISTD Amount	0
Dil Factor	0

Chromatogram Parameters


Use Restricted Time	False
Time Limits	0.023 - 7.169 minutes
Scan Range	1 - 334
m/z Range	2000 - 20000
Chromatogram Trace Type	TIC
Sensitivity	High
Rel. Intensity Threshold (%)	1

Chromatogram



Main Parameters (ReSpect™)	
Deconvolution Results Filter	
Output Mass Range	5000 - 100000
Deconvoluted Spectra Display Mode	Isotopic Profile (new)
Charge State Distribution	
Deconvolution Mass Tolerance	30 ppm
Choice of Peak Model	
Choice of Peak Model	Intact Protein
Resolution at 400 m/z	
Raw File Specific	6187
Generate XIC for Each Component	
Calculate XIC	False
Advanced Parameters (ReSpect™)	
Charge State Distribution	
Model Mass Range	10000 - 120000
Charge State Range	5 - 40
Minimum Adjacent Charges (low & high model mass)	3 - 3
Noise Parameters	
Rel. Abundance Threshold (%)	15
Deconvolution Quality	
Quality Score Threshold	0
Choice of Peak Model	
Target Mass	120000 Da
Peak Model Parameters	
Number of Peak Models	1
Left/Right Peak Shape	2:2
Peak Filter Parameters	
Peak Detection Minimum Significance Measure	1 Standard Deviations
Peak Detection Quality Measure	95%
Specialized Parameters	
Peak Model Width Factor	1
Intensity Threshold Scale	0.01
Deconvolution Parameters	
Noise Compensation	True
Charge Carrier	H
Negative Charge	False
Source Spectra Parameters	
Source Spectra Method	Sliding Windows
Sliding Windows Definition	
RT Range	0.440 - 3.000 minutes
Target Avg Spectrum Width	0.2 minutes
Target Avg Spectrum Offset Scan	1
Sliding Windows Merging Parameters	
Merge Tolerance	30 ppm
Max RT Gap	0.5 minutes
Min. Number of Detected Intervals	40

NativeLESA_undosed_rat_kidney_run5



Sliding Windows ReSpect Masses Table

Row Number	Average Mass	Sum Intensity	Relative Abundance	Fractional Abundance	Score	Number of Charge States	Charge State Distribution	Number of Detected Intervals	Mass Std Dev	PPM Std Dev	Delta Mass	Scan Range	Start Time (min)	Stop Time (min)	Apex RT
1	42641.66	29486568.19	100.00	61.60	25.46	10	6 - 15	114	1.58	37.01	0.00	22 - 143	0.449	2.995	0.815
2	17582.36	9383396.46	31.82	19.60	19.18	3	5 - 7	73	0.59	33.58	-25059.30	22 - 104	0.449	2.139	1.426
3	17511.65	9001058.95	30.53	18.80	17.51	3	5 - 7	59	0.40	22.94	-25130.01	22 - 89	0.449	1.833	1.447

Supplemental File 6: Biopharma Finder search parameters and outputs for LESA mass spectrum shown in Suppl Fig 3. (Kidney outer cortex, 10 mM ammonium acetate, 5% MeOH).

BioPharma Finder Report

Created: 08/04/2019 14:03:48


Sample Information

Raw File Name	C:\Xcalibur\Data\ms2LESA\NativeLESA_method_glassslide_undosed_kidney_5pMeOH_P4_run1_181107142027.RAW
Instrument Method	
Vial	
Injection Volume (µL)	0
Sample Weight	0
Sample Volume (µL)	0
ISTD Amount	0
Dil Factor	0

Chromatogram Parameters

Use Restricted Time	False
Time Limits	0.023 - 7.169 minutes
Scan Range	1 - 334
m/z Range	400 - 20000
Chromatogram Trace Type	TIC
Sensitivity	High
Rel. Intensity Threshold (%)	1

Chromatogram



Main Parameters (ReSpect™)

Deconvolution Results Filter	
Output Mass Range	5000 - 100000
Deconvoluted Spectra Display Mode	Isotopic Profile (new)
Charge State Distribution	
Deconvolution Mass Tolerance	30 ppm
Choice of Peak Model	
Choice of Peak Model	Intact Protein
Resolution at 400 m/z	
Raw File Specific	6187
Generate XIC for Each Component	
Calculate XIC	False


Advanced Parameters (ReSpect™)

Charge State Distribution	
Model Mass Range	10000 - 120000
Charge State Range	5 - 40
Minimum Adjacent Charges (low & high model mass)	3 - 3
Noise Parameters	
Rel. Abundance Threshold (%)	15
Deconvolution Quality	
Quality Score Threshold	0
Choice of Peak Model	
Target Mass	120000 Da
Peak Model Parameters	
Number of Peak Models	1
Left/Right Peak Shape	2:2
Peak Filter Parameters	
Peak Detection Minimum Significance Measure	1 Standard Deviations
Peak Detection Quality Measure	95%
Specialized Parameters	
Peak Model Width Factor	1
Intensity Threshold Scale	0.01
Deconvolution Parameters	
Noise Compensation	True
Charge Carrier	H
Negative Charge	False

Source Spectra Parameters

Source Spectra Method	Sliding Windows
Sliding Windows Definition	
RT Range	0.020 - 0.260 minutes
Target Avg Spectrum Width	0.05 minutes
Target Avg Spectrum Offset Scan	1
Sliding Windows Merging Parameters	
Merge Tolerance	30 ppm
Max RT Gap	0.5 minutes
Min. Number of Detected Intervals	3

NativeLESA_method_glassslide_undosed_kidney_5pMeOH_P4_run1_181107142027



Sliding Windows ReSpec Masses Table

Row Number	Average Mass	Sum Intensity	Relative Abundance	Fractional Abundance	Score	Number of Charge States	Charge State Distribution	Number of Detected Intervals	Mass Std Dev	PPM Std Dev	Delta Mass	Scan Range	Start Time (min)	Stop Time (min)	Apex RT
1	28428.52	1711541.75	100.00	70.77	23.86	4	5 - 8	9	0.60	20.96	0.00	1 - 10	0.023	0.237	0.130
2	18336.19	257555.72	15.05	10.65	16.77	3	6 - 8	8	1.70	92.50	-10092.33	1 - 10	0.023	0.237	0.173
3	51711.96	186325.26	10.89	7.70	26.05	5	13 - 17	6	2.03	39.30	23283.44	4 - 10	0.087	0.237	0.216
4	31889.50	174095.60	10.17	7.20	15.17	3	10 - 12	6	1.65	51.75	3460.98	1 - 8	0.023	0.194	0.087
5	42643.41	88801.55	5.19	3.67	21.36	4	12 - 15	3	2.52	59.05	14214.88	1 - 8	0.023	0.194	0.173

Supplemental File 7: Biopharma Finder search parameters and outputs for LESA mass spectrum shown in Fig 2C. (Kidney medulla, 10 mM ammonium acetate, 0.125% C8E4).

BioPharma Finder Report

Created: 08/04/2019 14:41:38


Sample Information

Raw File Name	C:\Xcalibur\Data\ms2LESA\NativeLESA_undosed_rat_kidney_run9.RAW
Instrument Method	
Vial	
Injection Volume (µL)	0
Sample Weight	0
Sample Volume (µL)	0
ISTD Amount	0
Dil Factor	0

Chromatogram Parameters

Use Restricted Time	False
Time Limits	0.023 - 7.169 minutes
Scan Range	1 - 334
m/z Range	2000 - 20000
Chromatogram Trace Type	TIC
Sensitivity	High
Rel. Intensity Threshold (%)	1

Chromatogram



Main Parameters (ReSpect™)

Deconvolution Results Filter	
Output Mass Range	5000 - 100000
Deconvoluted Spectra Display Mode	Isotopic Profile (new)
Charge State Distribution	
Deconvolution Mass Tolerance	25 ppm
Choice of Peak Model	
Choice of Peak Model	Intact Protein
Resolution at 400 m/z	
Raw File Specific	6187
Generate XIC for Each Component	
Calculate XIC	False


Advanced Parameters (ReSpect™)

Charge State Distribution	
Model Mass Range	10000 - 120000
Charge State Range	5 - 40
Minimum Adjacent Charges (low & high model mass)	3 - 3
Noise Parameters	
Rel. Abundance Threshold (%)	15
Deconvolution Quality	
Quality Score Threshold	0
Choice of Peak Model	
Target Mass	120000 Da
Peak Model Parameters	
Number of Peak Models	1
Left/Right Peak Shape	2:2
Peak Filter Parameters	
Peak Detection Minimum Significance Measure	1 Standard Deviations
Peak Detection Quality Measure	95%
Specialized Parameters	
Peak Model Width Factor	1
Intensity Threshold Scale	0.01
Deconvolution Parameters	
Noise Compensation	True
Charge Carrier	H
Negative Charge	False

Source Spectra Parameters

Source Spectra Method	Sliding Windows
Sliding Windows Definition	
RT Range	0.300 - 3.300 minutes
Target Avg Spectrum Width	0.2 minutes
Target Avg Spectrum Offset Scan	1
Sliding Windows Merging Parameters	
Merge Tolerance	30 ppm
Max RT Gap	0.5 minutes
Min. Number of Detected Intervals	40

NativeLESA_undosed_rat_kidney_run9



Sliding Windows ReSpect Masses Table

Row Number	Average Mass	Sum Intensity	Relative Abundance	Fractional Abundance	Score	Number of Charge States	Charge State Distribution	Number of Detected Intervals	Mass Std Dev	PPM Std Dev	Delta Mass	Scan Range	Start Time (min)	Stop Time (min)	Apex RT
1	20711.44	19849421.89	100.00	22.00	22.63	4	5 - 8	138	0.63	30.29	0.00	15 - 161	0.307	3.299	2.383
2	33857.16	12522848.07	63.09	13.88	23.91	6	7 - 12	137	1.55	45.76	13145.73	16 - 161	0.327	3.299	2.200
3	47036.59	5397511.39	27.19	5.98	24.33	6	9 - 14	107	3.48	73.91	26325.15	40 - 161	0.815	3.299	2.749
4	20734.29	5151119.87	25.95	5.71	18.78	4	5 - 8	99	0.43	20.74	22.86	23 - 146	0.469	2.994	2.037
5	31148.81	5146309.42	25.93	5.70	21.48	5	7 - 11	85	2.07	66.36	10437.37	16 - 161	0.327	3.299	1.854
6	36523.14	4335162.62	21.84	4.80	15.54	3	11 - 13	100	1.07	29.29	15811.70	35 - 160	0.714	3.279	1.039
7	29206.24	4050590.31	20.41	4.49	26.29	5	6 - 10	109	1.68	57.39	8494.81	25 - 161	0.510	3.299	2.403
8	15196.26	3813644.23	19.21	4.23	17.10	3	5 - 7	102	0.58	38.46	-5515.18	17 - 154	0.347	3.157	2.403
9	91177.05	3467631.17	17.47	3.84	22.89	4	29 - 32	93	4.21	46.19	70465.61	27 - 155	0.551	3.177	1.670
10	29326.39	3235284.87	16.30	3.59	23.00	5	6 - 10	93	1.86	63.57	8614.96	33 - 161	0.673	3.299	3.116
11	29735.99	3232881.38	16.29	3.58	15.26	6	6 - 12	55	1.69	56.70	9024.55	35 - 147	0.714	3.014	1.691
12	95852.40	3050618.05	15.37	3.38	15.62	4	24 - 27	65	5.02	52.41	75140.96	40 - 150	0.815	3.075	1.752
13	31146.56	2932689.72	14.77	3.25	25.64	6	6 - 11	43	2.38	76.49	10435.13	42 - 159	0.856	3.258	2.342
14	68482.42	2868304.85	14.45	3.18	18.49	3	17 - 19	78	2.36	34.43	47770.98	15 - 106	0.307	2.179	0.673
15	24798.72	2298792.11	11.58	2.55	18.71	4	6 - 9	77	1.04	42.02	4087.28	23 - 158	0.469	3.238	1.284
16	15952.35	2157903.03	10.87	2.39	20.72	3	5 - 7	74	0.46	28.75	-4759.09	58 - 158	1.182	3.238	2.403
17	30305.48	1965608.27	9.90	2.18	20.90	7	6 - 12	58	1.44	47.64	9594.05	43 - 150	0.877	3.075	2.485
18	41669.29	1640651.37	8.27	1.82	22.08	5	8 - 12	54	2.92	69.97	20957.86	18 - 147	0.368	3.014	1.956
19	69570.76	1610236.26	8.11	1.78	14.18	3	21 - 23	46	5.44	78.18	48859.32	66 - 146	1.345	2.994	2.729
20	62576.46	1514752.34	7.63	1.68	14.13	4	15 - 18	49	4.11	65.69	41865.02	15 - 160	0.307	3.279	1.548

Supplemental File 8: Biopharma Finder search parameters and outputs for LESA mass spectrum obtained from a separate medulla location. (Kidney medulla, 10 mM ammonium acetate, 0.125% C8E4).

BioPharma Finder Report

Created: 08/04/2019 14:48:25


Sample Information

Raw File Name	C:\Xcalibur\Data\ms2LESA\NativeLESA_undosed_rat_kidney_run3.RAW
Instrument Method	
Vial	
Injection Volume (µL)	0
Sample Weight	0
Sample Volume (µL)	0
ISTD Amount	0
Dil Factor	0

Chromatogram Parameters

Use Restricted Time	False
Time Limits	0.023 - 7.169 minutes
Scan Range	1 - 334
m/z Range	2000 - 20000
Chromatogram Trace Type	TIC
Sensitivity	High
Rel. Intensity Threshold (%)	1

Chromatogram



Main Parameters (ReSpect™)

Deconvolution Results Filter	
Output Mass Range	5000 - 100000
Deconvoluted Spectra Display Mode	Isotopic Profile (new)
Charge State Distribution	
Deconvolution Mass Tolerance	30 ppm
Choice of Peak Model	
Choice of Peak Model	Intact Protein
Resolution at 400 m/z	
Raw File Specific	6187
Generate XIC for Each Component	
Calculate XIC	False


Advanced Parameters (ReSpect™)

Charge State Distribution	
Model Mass Range	10000 - 120000
Charge State Range	5 - 40
Minimum Adjacent Charges (low & high model mass)	3 - 3
Noise Parameters	
Rel. Abundance Threshold (%)	15
Deconvolution Quality	
Quality Score Threshold	0
Choice of Peak Model	
Target Mass	120000 Da
Peak Model Parameters	
Number of Peak Models	1
Left/Right Peak Shape	2:2
Peak Filter Parameters	
Peak Detection Minimum Significance Measure	1 Standard Deviations
Peak Detection Quality Measure	95%
Specialized Parameters	
Peak Model Width Factor	1
Intensity Threshold Scale	0.01
Deconvolution Parameters	
Noise Compensation	True
Charge Carrier	H
Negative Charge	False

Source Spectra Parameters

Source Spectra Method	Sliding Windows
Sliding Windows Definition	
RT Range	0.500 - 3.000 minutes
Target Avg Spectrum Width	0.2 minutes
Target Avg Spectrum Offset Scan	1
Sliding Windows Merging Parameters	
Merge Tolerance	30 ppm
Max RT Gap	0.5 minutes
Min. Number of Detected Intervals	40

NativeLESA_undosed_rat_kidney_run3



Sliding Windows ReSpect Masses Table

Row Number	Average Mass	Sum Intensity	Relative Abundance	Fractional Abundance	Score	Number of Charge States	Charge State Distribution	Number of Detected Intervals	Mass Std Dev	PPM Std Dev	Delta Mass	Scan Range	Start Time (min)	Stop Time (min)	Apex RT
1	33857.56	12656058.89	100.00	22.05	30.47	6	7 - 12	113	1.29	38.12	0.00	25 - 146	0.510	2.994	0.612
2	20711.77	11932977.46	94.29	20.79	22.73	4	5 - 8	113	0.66	31.82	-13145.79	25 - 146	0.510	2.994	0.612
3	24799.20	4797751.65	37.91	8.36	20.59	4	6 - 9	113	0.92	36.92	-9058.36	25 - 146	0.510	2.994	0.612
4	14683.65	4760163.36	37.61	8.29	16.50	3	5 - 7	96	0.31	21.36	-19173.91	32 - 142	0.653	2.912	1.365
5	33716.86	4135550.76	32.68	7.20	16.57	3	8 - 10	113	1.36	40.36	-140.70	25 - 146	0.510	2.994	0.938
6	22528.56	2426025.40	19.17	4.23	18.05	4	5 - 8	84	1.75	77.73	-11329.00	31 - 142	0.632	2.912	0.917
7	27374.65	2389225.53	18.88	4.16	21.79	4	6 - 9	92	1.59	57.92	-6482.91	44 - 146	0.897	2.994	1.935
8	69709.47	2217647.63	17.52	3.86	14.49	3	21 - 23	82	2.88	41.25	35851.91	32 - 131	0.653	2.688	0.917
9	14686.72	2028902.20	16.03	3.53	21.67	3	5 - 7	61	0.21	14.12	-19170.84	64 - 141	1.304	2.892	1.691
10	19872.59	1871867.97	14.79	3.26	16.39	3	6 - 8	57	0.57	28.78	-13984.96	29 - 121	0.592	2.485	1.650
11	47035.70	1773721.29	14.01	3.09	27.78	6	9 - 14	56	3.31	70.43	13178.14	52 - 146	1.060	2.994	2.017
12	29205.97	1715512.18	13.55	2.99	25.04	5	6 - 10	87	1.72	58.72	-4651.59	25 - 146	0.510	2.994	1.365
13	29734.57	1227954.21	9.70	2.14	15.04	3	10 - 12	43	2.09	70.40	-4122.99	36 - 101	0.734	2.078	1.019
14	43250.03	1223301.38	9.67	2.13	19.46	3	12 - 14	49	2.09	48.40	9392.47	28 - 109	0.571	2.240	1.162
15	78165.98	1202540.78	9.50	2.10	13.95	3	21 - 23	41	7.38	94.41	44308.42	33 - 90	0.673	1.854	1.162
16	45659.72	1039819.48	8.22	1.81	16.80	3	10 - 12	41	3.76	82.28	11802.16	25 - 85	0.510	1.752	0.612