

**Randomized controlled trial of medium cut-off versus high-flux dialyzers on quality of life outcomes in maintenance hemodialysis patients**

Jeong-Hoon Lim<sup>1</sup>, Yeongwoo Park<sup>2</sup>, Ju-Min Yook<sup>1</sup>, Soon-Youn Choi<sup>1</sup>, Hee-Yeon Jung<sup>1</sup>, Ji-Young Choi<sup>1</sup>, Sun-Hee Park<sup>1</sup>, Chan-Duck Kim<sup>1</sup>, Yong-Lim Kim<sup>1</sup> and Jang-Hee Cho<sup>1</sup> \*

<sup>1</sup>Department of Internal Medicine, School of Medicine, Kyungpook National University, Kyungpook National University Hospital, Daegu, South Korea. <sup>2</sup>Department of Statistics, Kyungpook National University, Daegu, South Korea

Supplementary Information Listing:

Supplementary Table S1

\*Corresponding author:

Jang-Hee Cho, MD, PhD

Associate Professor

Division of Nephrology, Department of Internal Medicine

Kyungpook National University Hospital, 130 Dongdeok-ro, Jung-gu, Daegu, 41944, Korea

Tel: +82-53-200-5550

Fax: +82-53-426-2046

E-mail: [jh-cho@knu.ac.k](mailto:jh-cho@knu.ac.k)

**Supplementary Table S1. Characteristics of dialyzers**

Dialyzer	Membrane polymer	Membrane type	Inner diameter (µm)	Wall thickness (µm)	Effective surface area (m <sup>2</sup> )	Ultrafiltration coefficient (mL/h/mmHg)	Priming volume (mL)	Sterilization
Therenova 400	Polyarylethersulfone-polyvinylpyrrolidone	Medium cut-off	180	35	1.7	48	91	Steam
FX CorDiax 80	Polysulfone-polyvinylpyrrolidone	High-flux	185	35	1.8	64	95	Steam
FX CorDiax 60	Polysulfone-polyvinylpyrrolidone	High-flux	185	35	1.4	47	74	Steam