

Supporting Information

Evaluation of the novel antimicrobial BCP3 in a coating for endotracheal tubes

*Berkay Ozcelik^{1†}, Paul Pasic¹, Parveen Sangwan¹, Cheang Ly Be¹, Veronica Glattauer¹,
Helmut Thissen^{1*}, Ramiz A. Boulos^{2*}*

¹ Commonwealth Scientific and Industrial Research Organisation (CSIRO) Manufacturing,
Clayton, Victoria 3168, Australia.

² Boulos & Cooper Pharmaceuticals Pty. Ltd. 16/45 Delawney Street, Balcatta, Western
Australia 6021, Australia.

Supplementary Table S1: BCP3 MIC* values for MSSA, MRSA and *P. aeruginosa*.

Bacterial strain	MIC (µg/mL)
MRSA	15.6
MSSA	15.6
<i>P. aeruginosa</i>	> 1000

*lowest concentration at which greater than 90 % inhibition of bacterial growth was observed.