

On-line Table: Acquisition parameters for MR imaging sequences

Parameters	PD-FSE	T2-FSE
Plane	Sagittal	Sagittal
FOV (mm)	220	210
Acquisition matrix	384 × 384	448 × 448
Section thickness (mm)	3	3
No. of sections	12	12
Voxel size (mm)	0.8 × 0.6 × 3.0	0.7 × 0.5 × 3.0
TR (ms)	2500	3130
TE (ms)	31	96
Signal averages	2	4
Echo spacing (ms)	10.4	12
Turbo factor	8	20
Acquisition time (min:sec)	2:54	3:25