

On-line Table: Summary of series conducted on the WEB device

	Pierot et al, ²⁰ Lubicz et al ²⁹	Pierot et al ¹⁹	Liebig et al ¹⁸	Bheme et al ²³	Caroff et al ³⁰	Pierot et al ²⁶
	Oct 2010 to May 2012	Nov 2012 to Jan 2014	May 2011 to Dec 2013	Apr 2012 to Aug 2014	Jun 2013 to May 2014	Dec 2011 to Feb 2014
Centers (No.)	12	10	10	2	10	10
Patients (No.)	45	62	47	52	90	51
Aneurysms (No.)	45	63	52	55	98	51
Mean IA size						
Width (mm)	NA	NA	NA	7 ^a	8.9	8.2
Dome (mm)	NA	NA	NA	7 ^a	NA	NA
Neck (mm)	NA	NA	4.7	5 ^a	5.2	5.6
Depth/neck (mm)	NA	NA	NA	NA	1.6	NA
Aneurysm size <10/≥10 mm (%/%)	96/4	83/17	87/13	NA	NA	80/20
Neck size <4/≥4 mm (%/%)	0/100	10/90	6/94	20/80	NA	0/100
Main locations of IA						
MCA	26 (58)	32 (51)	18 (35)	19 (37)	38 (39)	29 (57)
AcomA	5 (11)	16 (25)	17 (33)	9 (18)	21 (21)	4 (8)
BA	11 (24)	9 (14)	8 (15)	10 (20)	19 (19)	12 (24)
ICA	1 (2)	6 (10)	2 (4)	4 (6)	15 (15)	6 (12)
Ruptured (No.) (%)	3 (7)	7 (11)	52 (100)	14 (25)	33 (34)	3 (6)
Failed implantation (No.)	0 (0)	1 (2)	0 (0)	4 (7)	5 (5)	3 (6)
WEB-DL/WEB-SL (%/%)	100/0	49/51	71/29	NA	0/100	100/0
Additional treatment (No.) (%)	5 (11)	7 (11)	8 (15)	NA	12 (13)	4 (8)
Complications (No.) (%)	5 (11)	16 (25)	16 (34)	6 (12)	13 (14)	11 (22)
Main complication (No.) (%)	TE, ^b 4 (9)	TE, 10 (16)	TE, 5 (11)	TE, 4 (8)	TE, 6 (7)	TE, 9 (18)
Postprocedural antiplatelet therapy, none/single/dual (%/%%/%)	NA	23/50/27	31/51/18	NA	33/NA/NA	16/63/22
Time at last clinical follow-up (mo)	NA	1	Discharge	Discharge	3.8 ^c	6
mRS >1 at last follow-up (%)	6.7	3.2	NA	5.9	2.2	6.2
Mortality at last follow-up (%)	2.2	0	17.0	5.9	1.1	4.2
Angiographic follow-up (mo)	27.4 ^a	NA	4	3.3	3.3	6
Adequate occlusion at last follow-up (%)	84.2	NA	80	66	65	85.4

Note:—IA indicates intracranial aneurysms; AcomA, anterior communicating artery; BA, basilar artery; NA, not applicable; TE, thromboembolic.

^a Median.

^b Thromboembolic.

^c Average follow-up time.