

Supplementary Table S7. Paired comparisons in sensitivity and specificity between grading of images by ophthalmic graders using ultra-wide field images and 7 field Early Treatment Diabetic Retinopathy Study fundus images (Main Analysis)

Graders ultrawide field	Graders 7-field ETDRS		Graders 7-field ETDRS	
	Positive	Negative	Positive	Negative
Positive	72	11	45	27
Negative	14	4	35	49
	Sensitivity (n=101)		Specificity (n=156)	
Ultrawide field (95% CI Wilson)	82% (74–88%)		54% (46–61%)	
7-field ETDRS (95% CI Wilson)	85% (76–90%)		49% (41–56%)	
Difference (95% CI)	-3% (-14–8%)		5% (-5–16%)	
McNemar's test p-value	P=0.55		P=0.31	

Note: ETDRS = Early treatment diabetic retinopathy study; Graders ultrawide field = graders assessment based on ultrawide field images; Graders 7-field ETDRS = graders assessment based on 7-field ETDRS images.