

Diagnostic Accuracy of HemotypeSC as a Point-of-Care Testing Device for Sickle Cell Disease: Findings from a Southwestern State in Nigeria and Implications for Patient Care in Resource-Poor Settings of sub-Saharan Africa

Supplement Table 1

Table 1a:

Interrater agreement

Number of subjects = 145
Ratings per subject = 3
Number of rating categories = 5

	Coefficient	Std. Err	t	P > t	95% confidence interval	
Percent Agreement	1.0000	0.0000	-	-	1.0000	1.0000
Brennan and Prediger	1.0000	0.0000	-	-	1.0000	1.0000
Cohen/Conger's Kappa	1.0000	0.0000	-	-	1.0000	1.0000
Scott/ Fleiss' Kappa	1.0000	0.0000	-	-	1.0000	1.0000
Gwet's AC	1.0000	0.0000	-	-	1.0000	1.0000
Krippendorff's Alpha	1.0000	0.0000	-	-	1.0000	1.0000
Confidence intervals are clipped at the upper limit						

Table 1b:

Interrater agreement

Number of subjects = 156
Ratings per subject: min = 3
 avg = 2.9295
 max = 3
Number of rating categories = 5

	Coefficient	Std. Err	t	P > t	[95% confidence interval]	
Percent Agreement	1.0000	0.0000	-	-	1.0000	1.0000
Brennan and Prediger	1.0000	0.0000	-	-	1.0000	1.0000
Cohen/Conger's Kappa	1.0000	0.0000	-	-	1.0000	1.0000
Scott/ Fleiss' Kappa	1.0000	0.0000	-	-	1.0000	1.0000
Gwet's AC	1.0000	0.0000	-	-	1.0000	1.0000
Krippendorff's Alpha	1.0000	0.0000	-	-	1.0000	1.0000

Confidence intervals are clipped at the upper limit