

Supporting Information

Rapid Antimicrobial Susceptibility Testing on Clinical Urine Samples by Video-based Object Scattering Intensity Detection

*Fenni Zhang^{1,6}, Jiapei Jiang^{1,2}, Michelle McBride¹, Xinyu Zhou^{1,2}, Yunze Yang¹, Manni Mo^{1,3}, Joseph Peterman¹, Thomas Gryś^{*4}, Shelley E. Haydel^{*1,5}, Nongjian Tao^{§1}, and Shaopeng Wang^{*1}*

1. Biodesign Center for Bioelectronics and Biosensors, Arizona State University, Tempe, AZ 85287, USA
2. School of Biological and Health Systems Engineering, Tempe, Arizona 85287, USA
3. School of Molecular Sciences, Arizona State University, Tempe, Arizona 85287, USA
4. Department of Laboratory Medicine and Pathology, Mayo Clinic, Phoenix, AZ 85054, USA
5. School of Life Sciences, Arizona State University, Tempe, Arizona 85287, USA
6. Biosensor National Special Laboratory, Key Laboratory for Biomedical Engineering of Education Ministry, Department of Biomedical Engineering, Zhejiang University, Hangzhou, 310027, PR China

*Corresponding authors:

Shaopeng Wang: shaopeng.wang@asu.edu

Shelley E. Haydel: shelley.haydel@asu.edu

Thomas E. Gryś: Grys.Thomas@mayo.edu

§Deceased in March 2020.

S1. Background removal processing flow chart




Figure S1. Background removal processing flow chart. Step 1: Raw LVSi images recorded at 10 frames per second (A) are averaged for every 4 frames (B) to reduce noise and data size. The size of the local stack average is set to avoid cell motion induced blur. Step 2: Subtract stack local minimum (C) from (B) to remove static and slow drifting background noises (D). Stack local minimum, the minimal intensity over a short time duration for each pixel, is calculated by minimum intensity projection of a small image stack. The stack size is set to be small to remove static and slow drift noises, while retaining bacteria that move beyond the original position in order to avoid loss of signal. Step 3: Subtract stack median of all frames (E) from (D) to remove drift noises (F). Stack median is calculated by median intensity projection of each pixel for the entire video stacks. Step 4: Subtract local spatial background (G) from (F) to get the background free LVSi image (H). Spatial local background of each pixel in the image is calculated by averaging over a large ball with radius of 10 pixels around the pixel to remove large spatial variations of the background intensities. The radius should be set to at least the size of the largest object that is not part of the background.

S2. ROC curve for infection threshold determination

To determine the infection threshold, the results were evaluated using the receiver operating characteristic (ROC) curve constructed using I_C/I_{C0} as a predictor. From the ROC curve for the first 20 samples, of which

10 were positive and 10 were negative from the clinical validation, we determined the infection threshold (T_1) of 1.1 with a sensitivity of 100% and a specificity of 100% at a 90 min testing time.




Figure S2. ROC curve reveals 100% sensitivity and 100% specificity at 90 min with threshold of 1.1.

S3. OSID-AST with *E. coli* cultures at different concentrations

To determine the dynamic range of OSID-AST method, we performed AST with *E. coli* cultures (concentrations ranging from 10^3 to 10^7 CFU/mL). The AST results and the corresponding raw intensity and the object intensity are plotted in Figure S3. At the low concentration of 10^3 CFU/mL, both raw intensity and OSID analysis do not detect an obvious increase within 90 min. With longer testing times (170 min), the OSID analysis detects intensity increases, while raw intensity alone does not show obvious increases at the maximum time tested, 210 min. At concentrations between 10^4 and 10^7 CFU/mL, OSID method works well with the total AST time decreasing with increasing cell concentrations. Thus, the detection range of object intensity method is between 10^4 and 10^7 CFU/mL, while the raw intensity detection only accurately detects growth with concentrations above 10^6 CFU/mL. In contrast, the single cell counting method needs lower concentrations, between 10^4 and 10^5 CFU/mL, to accurately enumerate increases in bacterial cells. Therefore, OSID-AST can accept a wider dynamic range of bacterial loads, which simplifies the sample preparation process while providing robust results.




Figure S3. OSID-AST with *E. coli* cultures of concentrations of 10^3 (A), 10^4 (B), 10^5 (C), 10^6 (D), 10^7 (E) CFU/mL.

S4. Calibration curve between bacterial CFU concentrations and AST time




Figure S4. Calibration between bacterial concentrations and AST time. The plating validation of initial *E. coli* concentrations at different dilutions (A). *E. coli* growth curves with different initial cell concentrations (10^3 - 10^7 CFU/mL) (B). The calibration curve between bacterial concentration and AST time (C). The threshold for AST time determination is set at 1.1 as in panel B.

S5. Comparison of object intensity detection and single cell counting for *E. coli* and *S. saprophyticus* cultures




Figure S5. Representative results of object intensity detection and single cell counting for pure *E. coli* (A) and *S. saprophyticus* (B) cultures. Single cell counting accurately detected growth and susceptibility with *E. coli*, while object intensity accurately detected growth and susceptibility with both species.

S6. Flow chart of the clinical sample preparation, testing and validation process




Figure S6. Workflow of sample preparation, OSID-AST test and plating-based validations for clinical urine samples.

S7. Clinical urine sample ID results

Table S1. OSID-AST detection (I_{c90}/I_{c0}) results of bacteriuria infection for 130 human patient samples compared with clinical results and on-site plating validation results

Sample #	Sample ID	Color	I _{CSB} /I _{CO}	OSID-AST Call	Clinical Lab Call*	Plating Validation Call*
1	ATU090319_01	Yellow	0.73776	Negative†	Positive	Positive
2	ATU090319_02	Yellow	5.38214	Positive	Positive	Positive
4	ATU090319_04	Brown	2.21285	Positive	Positive	Positive
6	ATU090319_06	Clear	4.48616	Positive	Positive	Positive
8	ATU090319_08	Clear	4.5611	Positive	Positive	Positive
10	ATU090319_10	Yellow	3.1863	Positive	Positive	Positive
11	ATU091019_01	Yellow	1.16772	Positive	Positive	Positive
13	ATU091019_03	Yellow	5.20085	Positive	Positive	Positive
15	ATU091019_05	Milky	0.8132	Negative†	Positive	Positive
17	ATU091019_07	Milky	2.04085	Positive	Positive	Positive
18	ATU091019_08	Cloudy	4.8725	Positive	Positive	Positive
21	ATU092419_01	Milky	6.48988	Positive	Positive	Positive
22	ATU092419_02	Milky	7.17449	Positive	Positive	Positive
26	ATU092419_06	Milky	2.53505	Positive	Positive	Positive
27	ATU092419_07	Yellow	0.80355	Negative†	Positive	Positive
28	ATU092419_08	Milky	0.63273	Negative†	Positive	Positive
31	ATU100119_01	Milky	4.073	Positive	Positive	Positive
33	ATU100119_03	Clear	3.6801	Positive	Positive	Positive
35	ATU100119_05	Clear	5.3097	Positive	Positive	Positive
36	ATU100119_06	Cloudy	6.70157	Positive	Positive	Positive
38	ATU100119_08	Cloudy	0.72527	Negative†	Positive	Positive
44	ATU100819_04	Milky	7.42181	Positive	Positive	Positive
45	ATU100819_05	Milky	7.68108	Positive	Positive	Positive
46	ATU100819_06	Milky	4.12703	Positive	Positive	Positive
49	ATU100819_09	Milky	5.10955	Positive	Positive	Positive
50	ATU100819_10	Milky	0.75666	Negative†	Positive	Positive
51	ATU101519_01	Yellow	0.89931	Negative†	Positive	Negative†
52	ATU101519_02	Milky	4.23241	Positive	Positive	Positive
53	ATU101519_03	Clear	5.32849	Positive	Positive	Positive
57	ATU101519_07	Milky	5.19027	Positive	Positive	Positive
58	ATU101519_08	Milky	6.19954	Positive	Positive	Positive
61	ATU102219_01	Yellow	5.12761	Positive	Positive	Positive
62	ATU102219_02	Clear	8.17162	Positive	Positive	Positive
63	ATU102219_03	Clear	2.50728	Positive	Positive	Positive
64	ATU102219_04	Clear	7.79974	Positive	Positive	Positive
69	ATU102219_09	Milky	5.60933	Positive	Positive	Positive
71	ATU102919_01	Milky	1.96968	Positive	Positive	Positive
72	ATU102919_02	Yellow	1.13846	Positive	Positive	Positive
73	ATU102919_03	Yellow	5.47704	Positive	Positive	Positive
79	ATU102919_09	Milky	6.25799	Positive	Positive	Positive
80	ATU102919_10	Milky	2.46932	Positive	Positive	Positive
83	ATU111219_03	Cloudy	2.51224	Positive	Positive	Positive
84	ATU111219_04	Yellow	6.27751	Positive	Positive	Positive
85	ATU111219_05	Milky	0.88083	Negative†	Positive	Positive
89	ATU111219_09	Cloudy	0.72358	Negative†	Positive	Positive
90	ATU111219_10	Yellow	4.27291	Positive	Positive	Positive
93	ATU120319_03	Milky	4.75435	Positive	Positive	Positive
96	ATU120319_06	Milky	4.74752	Positive	Positive	Positive
97	ATU120319_07	Clear	1.44209	Positive	Positive	Positive
99	ATU120319_09	Milky	1.22829	Positive	Positive	Positive
100	ATU120319_10	Yellow	2.89034	Positive	Positive	Positive
102	ATU121019_02	Clear	6.36563	Positive	Positive	Positive
104	ATU121019_04	Clear	0.87916	Negative†	Positive	Positive
106	ATU121019_06	Cloudy	3.58189	Positive	Positive	Positive
108	ATU121019_08	Cloudy	1.6952	Positive	Positive	Positive
110	ATU121019_10	Milky	2.72788	Positive	Positive	Positive
113	ATU010720_03	Milky	2.03288	Positive	Positive	Positive
114	ATU010720_04	Milky	3.19438	Positive	Positive	Positive
116	ATU010720_06	Milky	2.32701	Positive	Positive	Positive
117	ATU010720_07	Cloudy	1.97967	Positive	Positive	Positive
118	ATU010720_08	Cloudy	5.8613	Positive	Positive	Positive
124	ATU011420_04	Milky	5.88689	Positive	Positive	Positive
125	ATU011420_05	Milky	0.96676	Negative†	Positive	Negative†
128	ATU011420_08	Clear	2.38955	Positive	Positive	Positive
129	ATU011420_09	Cloudy	5.39929	Positive	Positive	Positive
130	ATU011420_10	Milky	1.24704	Positive	Positive	Positive

3	ATU090319_03	Clear	0.46777	Negative	Negative	Negative
5	ATU090319_05	Clear	0.6696	Negative	Negative	Negative
7	ATU090319_07	Clear	0.81648	Negative	Negative	Negative
9	ATU090319_09	Clear	0.76041	Negative	Negative	Negative
12	ATU091019_02	Yellow	0.86645	Negative	Negative	Negative
14	ATU091019_04	Yellow	0.93563	Negative	Negative	Negative
16	ATU091019_06	Clear	0.80403	Negative	Negative	Negative
19	ATU091019_09	Clear	0.90393	Negative	Negative	Negative
20	ATU091019_10	Clear	0.88178	Negative	Negative	Negative
23	ATU092419_03	Cloudy	0.77362	Negative	Negative	Negative
24	ATU092419_04	Clear	0.57816	Negative	Negative	Negative
25	ATU092419_05	Clear	0.78313	Negative	Negative	Negative
29	ATU092419_09	Clear	0.79639	Negative	Negative	Negative
30	ATU092419_10	Cloudy	0.78	Negative	Negative	Negative
32	ATU100119_02	Clear	0.98838	Negative	Negative	Negative
34	ATU100119_04	Clear	0.69371	Negative	Negative	Negative
37	ATU100119_07	Clear	0.90024	Negative	Negative	Negative
39	ATU100119_09	Milky	0.84636	Negative	Negative	Negative
40	ATU100119_10	Clear	0.86164	Negative	Negative	Negative
41	ATU100819_01	Yellow	0.56269	Negative	Negative	Negative
42	ATU100819_02	Clear	0.83746	Negative	Negative	Negative
43	ATU100819_03	Yellow	0.87212	Negative	Negative	Negative
47	ATU100819_07	Clear	0.87622	Negative	Negative	Negative
48	ATU100819_08	Clear	0.90836	Negative	Negative	Negative
54	ATU101519_04	Clear	0.90369	Negative	Negative	Negative
55	ATU101519_05	Milky	0.61539	Negative	Negative	Negative
56	ATU101519_06	Cloudy	0.1912	Negative	Negative	Negative
59	ATU101519_09	Cloudy	0.48572	Negative	Negative	Negative
60	ATU101519_10	Cloudy	0.49874	Negative	Negative	Negative
65	ATU102219_05	Yellow	0.41149	Negative	Negative	Negative
66	ATU102219_06	Yellow	0.6527	Negative	Negative	Negative
67	ATU102219_07	Clear	0.82163	Negative	Negative	Negative
68	ATU102219_08	Clear	0.67353	Negative	Negative	Negative
70	ATU102219_10	Clear	0.5757	Negative	Negative	Negative
74	ATU102919_04	Yellow	0.85456	Negative	Negative	Negative
75	ATU102919_05	Milky	0.33682	Negative	Negative	Negative
76	ATU102919_06	Yellow	0.83996	Negative	Negative	Negative
77	ATU102919_07	Clear	0.8318	Negative	Negative	Negative
78	ATU102919_08	Clear	0.91851	Negative	Negative	Negative
81	ATU111219_01	Yellow	0.85783	Negative	Negative	Negative
82	ATU111219_02	Yellow	0.9487	Negative	Negative	Negative
86	ATU111219_06	Clear	0.8654	Negative	Negative	Negative
87	ATU111219_07	Yellow	0.44171	Negative	Negative	Negative
88	ATU111219_08	Yellow	0.89734	Negative	Negative	Negative
91	ATU120319_01	Milky	0.91473	Negative	Negative	Negative
92	ATU120319_02	Yellow	0.68496	Negative	Negative	Negative
94	ATU120319_04	Yellow	0.82697	Negative	Negative	Negative
95	ATU120319_05	Yellow	0.90686	Negative	Negative	Negative
98	ATU120319_08	Clear	0.8935	Negative	Negative	Negative
101	ATU121019_01	Milky	0.97041	Negative	Negative	Negative
103	ATU121019_03	Yellow	0.90842	Negative	Negative	Negative
105	ATU121019_05	Clear	0.93901	Negative	Negative	Negative
107	ATU121019_07	Clear	0.97536	Negative	Negative	Negative
109	ATU121019_09	Yellow	0.88897	Negative	Negative	Negative
111	ATU010720_01	Clear	0.92131	Negative	Negative	Negative
112	ATU010720_02	Clear	0.94069	Negative	Negative	Negative
115	ATU010720_05	Clear	0.95538	Negative	Negative	Negative
119	ATU010720_09	Clear	0.90791	Negative	Negative	Negative
120	ATU010720_10	Yellow	0.9299	Negative	Negative	Negative
121	ATU011420_01	Yellow	0.91915	Negative	Negative	Negative
122	ATU011420_02	Clear	0.92376	Negative	Negative	Negative
123	ATU011420_03	Cloudy	0.76016	Negative	Negative	Negative
126	ATU011420_06	Clear	0.95965	Negative	Negative	Negative
127	ATU011420_07	Yellow	0.94512	Negative	Negative	Negative

* Reference method (Traditional plating followed by BD Phoenix™ automated AST) results generated by the Mayo Clinic microbiology lab.

On-site validation results generated by overnight plating upon sample receiving.

† Disagreement between OSID-AST and reference method results.

S8. Initial sample validation results

On-site initial bacterial load validation is performed with sample plating and colony counting. Upon urine sample reception, samples were subjected to serial dilutions and plated on LB agar for colony enumeration. This plating validation provides initial bacterial concentration references and reveals any viability changes during sample storage and transportation. While 66 of 130 clinical samples were confirmed to have greater than 10^3 CFU per mL, two of these contained concentrations below the clinical threshold of 10^4 CFU/mL, and six had bacterial concentrations that were 10-100 times less than those initially determined by Mayo Clinic (Figure S7), before storage and transport. Therefore, we anticipate greater accuracy when rapid AST is performed in POC settings and this loss in bacterial viability is avoided.




Figure S7. Initial plating validation of all 66 clinically determined positive sample. The blue bar shows the initial bacterial concentration of 11 false negative samples. The red dashed lines indicate the clinical infection threshold (10^4 - 10^5 CFU/mL). The stars indicate the urine samples with initial plating concentrations less than those determined by Mayo Clinic (prior to storage and transport).

S9. Initial and parallel plating validation result of 11 false negative samples

Parallel plating validation was performed along with LVSi detection to test the samples post-preparation. Initial plating CFU/mL determinations, calculated CFU/mL based on sample dilution, and parallel plating of sample post-preparation of 11 false negatives samples are presented here. The parallel plating validation results show low counts of bacterial cells (below 1000 cells/mL), after all sample handling, including prewarming, filtration and dilution. Dilutions of these samples ranged from 10 to 1000 times such that both single cell counting and object scattering intensity could be performed. Since the OSID-AST method functions at a higher particle concentration range (10^4 to 10^7 CFU/mL), we estimate that 7 out of 11 of these false negative results could be avoided with an optimized dilution scheme and quicker handling process at the POC settings.




Figure S8. The comparison of initial plating, calculated CFU/mL by dilution, and the parallel plating results of 11 false negatives samples. The plating results are the mean value of three replicates. The limit of detection (LOD, blue hatched line) for initial plating (blue bar) is 100 CFU/mL. The limit of detection (LOD, green hatched line) for parallel plating (green bar) is 200 CFU/mL. When parallel plating was performed but colonies were not detected in all three replicates (sample #1, #51, #105), the samples are marked as 'Below LOD'.

S10. Examples of false negative samples by single cell counting, positive by OSID

Each sample was analyzed with both the single cell counting method and the OSID method. Two examples are presented here. The single cell counting detection, which needs extra manual cell detection and tracking processing, showed 17 false negative samples out of 130 tested samples, a 6 sample increase compared to the object intensity detection method. The OSID method measures both cell growth and elongation, while cell counting alone tracks the sum of distinct individual particles and thus detects an increase only after elongated cells divide and daughter cells separate.




Figure S9. Two examples of false negative results as determined by single cell counting, which were correctly identified as positive for infection by object intensity detection. Example of an infection positive sample that is resistant to ciprofloxacin (2 $\mu\text{g}/\text{mL}$) (A). Example of an infection positive sample that is susceptible to ciprofloxacin (2 $\mu\text{g}/\text{mL}$) (B).

S11. Clinical urine sample AST results

Table 2. OSID-AST of 55 bacteriuria-positive human urine samples compared to clinical and onsite plating validation results

Sample #	Sample ID	Comments	$\Delta I_{ABX}/\Delta I_c$	OSID-AST Call	Clinical Lab Call*	Plating Validation Call#
1	ATU090319_01	2 Isolates: >100,000 cfu/mL KLEBSIELLA PNEUMONIAE ; 10,000-100,000 cfu/mL ENTEROCOCCUS FAECALIS	N/A	Negative†	S	S
2	ATU090319_02	>100,000 cfu/mL ESCHERICHIA COLI	0.94068	R	R	R
4	ATU090319_04	>100,000 cfu/mL ESCHERICHIA COLI	0.03015	S	S	NT
6	ATU090319_06	>100,000 cfu/mL ESCHERICHIA COLI	1.26833	R	R	R
8	ATU090319_08	>100,000 cfu/mL ESCHERICHIA COLI	-0.05981	S	S	NT
10	ATU090319_10	>100,000 cfu/mL KLEBSIELLA AEROGENES	0.24834	S	S	S
11	ATU091019_01	>100,000 cfu/mL ESCHERICHIA COLI	-0.35376	S	S	NT
13	ATU091019_03	>100,000 cfu/mL ESCHERICHIA COLI	0.01426	S	S	S
15	ATU091019_05	>100,000 cfu/mL ESCHERICHIA COLI	N/A	Negative†	S	S
17	ATU091019_07	>100,000 cfu/mL GRAM NEGATIVE BACILLUS	1.76005	R	R	R
18	ATU091019_08	>100,000 cfu/mL GRAM NEGATIVE BACILLUS	0.56815	R	R	R
21	ATU092419_01	>100,000 cfu/mL ESCHERICHIA COLI	0.01376	S	S	NT
22	ATU092419_02	>100,000 cfu/mL ESCHERICHIA COLI	0.1729	S	S	NT
26	ATU092419_06	>100,000 cfu/mL ENTEROBACTER CLOACAE	0.01948	S	S	S
27	ATU092419_07	>100,000 cfu/mL GRAM NEGATIVE BACILLUS	N/A	Negative†	S	S
28	ATU092419_08	10,000-100,000 cfu/mL GRAM NEGATIVE BACILLUS	N/A	Negative†	R	R
31	ATU100119_01	>100,000 cfu/mL ESCHERICHIA COLI	2.30E-04	S	S	NT
33	ATU100119_03	>100,000 cfu/mL ESCHERICHIA COLI	-0.09991	S	S	NT
35	ATU100119_05	>100,000 cfu/mL ESCHERICHIA COLI	0.95614	R	R	NT
36	ATU100119_06	>100,000 cfu/mL ESCHERICHIA COLI	0.04334	S	S	S
38	ATU100119_08	>100,000 cfu/mL GRAM NEGATIVE BACILLUS	N/A	Negative†	S	S
44	ATU100819_04	>100,000 cfu/mL ESCHERICHIA COLI	0.98483	R	R	NT
45	ATU100819_05	>100,000 cfu/mL ESCHERICHIA COLI	0.0119	S	S	NT
46	ATU100819_06	>100,000 cfu/mL ESCHERICHIA COLI	1.19128	R	R	R
49	ATU100819_09	>100,000 cfu/mL ESCHERICHIA COLI	0.02571	S	S	S
50	ATU100819_10	10,000-100,000 cfu/mL ESCHERICHIA COLI	N/A	Negative†	S	S
51	ATU101519_01	>100,000 cfu/mL ESCHERICHIA COLI	N/A	Negative†	S	NT
52	ATU101519_02	>100,000 cfu/mL ESCHERICHIA COLI	-0.00858	S	S	NT
53	ATU101519_03	>100,000 cfu/mL KLEBSIELLA PNEUMONIAE	0.20227	S	S	NT
57	ATU101519_07	>100,000 cfu/mL ESCHERICHIA COLI	0.14571	S	S	S
58	ATU101519_08	>100,000 cfu/mL GRAM NEGATIVE BACILLUS	0.30196	S	S	S
61	ATU102219_01	>100,000 cfu/mL ESCHERICHIA COLI	-0.07596	S	S	NT
62	ATU102219_02	>100,000 cfu/mL ESCHERICHIA COLI	0.85396	R	R	NT
63	ATU102219_03	>100,000 cfu/mL ESCHERICHIA COLI	0.8574	R	R	NT
64	ATU102219_04	>100,000 cfu/mL KLEBSIELLA PNEUMONIAE	0.1025	S	S	NT
69	ATU102219_09	>100,000 cfu/mL KLEBSIELLA sp	0.4051	S	S	S
71	ATU102919_01	>100,000 cfu/mL ESCHERICHIA COLI	-0.03695	S	S	NT
72	ATU102919_02	>100,000 cfu/mL ESCHERICHIA COLI	-1.14431	S	S	S
73	ATU102919_03	>100,000 cfu/mL ESCHERICHIA COLI	0.07311	S	S	S
79	ATU102919_09	>100,000 cfu/mL ESCHERICHIA COLI	0.09926	S	S	S
80	ATU102919_10	>100,000 cfu/mL CITROBACTER FREUNDII COMPLEX	1.14637	R ^a	S	S
83	ATU111219_03	>100,000 cfu/mL ESCHERICHIA COLI	-0.06239	S	S	NT
84	ATU111219_04	>100,000 cfu/mL KLEBSIELLA PNEUMONIAE	0.13194	S	S	NT
85	ATU111219_05	10,000-100,000 cfu/mL KLEBSIELLA PNEUMONIAE	N/A	Negative†	R	NT
89	ATU111219_09	>100,000 cfu/mL PSEUDOMONAS AERUGINOSA	N/A	Negative†	N/A	S
90	ATU111219_10	>100,000 cfu/mL ENTEROBACTER CLOACAE	0.2266	S	N/A	S
93	ATU120319_03	>100,000 cfu/mL ESCHERICHIA COLI	0.22734	S	S	S
96	ATU120319_06	>100,000 cfu/mL ESCHERICHIA COLI	1.12652	R	R	NT
97	ATU120319_07	10,000-100,000 cfu/mL ESCHERICHIA COLI	1.45551	R	R	NT
99	ATU120319_09	>100,000 cfu/mL ESCHERICHIA COLI	-0.07618	S	S	NT
100	ATU120319_10	>100,000 cfu/mL KLEBSIELLA PNEUMONIAE	0.10105	S	S	NT
102	ATU121019_02	>100,000 cfu/mL ESCHERICHIA COLI	0.40565	S	S	S
104	ATU121019_04	>100,000 cfu/mL ESCHERICHIA COLI	N/A	Negative†	S	S
106	ATU121019_06	>100,000 cfu/mL ESCHERICHIA COLI	0.87237	R	R	NT
108	ATU121019_08	>100,000 cfu/mL ESCHERICHIA COLI	0.27158	S	S	NT
110	ATU121019_10	>100,000 cfu/mL KLEBSIELLA PNEUMONIAE	0.4057	S	S	NT
113	ATU010720_03	>100,000 cfu/mL ESCHERICHIA COLI	0.03975	S	S	NT
114	ATU010720_04	>100,000 cfu/mL ESCHERICHIA COLI	-0.00101	S	S	NT
116	ATU010720_06	>100,000 cfu/mL ESCHERICHIA COLI	0.01551	S	S	S
117	ATU010720_07	>100,000 cfu/mL ESCHERICHIA COLI	0.95502	R	R	R
118	ATU010720_08	>100,000 cfu/mL KLEBSIELLA PNEUMONIAE	0.16467	S	S	S
124	ATU011420_04	>100,000 cfu/mL ESCHERICHIA COLI	1.03449	R	R	NT
125	ATU011420_05	10,000-100,000 cfu/mL ESCHERICHIA COLI	N/A	Negative†	S	NT
128	ATU011420_08	10,000-100,000 cfu/mL ESCHERICHIA COLI	-0.16487	S	S	S
129	ATU011420_09	>100,000 cfu/mL ESCHERICHIA COLI	1.29926	R	R	R
130	ATU011420_10	>100,000 cfu/mL ESCHERICHIA COLI	0.91598	R	R	R

* Reference method (Traditional plating followed by BD Phoenix™ automated AST) results generated by the Mayo Clinic microbiology lab.

On-site validation results generated by overnight plating upon sample receiving.

† False negative results by OSID-AST.


^a Mis-categorization occurred if reference method indicated susceptible and OSID-AST indicated resistant.

S12. Clinical Sample # 80

For AST of 55 positive samples, one susceptible sample was mis-categorized as resistant by OSID-AST (Sample # 80, Supporting Information S11, Table S2), which contains >100,000 CFU/mL *Citrobacter freundii* Complex. To elucidate this inconsistency, a validation experiment was performed. Since the original clinical sample was no longer available, we used cryogenically-preserved clinical sample # 80 for this validation experiment. Briefly, frozen cells were inoculated onto LB agar and incubated overnight (~ 15 h) at 37 °C. Then, bacteria were spiked into pooled urine (BioIVT Human Urine Pooled Gender) with a concentration of ~ 10⁹ CFU/mL and stored at 4 °C overnight to simulate clinical sample conditions. For LVS_i imaging, the spiked urine sample were processed as the flowchart in Figure S6 for OSID-AST detection, and high magnification microscopy (80×) images were recorded during AST detection for morphology validation.

LVS_i video and high magnification images (80× microscopic imaging) of the cultured isolates from the sample were shown in Figure S10, and obvious elongation of the bacteria under 2 µg/mL ciprofloxacin were detected, which induced object intensity increase in OSID-AST system similar to the control (antibiotic free) sample within the 90 min measurement time (Figure S10 A & B). With longer measurement times, the normalized growth curve with 2 µg/mL ciprofloxacin (Figure S10 C & D) plateaued around 2 h after the intensity roughly doubled, which means the cells stopped growing after reaching maximum length, as the antibiotics prevented DNA replication and cell division. In contrast, the cells without antibiotics grew exponentially. Therefore, for this elongation case, longer detection time is needed for accurate OSID-AST detection (Figure S10 C & D).

Data from Clinical Sample # 80



Data from isolates of Clinical Sample # 80




Figure S10. Background removed LVS images at different time points and object intensity curves for clinical sample # 80 (A, B) and cryogenically-preserved clinical sample #80 (C, D) without (A, C) and with antibiotics (B, D). The insets in C and D are the corresponding high magnification (80×) cell image for morphology validation.

S13. MIC determination for OSID-AST

The minimum inhibitory concentration (MIC) of ciprofloxacin and nitrofurantoin were determined for OSID-AST with 90 min exposure of different concentrations of the antibiotic. The normalized object

intensity (I_{90}/I_0) of *E. coli* cells with 0 $\mu\text{g/mL}$, 0.5 $\mu\text{g/mL}$, 1 $\mu\text{g/mL}$, 2 $\mu\text{g/mL}$, 4 $\mu\text{g/mL}$, and 8 $\mu\text{g/mL}$ ciprofloxacin were calculated for the inhibition curve plot (Supporting Information Figure S11A), while the nitrofurantoin concentrations tested were 0 $\mu\text{g/mL}$, 4 $\mu\text{g/mL}$, 8 $\mu\text{g/mL}$, 16 $\mu\text{g/mL}$, 32 $\mu\text{g/mL}$, 64 $\mu\text{g/mL}$ and 128 $\mu\text{g/mL}$ (Supporting Information Figure S11B). The MIC value is defined as no growth in object intensity where normalized object intensity equals 1 (dashed lines in Figure S11). The MIC value is determined to be 1 $\mu\text{g/mL}$ for ciprofloxacin and 4 $\mu\text{g/mL}$ for nitrofurantoin.




Figure S11. The inhibition curves of OSID-AST after 90 min exposure of ciprofloxacin (A) and nitrofurantoin (B). The MIC value is determined by the minimum concentration that no growth induced object intensity increase (dashed lines) detected by OSID-AST.