

Supplementary Table 1. Comparison of the Lesion-Based Diagnostic Performances of Two DLLDs in the Detection of Colorectal Liver Metastasis

	3-Class DLLD (95% CI)	1-Class DLLD (95% CI)	<i>P</i>
AUAFROC	0.631 (0.520, 0.737)	0.432 (0.346, 0.524)	< 0.001*
Sensitivity (%)	81.82 [81/99] (72.68, 88.39)	85.86 [85/99] (78.48, 91.00)	0.40
FPP	1.330 [113/85] (1.052, 1.681)	3.365 [286/85] (2.748, 4.120)	< 0.001*

*Statistically significant. AUAFROC = area under the alternative free-response receiver operating characteristic curve, CI = confidence interval, DLLD = deep learning-based lesion detection algorithm, FPP = false positives per patient