

Online Resource 9

Table S3 Correlation of inclination angle and distance-to-monitor errors with vertex and centroid position and area errors

	Inclination angle	Distance-to-monitor
Vertex position	0.272*	0.441*
Centroid position	0.348*	0.473*
Area	0.214*	0.612*

^aAll marker positions (1-9) and 8x8 and 12x12 cm markers included ($n = 972$)

*All correlations were significant at the 0.01 level (2-tailed)

Title: Effect of marker position and size on the registration accuracy of HoloLens in a non-clinical setting with implications for high-precision surgical tasks

Journal: International Journal of Computer Assisted Radiology and Surgery

Authors: Laura Pérez-Pachón¹, Parivrudh Sharma¹, Helena Brech¹, Jenny Gregory¹, Terry Lowe^{1,3}, Matthieu Poyade², Flora Gröning¹

¹ School of Medicine, Medical Sciences and Nutrition, University of Aberdeen, Aberdeen, United Kingdom

² School of Simulation and Visualisation, Glasgow School of Art, Glasgow, United Kingdom

³ Head and Neck Oncology Unit, Aberdeen Royal Infirmary (NHS Grampian), Aberdeen, United Kingdom

Corresponding author: laura.perezpachon@gmail.com (LP)