

Supplemental Table 2. Inter-assay precision

ANALYTE COPIES PER MILLION CELLS	AVERAGE MEASURED HIV- LTRGAG COPY PER MILLION CELLS (RANGE)	STANDARD DEVIATION	COEFFICIENT OF VARIATION
1,000	429.49 (266.52-585.02)	104.00	24.22%
100	99.37 (67.11-135.36)	28.32	28.50%
25	22.53 (14.20-36.38)	7.80	34.63%
10	12.91 (4.37-20.34)	6.89	53.39%
7.5	5.99 (2.48-11.00)	2.80	46.70%
5	6.46 (1.61-9.57)	3.02	46.79%
2.5	3.10 (0-6.69)	2.62	84.55%
1.25	0.52 (0-1.57)	0.81	154.92%
0	0 (0-0)	0	N/A