

**Supplemental Table 3: Intra-assay Precision**

<b>ANALYTE COPIES PER MILLION CELLS</b>	<b>STANDARD DEVIATION, DONOR 1</b>	<b>COEFFICIENT OF VARIATION DONOR 1</b>	<b>STANDARD DEVIATION DONOR 2</b>	<b>COEFFICIENT OF VARIATION DONOR 2</b>
<b>1,000</b>	89.56	18.26%	88.72	24.08%
<b>100</b>	26.19	28.01%	34.85	33.09%
<b>25</b>	8.08	29.86%	5.01	27.83%
<b>10</b>	5.40	69.39%	3.25	18.03%
<b>7.5</b>	2.00	42.71%	3.23	44.26%
<b>5</b>	3.30	64.74%	2.53	32.31%
<b>2.5</b>	2.25	45.21%	1.25	101.95%
<b>1.25</b>	0.91	173.21%	0.90	173.21%
<b>0</b>	0	N/A	0	N/A