

Supplementary Table SIV Analytical assays for the hormones measured in serum.

Hormone	Assay	Sensitivity (limit of detection)	Precision (coefficient of variation, %)
AMH	ECLIA	0.07 pmol/l	≤4.4 ^a
Estradiol	CLIA	11.8 pg/ml (43.6 pmol/l)	≤13.3 ^b
Inhibin B	ELISA	2.91 pg/ml	≤6.6 ^a
Inhibin A	CLIA	<1 pg/ml	≤6.05 ^b
Progesterone	CLIA	0.1 ng/ml (0.32 nmol/l)	≤14.7 ^b

AMH, anti-Müllerian hormone; CLIA, chemiluminescence immunoassay; ECLIA, electrochemiluminescence immunoassay.

^aInter-assay precision (intermediate precision).

^bTotal precision.