

Supplemental Online Content

Young-Wolff KC, Ray GT, Alexeeff SE, et al. Rates of prenatal cannabis use among pregnant women before and during the COVID-19 pandemic. *JAMA*. doi:10.1001/jama.2021.16328

eAppendix. Laboratory Methods

This supplemental material has been provided by the authors to give readers additional information about their work.

Laboratory Methods

Toxicology tests were performed on a Beckman Coulter AU680 chemistry analyzer using the EMIT II Plus Cannabinoid Assay (Siemens) with a cutoff of 45 ng/mL. Confirmatory testing for the presence of the cannabis metabolite, 11-nor-9-carboxy- Δ 9-tetrahydrocannabinol, was performed by liquid chromatography-tandem mass spectrometry (LC-MS/MS) for all positive immunoassay results. The confirmation test methodology was an LC-MS/MS on a triple quadrupole system with a cutoff for positivity of 15 ng/mL.