

Group	Strain	Source	PCR result
Virus	Dengue virus type 2	ATCC VR-1584	Negative
	Dengue virus type 4	ATCC VR-1490	Negative
	Herpes virus type 2	ATCC-VR-734	Negative
	Echovirus 30	KBPV-VR-25	Negative
	Human Adenovirus type 40	ATCC VR-931	Negative
	Human Parvovirus-B19	NIBSC-12/208	Negative
	BK virus	ATCC VR-837	Negative
	Coxsackie virus B3	ATCC VR-30	Negative
	Coxsackie virus B5	ATCC VR-185	Negative
Bacteria	<i>Enterococcus faecium</i>	KCTC-13225	Negative
	<i>Enterococcus faecalis</i>	KCTC-5290	Negative
	<i>Staphylococcus Saprophyticus</i>	KCTC-3345	Negative
	<i>Streptococcus agalactiae</i>	ATCC-12928	Negative
	<i>Klebsiella quasipneumoniae sub</i>	ATCC-700603	Negative
	<i>Lactobacillus jensenii</i>	ATCC-25258	Negative
	<i>Lactobacillus gasseri</i>	ATCC-33323	Negative
	<i>Lactobacillus crispatus</i>	ATCC-33820	Negative
	<i>Proteus mirabilis</i>	ATCC-29906	Negative
	<i>Haemophilus ducreyi</i>	ATCC-33940	Negative