



Supplement File 1.

 COLLEGE of AMERICAN PATHOLOGISTS <small>325 Waukegan Road, Northfield, Illinois 60093-2750 800-323-4040 • cap.org</small>		CAP Number: 1650514-01 Kit# 1 Institution: Cleveland Clinic Attention: Sajal Gupta MD,TS(ABB) City / State: Cleveland OH 44106-3018			Kit ID: 34613141 Kit Mailed: 4/19/2021 Original Evaluation: 5/28/2021					
EVALUATION ORIGINAL		SEM-A 2021 Semen Analysis								
Test Unit of Measure Peer Group	Evaluation and Comparative Method Statistics							Plot of the Relative Distance of Your Results from Target as Percentages of allowed Deviation		
	Specimen	Your Result	Mean	S.D.	No. of Labs	S.D.I	Limits of Acceptability Lower Upper	Your Grade	Survey	-100-----Mean-----+100
Sperm viability	SEM-03	52	30.0	23.9	359	+0.9	0 102	Acceptable		
% Sperm Viability (% Viable)	SEM-04	15	26.1	25.4	359	-0.4	0 103	Acceptable		
The College of American Pathologists recommends that the result of this interlaboratory comparison not be used as a sole criterion for judging the performance of any individual clinical laboratory.										