

# Standardization of molecular monitoring of CML: results and recommendations from the European Treatment and Outcome Study

## Supplementary information

### Supplementary Methods

#### Protocol for preparation of high- and low-level internal quality control cell line samples

1. Grow enough HL60 and K562 cells for your requirements. We estimate that  $5 \times 10^5$  HL60 cells per vial give approximately 40,000 *ABL1* copies per assay. A 1:50 dilution of K562 cells will generate a high-level control standard of approximately 10% BCR::*ABL1*<sup>IS</sup>
2. Spin HL60 cultures at 1500rpm for 15 mins and resuspend pellets in 1 x PBS
3. Count the HL60 cells and resuspend to  $1.5 \times 10^6$  cells / ml in 1 X PBS
4. Count the K562 cells and spin down required number of cells for a 1:50 dilution.
5. The K562 cells should be resuspended in HL60 cells ( $1.5 \times 10^6$  cells / ml in 1 X PBS) diluting the K562 cells 1:50 to generate the high-level control standard e.g. Spin down  $1.5 \times 10^7$  K562 cells and resuspend the pellet in 500ml HL60 ( $7.5 \times 10^8$  HL60)
6. The high-level standard can then be diluted 1:100 to generate a low-level control standard of approximately 0.1% BCR::*ABL1*<sup>IS</sup>. Perform the dilution using the HL60 cells ( $1.5 \times 10^6$  cells / ml in 1 X PBS) as the diluent.
7. For preparation of RLT lysates spin batches of 20ml of culture in 50ml Falcon tubes and resuspend each pellet in 36 ml RLT. The solution should be lysed by shearing with a 20ml syringe and wide gauge needle. This can be aliquoted into 60 vials of 600 $\mu$ l ( $5 \times 10^5$  cells per vial)
8. For preparation of Maxwell (Promega) buffer lysates spin batches of 20ml of culture in 50ml Falcon tubes and resuspend each pellet in of 12ml Maxwell buffer. The solution should be lysed by shearing with a 20ml syringe and wide gauge needle. This can be aliquoted into 60 vials of 200 $\mu$ l ( $5 \times 10^5$  cells per vial)
9. For preparation of Trizol lysates spin batches of 15ml of culture in 50ml Falcon tubes and resuspend each pellet in 45ml Trizol. The solution should be lysed by shearing with a 20ml syringe and wide gauge needle. This can be aliquoted into 45 vials of 1ml ( $5 \times 10^5$  cells per vial).

**Supplementary Table 1:** Details of cell line lysates distributed, control genes analysed, plasmid type used for standard curves, PCR protocol and PCR machines used for all data sets for each year.

	2016	2017	2019	2020	2021
<b>Lysate</b>					
RLT	24	23	25	24	29
Trizol	29	28	26	21	19
Maxwell	0	0	7	11	10
<b>Control gene</b>					
ABL1	41	45	43	42	42
GUSB	11	13	14	14	15
GUSB and BCR	1	1	1	0	1
<b>Plasmid</b>					
ERM-AD623	19	21	29	27	28
Ipsogen	27	24	25	23	24
pME-2	4	4	1	1	1
Other / None	2	2	3	5	5
<b>PCR protocol</b>					
EAC	37	35	42	40	38
Ipsogen BCR-ABL1 Mbc Kit	7	7	6	6	7
Emig	3	3	2	2	2
Other / In house/ Not specified	6	6	8	8	11

	2016	2017	2019	2020	2021
<b>PCR machine</b>					
ABI StepOne	2	2	1	1	1
ABI 7000	1	1	1	0	0
ABI 7300	1	1	1	1	1
ABI 7500	12	11	17	16	16
ABI 7900	5	5	4	2	2
BioRad CFX96	1	1	1	1	1
Bio-Rad QX200	1	1	1	0	2
LightCycler 1.2	3	2	0	0	0
LightCycler 1.5	1	1	1	0	0
LightCycler 2.0	3	3	3	3	3
LightCycler 480	11	11	9	10	10
Stratagene MX3005	1	1	1	0	0
QuantStudio 3	0	0	0	1	1
QuantStudio 5	1	3	5	9	9
QuantStudio 7	1	1	3	2	3
Flex Quant Studio 12k	0	0	0	1	1
RotorGene 3000	1	1	1	1	2
Rotor Gene 6000	6	5	6	5	4
Viiia 7DX	2	2	3	2	2
Not specified	0	0	0	1	0

**Supplementary Table 2: Scoring criteria for MR<sup>4.5</sup> detection**

Category	Score
<b>1 Median cell line lysate control gene copy number</b>	
The median total control gene copy number reported was calculated for every replicate of every cell line lysate sample (n=18). A score was assigned based on the deviation from the expected median copy number (ABL1 1.4E+05, GUSB 3.35E+05)	
2-fold lower than expected value	6
3-fold lower	5
5-fold lower	2
>5-fold lower	0
<b>2 % MR4.5 Detection</b>	
A point was awarded for each MR4.5 sample analysed and detected (n=12, 6 replicates of the 0.0032 BCR::ABL <sup>15</sup> cell line lysate sample and 6 replicates of the MR4.5 sample in the secondary reference panel). When a sample failed for technical reasons this was excluded from the analysis. The percentage detection was defined as the number of MR4.5 detected / number of samples analysed	
100% detected	6
>85% detected	5
>65% detected	2
<65% detected	0
<b>3 cDNA copy number</b>	
The mean absolute ABL1 and BCR::ABL1 copy numbers per ul of cDNA analysed were calculated per batch. Each replicate(n=6) was assigned a score based on the deviation from the exact copy number	
2-fold higher or lower	3
3-fold higher or lower	2
5-fold higher or lower	1
>5 fold higher or lower	0
The <b>final cDNA copy number score</b> was assigned as : 100*(sum of score for all replicates / 18) and scored as follows:	
100%	3
>85%	2
>65%	1
<65%	0
<b>4 cDNA ratio</b>	
The % BCR::ABL1/ABL1 for the plasmid sample for each batch was calculated. The plasmid sample contained equal copies of BCR::ABL1 and control genes and therefore the expected % is 100%. Each replicate (n=3) was assigned a score based on the deviation from 100%.	
80% - 120%	2
60 - 80% or 120% - 140%	1
<60% or >140%	0
The <b>final cDNA ratio score</b> was assigned as : 100*(sum of score for all replicates / 6) and scored as follows:	
100%	3
>85%	2
>60%	1
<60%	0
<b>5 Audit of control gene values for laboratory samples</b>	
The percentage of control gene values >32000 (ABL1) or 76,800 (GUSB) were calculated for each laboratory and the following scores assigned:	
>80% of samples	5
>60% of samples	3
<60% of samples	0
<b>FINAL SCORE</b>	
The final score = 100* (sum of scores for 5 categories / 23)	
>80%	<b>Green: Can detect MR4.5 in a high proportion of samples</b>
>60%	<b>Orange: Can detect MR4.5 in most samples</b>
<60%	<b>Red: Unable to detect MR4.5 in most samples</b>

**Supplementary Table 3:** Summary statistics for CF values per year from all data sets analysed from 2016 – 2021 for the control genes *ABL1* and *GUSB*

<b>GUSB CF</b>	<b>2016</b>	<b>2017</b>	<b>2019</b>	<b>2020</b>	<b>2021</b>	<b>Overall</b>
<b>Mean</b>	1.704	1.885	1.698	2.076	2.042	1.891
<b>Median</b>	1.525	1.602	1.645	1.485	1.444	1.576
<b>Max</b>	3.304	4.392	2.686	6.152	8.040	8.040
<b>Min</b>	0.914	0.585	0.904	0.906	0.965	0.585

<b>ABL1 CF</b>	<b>2016</b>	<b>2017</b>	<b>2019</b>	<b>2020</b>	<b>2021</b>	<b>Overall</b>
<b>Mean</b>	0.645	0.679	0.608	0.672	0.690	0.659
<b>Median</b>	0.648	0.600	0.547	0.623	0.647	0.604
<b>Max</b>	1.364	1.507	1.754	1.328	1.407	1.754
<b>Min</b>	0.199	0.246	0.271	0.418	0.372	0.199

**Supplementary Table 4:** To assess whether CFs were converting data to the IS reliably, the raw data (%BCR::ABL1 / reference gene) from each laboratory were converted to BCR::ABL1<sup>IS</sup> using the newly derived laboratory specific CF. As an example, for the 2017 round 72.3% of results for 3 test samples were within 2 fold of the expected IS value for the raw, unconverted data (left). This increased to 95.5% of results when the data were converted to BCR::ABL1<sup>IS</sup> using the newly derived CF (right).

Sample	Raw data %BCR::ABL1 / CG			BCR::ABL1 <sup>IS</sup> (2017 CF)		
	'10z'	'MMR'	'DMR'	'10z'	'MMR'	'DMR'
IS of Sample	5.3775	0.0534	0.0052	5.3775	0.0534	0.0052
2 fold upper	10.7550	0.1068	0.0105	10.7550	0.1068	0.0105
2 fold lower	2.6888	0.0267	0.0026	2.6888	0.0267	0.0026
5 fold upper	26.8876	0.2669	0.0261	26.8876	0.2669	0.0261
5 fold lower	1.0755	0.0107	0.0010	1.0755	0.0107	0.0010
Lab 1	13.6912	0.0785	0.0054	9.7539	0.0559	0.0038
Lab 4	6.7585	0.0517	0.0042	6.7045	0.0513	0.0042
Lab 5	3.0189	0.0339	0.0034	4.5502	0.0510	0.0051
Lab 6	7.2739	0.0955	0.0114	3.8961	0.0510	0.0061
Lab 7	15.2580	0.1526	0.0202	5.4387	0.0544	0.0072
Lab 8	9.0541	0.0957	0.0089	5.2426	0.0554	0.0052
Lab 9	5.9720	0.0774	0.0084	3.7037	0.0480	0.0052
Lab 10	9.4118	0.0953	0.0092	3.7273	0.0377	0.0037
Lab 11	12.5683	0.1093	0.0136	4.7884	0.0416	0.0052
Lab 12	6.3600	0.0590	0.0061	5.5201	0.0512	0.0053
Lab 14	8.6632	0.0833	0.0098	5.4820	0.0527	0.0062
Lab 15	13.3761	0.1206	0.0124	5.6868	0.0513	0.0053
Lab 16	10.7054	0.1071	0.0111	5.8398	0.0584	0.0061
Lab 18	12.6887	0.1360	0.0113	5.3617	0.0575	0.0048
Lab 20	8.3192	0.1098	0.0095	4.5667	0.0603	0.0052
Lab 21	11.9664	0.1165	0.0096	4.9271	0.0480	0.0040
Lab 24	7.2238	0.1345	0.0257	2.9034	0.0541	0.0103
Lab 25	6.9125	0.0640	0.0062	5.8869	0.0545	0.0053
Lab 26	18.3331	0.2097	0.0224	4.5156	0.0516	0.0055
Lab 27	9.3404	0.1160	0.0100	4.2409	0.0527	0.0045
Lab 28	4.3193	0.0382	0.0039	5.2528	0.0464	0.0048
Lab 29	8.7263	0.0880	0.0101	4.7157	0.0476	0.0055
Lab 30	4.3078	0.0383	0.0041	4.9341	0.0439	0.0047
Lab 31	8.4226	0.0949	0.0091	3.9898	0.0450	0.0043
Lab 33	9.9396	0.0943	0.0122	4.6845	0.0444	0.0057
Lab 35	9.4717	0.1017	0.0079	5.8462	0.0628	0.0049
Lab 36	6.5303	0.0762	0.0073	4.7161	0.0550	0.0052
Lab 38	7.6397	0.0902	0.0091	4.3123	0.0509	0.0051
Lab 39	8.1804	0.1197	0.0134	3.7515	0.0549	0.0062
Lab 40	8.6677	0.0926	0.0082	6.4750	0.0692	0.0061
Lab 41	13.2445	0.0822	0.0047	9.2605	0.0574	0.0033
Lab 42	4.8280	0.0373	0.0050	6.0293	0.0466	0.0062
Lab 43	7.2066	0.0717	0.0071	4.1715	0.0415	0.0041
Lab 45	13.6507	0.1343	0.0152	4.9769	0.0490	0.0055
Lab 46	7.1900	0.0598	0.0014	6.4194	0.0534	0.0012
Lab 47	10.3390	0.1168	0.0106	4.3562	0.0492	0.0045
Lab 49	7.4757	0.0780	0.0075	4.2198	0.0440	0.0042
Lab 51	7.6197	0.1094	ND	11.0298	0.1584	ND
Lab 52	7.2619	0.0674	0.0062	5.3775	0.0499	0.0046
Lab 53	3.2813	0.0539	0.0066	3.1005	0.0510	0.0062
Lab 55	6.6773	0.0670	0.0071	6.2374	0.0626	0.0067
Lab 57	8.1662	0.0629	0.0061	5.0872	0.0519	0.0051
Lab 58	6.4267	0.0692	0.0066	4.9896	0.0537	0.0051
Lab 59	6.9851	0.0755	0.0062	4.1912	0.0453	0.0037
Lab 60	7.4444	0.0766	0.0053	4.9226	0.0506	0.0035
Lab 2	2.3875	0.0236	0.0024	7.2586	0.0718	0.0074
Lab 3	4.1636	0.0427	0.0033	7.3370	0.0753	0.0059
Lab 13	3.7133	0.0432	0.0041	5.9191	0.0689	0.0066
Lab 17	7.9860	0.0550	0.0019	7.7767	0.0535	0.0019
Lab 19	2.6474	0.0203	0.0022	11.6270	0.0893	0.0097
Lab 22	4.3194	0.0362	0.0030	6.7902	0.0570	0.0047
Lab 32	3.1218	0.0377	0.0037	6.3076	0.0762	0.0075
Lab 34	5.4980	0.0639	0.0057	5.8364	0.0678	0.0061
Lab 37	2.6809	0.0209	0.0020	8.3277	0.0650	0.0062
Lab 44	3.3124	0.0309	0.0028	7.2125	0.0673	0.0061
Lab 48	4.1096	0.0303	0.0035	6.6150	0.0487	0.0056
Lab 50	4.7431	0.0490	0.0043	5.4605	0.0564	0.0049
Lab 54	12.7575	0.1849	0.0185	7.4633	0.1081	0.0108
Lab 56	4.0030	0.0405	0.0034	5.3932	0.0546	0.0046

**Supplementary Figure 1:** Three batches of samples (n=9) were distributed to each participating laboratory. On Day A, RNA was extracted from 5 vials of the lyophilised cell line secondary reference panel and the 3 cell line lysate samples from Batch 1. The eight RNA samples were divided into two aliquots; one set of samples were stored at -20° and the other were used to synthesize cDNA. RT-qPCR was performed on the eight cDNA samples and the additional 'cDNA' sample from Batch 1 using standard laboratory protocols. On Day B, cDNA was synthesized from the stored RNA from Day A. RT-qPCR was performed on the eight cDNA samples and the 'cDNA' sample from Batch 1 using standard laboratory protocols. After 28 days the whole process was repeated with the samples from Batch 2 (Days C and D) and after a further 28 days the Batch 3 samples were analyzed (Days E and F). The analysis of each batch was separated by 28 days if possible.

**Days A, C and E**



Secondary reference material lyophilised cell line panel n=5



Cell line lysates n=3



Extract RNA (n=8)



cDNA synthesis (n=8)




RT-qPCR (n=9)



'cDNA' Sample

**Days B, D and F**

Store RNA for >48 hr



cDNA synthesis (n=8)




RT-qPCR (n=9)




'cDNA' Sample

**Supplementary Figure 2:** Four *BCR::ABL1* negative samples were distributed. RNA was extracted from each sample and divided into 2 aliquots. cDNA1 was synthesized from RNA1 on Day A and cDNA2 from RNA2 on Day B. The cell line lysate samples provided sufficient material to generate enough RNA and cDNA to perform 18 RT-qPCR replicates (15 x *BCR::ABL1*, 3 x *ABL1*) per cDNA sample, per reagent lot on three days. (144 individual RT-qPCRs in total; 60 *BCR::ABL1* and 12 *ABL1* replicates for each reagent lot). *ABL1* replicates were included to monitor RNA and cDNA quality.




**Supplementary Figure 3:** Box plots and statistics for CF values from all data sets analysed from 2016 – 2021 for the reference genes *ABL1* and *GUSB*





**Supplementary Figure 4.** Application of Westgard rules to accept or reject each run based on the performance of high and low controls (adapted from Branford S, Hughes T. *Methods Mol Med* 2006;125:69–92; Branford S, et al. *Blood* 2008;112:3330–3338). Run 2 should be rejected as both the high and low level control results are  $>2SD$  from the established mean. Run 9 should be rejected as the high level control result is  $>3SD$ .



## EUTOS CF Spreadsheet v030322 Instructions for use

The spreadsheet has been produced to help laboratories calculate conversion factors using either sample exchange methods or secondary reference materials. The calculation is based on the linear regression approach suggested by the NIBSC (<https://www.nibsc.org/documents/ifu/09-138.pdf>; pdf document attached as supplementary material).

**PLEASE NOTE THAT THE SPREADSHEET IS NOT CE MARKED OR OTHERWISE CERTIFIED. IF YOU USE IT TO DERIVE CONVERSION FACTORS FOR CLINICAL USE THEN YOU ARE RESPONSIBLE FOR ENSURING THAT THE RESULTS ARE CORRECT.**

1. Enter the known BCR::ABL1<sup>IS</sup> value of the reference sample into column A starting in cell A2. The data can be pasted from another spreadsheet

2. Add the corresponding **unconverted** %BCR::ABL1/reference gene value obtained from your analysis of the reference sample into column B starting in cell B2. The data can be pasted in from another spreadsheet

**IMPORTANT: these values should be calculated with NO conversion factor applied.**

3. The spreadsheet allows you to add up to 180 paired values.

**Do not edit or move the data** once they have been added to the sheet. If you need to delete or move data then start a new spreadsheet.

**Do not manipulate the sheet in any way.**

Once all the data are added the **conversion factor** will be displayed in **cell I1**

Three quality parameters are also shown:

1) Lower 95% confidence interval of the slope (I11).

2) Upper 95% confidence interval of the slope (I12).

3) R<sup>2</sup> (I14).

Cells I11, I12 and I14 will appear green if the data are linear and show no bias. The 95% confidence interval of the slope must be fully contained within the range 0.83 - 1.20 and the R<sup>2</sup> should be >0.97. If the cells appear red then it may be helpful to visualise your data to see where any issues are occurring. For a visual representation of the data the log<sub>10</sub> transformed BCR::ABL1<sup>IS</sup> reference values can be plotted against the corresponding log<sub>10</sub> transformed %BCR::ABL1 / reference gene (laboratory derived non-IS) using the Chart function in Excel.

If you have an existing CF the newly derived CF can be validated using the following criteria:

Optimal (+/- 1.2 fold): Old CF / New CF = 0.83 – 1.2

Satisfactory (+/- 1.6 fold): Old CF / New CF = 0.63 – 1.58

Unvalidated: Old CF / New CF <0.63 or >1.58