

Figure S2: Percentage depth dose (PDD) (a) and lateral profiles at depths of 0 mm (b), 10 mm (c), 20 mm (d) as measured and as predicted by the eRT6 beam model in water. Measured and simulated ranges and 90% (R90%), 80% (R80%), and 50% (R50%) of the maximum dose are indicated in (a) for the PDD. Measured and simulated full-width at half-maximum (FWHM), full-width at 80% of the maximum (FW80%), and 80%-20% penumbra (P80-20%) of the lateral profiles are indicated in (b), (c), and (d).

