




E

	drilling template	navigation
angular deviation	4.6°	3°
median horizontal deviation base	0.34 mm	0.52 mm
median horizontal deviation tip	0.59 mm	0.75 mm
median vertical deviation base	0.275 mm	0.445 mm
median vertical deviation tip	0.23 mm	0.43 mm
median duration per implant	3.18 min	4.66 min



E

	drilling template	navigation
angular deviation	4.6°	3°
median horizontal deviation base	0.34 mm	0.52 mm
median horizontal deviation tip	0.59 mm	0.75 mm
median vertical deviation base	0.275 mm	0.445 mm
median vertical deviation tip	0.23 mm	0.43 mm
median duration per implant	3.18 min	4.66 min