

Table S1. The detailed prostate multi-parametric magnetic resonance imaging parameters.

Sequences	TSE T2-weighted imaging			Diffusion-weighted imaging	T1-weighted dynamic contrast-enhanced imaging
	Plane	Axial	Sagittal		
TR (msec)	5500	5040	3900	4800	4.81
TE (msec)	104	115	117	63	1.74
Time of acquisition (minutes)	3.21	2.33	3.05	4.02	2.52
FOV (mm)	200	220	240	200	260
Slice thickness (mm)	3	3	3	3	3
b-values (s/mm ²)	-	-	-	0, 50, 500, 1000	-
Matrix size	384x307	384x288	448x291	114x88	192x138

TSE: Turbo spin-echo; TR: Repetition time; TE: Echo time; FOV: Field of view.