



Published in final edited form as:

*BME Front.* 2021 ; 2021: . doi:10.34133/2021/9823268.

## Recent advances in photoacoustic tomography

Lei Li,

Lihong V. Wang\*

Caltech Optical Imaging Laboratory, Andrew and Peggy Cherng Department of Medical Engineering, Department of Electrical Engineering, California Institute of Technology, 1200 East California Boulevard, Mail Code 138-78, Pasadena, CA 91125, USA

### Abstract

Photoacoustic tomography (PAT) that integrates the molecular contrast of optical imaging with the high spatial resolution of ultrasound imaging in deep tissue has widespread applications in basic biological science, preclinical research and clinical trials. Recently, tremendous progress has been made in PAT regarding technical innovations, preclinical applications, and clinical translations. Here, we selectively review the recent progresses and advances in PAT, including the development of advanced PAT systems for small-animal and human imaging, newly engineered optical probes for molecular imaging, broad-spectrum PAT for label-free imaging of biological tissues, high-throughput snapshot photoacoustic topography, and integration of machine learning for image reconstruction and processing. We envision that PAT will have further technical developments and more impactful applications in biomedicine.

### Introduction

Biomedical imaging has played a significant role in modern medicine for diagnosing diseases, monitoring therapy, and providing biological insights into lives.<sup>1,2</sup> Photoacoustic tomography (PAT), also known as optoacoustic tomography, is an emerging biomedical imaging modality that provides cross-sectional or three-dimensional (3D) imaging of an object based on the photoacoustic (PA) effect—a physical phenomenon that converts absorbed light to sound.<sup>3</sup> Although discovery of the photoacoustic effect dates back to 1880,<sup>4</sup> PAT has enjoyed a rapid growth since early 2000s following the development in ultrasonic detectors, computations, and lasers.<sup>5</sup> Since 2010, PA imaging has become one of the largest research areas in biophotonics and still enjoys the rapid growth.<sup>6</sup>

Figure 1 illustrates the principle of PAT. Typically, PAT employs non-ionizing laser pulses directed to the object for excitation. When using microwave or radio-frequency pulses for illumination, it is referred to as thermoacoustic tomography (TAT).<sup>7,8</sup> The delivered photons are absorbed by the molecules in the object, elevating the molecules from the ground state to the excited state. Another photon or heat is emitted, when the excited molecule relaxes to the ground state. The absorbed photons can be converted to heat through non-radiative relaxation.<sup>9</sup> Via thermoelastic expansion, a pressure rise is induced by the heat.<sup>10</sup> The

\*Correspondence should be addressed to L.V.W. (LVW@caltech.edu).

pressure rise propagates (with a speed of ~1500 m/s) inside biological tissue as an acoustic wave, also mentioned as a PA wave.<sup>11,12</sup> By detecting the PA waves, we can form images that map the optical absorption of the object. Based on the image formation methods, PAT has two major implementations: PA microscopy (PAM) and PA computed tomography (PACT).<sup>6</sup> PAM forms images by raster scanning the focus of light and sound across the object, while PACT yields images by inverse reconstruct the detected signals induced by wide-field illumination.

PAT is a hybrid imaging method that harvests both optical and acoustic energy. Thus, PAT inherits the advantages of both optical and ultrasound imaging, offering rich optical contrast and high spatial resolution inside deep tissue.<sup>13</sup> PAT is directly sensitive to molecules' optical absorption.<sup>14</sup> By exciting molecules at preferred wavelengths according to the spectrum signatures, PAT achieves multicontrast imaging of molecules based on their chemical compositions.<sup>15</sup> Till now, PAT has spectroscopically imaged numerous endogenous molecules, including oxy- and deoxy-hemoglobin, oxy- and reduced myoglobin, melanin, cytochrome, DNA/RNA, bilirubin, lipid, and water, which enables PAT for anatomical, functional, metabolic, and histologic imaging (Fig. 2a).<sup>16–27</sup> Thanks to the strong optical absorption of exogenous contrast agents, e.g., micro/nanoparticles, organic dyes, and genetically encoded proteins, PAT enjoys superb sensitivity in deep tissue and offers molecular imaging (Fig. 2b).<sup>27–33</sup> Increasing the optical fluence improves the detection sensitivity, as long as the temperature increase is within the safety limit. To guarantee the safety, typically, the laser exposure to the skin surface and eye is regulated by the American National Standards Institute (ANSI) standard. Thanks to the PA effect, PAT detects ultrasonic waves induced by excitation photons, including both ballistic and scattered ones; thus, PAT achieves much deeper penetration than conventional optical microscopy relying on ballistic photons. Moreover, acoustic waves are orders of magnitude less scattered inside soft tissues; therefore, PAT provides far better spatial resolution than pure optical imaging methods in deep tissues (> 2 mm).<sup>34</sup> With a skin surface radiant exposure of 21 mJ/cm<sup>2</sup> at 1064 nm, which is only one fifth of the ANSI limit (100 mJ/cm<sup>2</sup>), PAT demonstrated an imaging thickness up to 7 cm *in vivo* with double-sided light illumination.<sup>35</sup> Because of the acoustic detection, PAT's spatial resolution and penetration depth are scalable with the detected acoustic frequency. The spatial resolution of PAT improves as the acoustic central frequency and bandwidth increase at the expense of penetration depth.<sup>36</sup> PAT has uniquely shown a capability of multiscale imaging using a consistent optical contrast, ranging from organelles, cells, and tissues to wholebody small animals and human organs, as shown in Fig. 2c.<sup>25,37–39</sup> Previously, preclinical small-animal imaging and clinical applications typically employ non-optical imaging modalities, including magnetic resonance imaging (MRI), X-ray computed tomography (X-ray CT), positron emission tomography (PET), single-photon emission computed tomography (SPECT), and ultrasound imaging (USI), which all can provide deep penetration.<sup>40</sup> However, those approaches still face challenges. For example, MRI necessitates a long data-acquisition time, not suitable for capturing fast dynamics.<sup>41,42</sup> PET and SPECT have a low spatial resolution to resolve detailed structures. X-ray CT, PET, and SPECT employ ionizing radiation, impeding longitudinal monitoring.<sup>43</sup> Conventional USI does not reveal molecular contrasts outside blood vessels.<sup>44</sup> PAT, as a noninvasive approach, achieves high-resolution imaging in deep tissues with

optical contrasts, providing a complementary approach for preclinical research and clinical translations.

In recent years, PAT has an even more rapid development, including technical innovations, various biomedical applications, and clinical translations. In this paper, we selectively review some of the recent progresses and advances in PAT, including small-animal wholebody imaging, novel molecular imaging, rapid assessments of brain functions, broad-spectrum imaging of tissues, human organ imaging, snapshot photoacoustic topography, and integration of machine learning for advanced image reconstruction. We envision that PAT will have more impactful applications in fundamental science, preclinical research, and clinical trials.

## Wholebody PACT of small animals

Small-animal wholebody imaging plays an indispensable role in preclinical study.<sup>56</sup> With optical contrasts and high spatiotemporal resolution, small-animal imaging can provide physiological understandings of biological processes and dynamics, advancing fundamental biology, preclinical study, and clinical translation.<sup>57</sup> A recent development in PACT for small-animal wholebody imaging, termed single-impulse panoramic PACT (SIP-PACT), permits anatomical, functional, and molecular wholebody imaging with exceptional quality and speed.<sup>58</sup> The schematic of SIP-PACT is shown in Fig. 3a. SIP-PACT employed full-ring ultrasonic detection (512 elements, 5-MHz central frequency, over 90% one-way bandwidth) with parallel amplification and digitization, maximizing the detection signal-to-noise ratio (SNR) and speed. The light illumination and acoustic detection are aligned confocally to optimize the detection sensitivity (Fig. 3b). It takes 50  $\mu$ s for SIP-PACT to acquire a 2-dimensional (2D) image of a cross-section *in vivo*. The in-plane panoramic acoustic detection of SIP-PACT offered an isotropic resolution ( $\sim$ 125  $\mu$ m) within the whole cross-section and full-view fidelity. Moreover, to better reveal detailed structures of internal organs, a half-time dual-speed-of-sound universal back-projection algorithm was developed to account for the acoustic inhomogeneity between the animal tissue and the surrounding coupling medium (water). Representative small-animal wholebody images, acquired at 1064 nm, from SIP-PACT are shown in Fig. 3c–f.<sup>58</sup>

Recently, a hemispherical transducer array based PACT system (Fig. 3g) has been built to image the wholebody of small animals.<sup>59</sup> By spiral scanning the hemispherical transducer array (256 elements, 4-MHz central frequency, 100% one-way bandwidth) around the animal, volumetric wholebody imaging can be achieved (Fig. 3g, the wholebody tissue is shown in gray; illumination wavelength, 800 nm). By parking the array at a given position and acquiring data continuously, high-speed imaging of the dynamics inside an organ can be realized. For example, the high imaging speed reveals cardiac dynamics *in vivo* in detail, as shown in Fig. 3h (the beating heart is shown in color).<sup>59</sup>

The recent progress in small-animal PACT promises more extensive preclinical studies, particularly when high spatiotemporal resolution and high sensitivity are necessary. For instance, tracking immune cells can visualize immune responses during cancer progression and treatment, advancing our understanding of immune system dynamics. Overall, the

development of small-animal PACT enables real-time imaging of biological processes in basic life science research and ultimately in clinical applications.

## Molecular PACT

Because endogenous contrast agents often lack specificity or sensitivity, scientists usually rely on exogenous contrast agents that can visualize biological phenomena with high sensitivity and specificity to study biological activities. Exogenous contrast agents for deep-penetrating PAT hold two key advantages: (1) they are optimized for high optical absorption at near-infrared (NIR) wavelengths for high detection sensitivity inside deep tissues; (2) they are specifically designed to conjugate with targeting moieties to selectively bind to receptors to achieve high specificity. Exogenous contrast agents, such as organic dyes<sup>60–62</sup>, genetically encoded proteins<sup>63–66</sup>, and micro/nanoparticles<sup>67–71</sup>, have been extensively explored for PA molecular imaging. With the innovations and advances in exogenous contrast agents, molecular PAT has recently enjoyed a prosperous development (Fig. 4).

A tyrosinase-based reporter that triggers human cells to express eumelanin has been developed to provide a strong PA contrast.<sup>64</sup> Combining the tyrosinase-based reporter and a Fabry-Perot based ultrasound sensor (22 MHz bandwidth), PACT has imaged transgenic cancer cells expressing eumelanin (Fig. 4a, illumination wavelength, 600 nm) and monitored tumor development *in vivo*.<sup>64</sup> An *in vivo* pH mapping technology, consisting of a photoacoustic pH indicator, SNARF-5F, has been developed.<sup>72</sup> Facilitated by multi-wavelength PACT, the pH levels inside a tumor have been quantitatively measured. Here, a linear transducer array (128 elements, 11.25-MHz central frequency, 75% one-way bandwidth) based PACT system was used. The SNARF-5F based nanoparticles were injected via tail-vein into a tumor-bearing mouse; multi-wavelength PACT of the mouse at 75 min post-injection has mapped the pH distribution inside the tumor *in vivo* (Fig. 4b, illumination wavelengths, 565 nm and 600 nm).<sup>72</sup> Bacterial outer membrane vesicles (OMVs) have been explored for delivering drugs, vaccines, and immunotherapy agents. Recently, bioengineered OMVs for contrast-enhanced PA imaging have been investigated.<sup>73</sup> OMVs encapsulating biopolymer-melanin (OMV<sup>Mel</sup>) were produced using a tyrosinase-expressing bacterial strain, which provides strong optical absorption at NIR wavelengths. Here, an arc-shaped transducer array (270° in-plane detection angle, 256 elements, 5-MHz central frequency, 90% one-way bandwidth) based PACT was used. Using multi-wavelength PACT, tumor-associated OMV<sup>Mel</sup> distribution has been monitored *in vivo* (Fig. 4c, illumination wavelengths, 680–900 nm, 10-nm interval).<sup>73</sup>

Genetically encoded photoswitchable proteins (e.g., RpBphP1, DrBphP-PCM, etc.) that effectively improve the PA imaging sensitivity and specificity have been reported recently.<sup>31,32,63,65,66,74,75</sup> Here, a 512-element full-ring transducer array (5-MHz central frequency, 90% one-way bandwidth) based PACT system was used. By analyzing the unique decay characteristics of different photoswitchable proteins, quantitative multicontrast imaging has been achieved to differentiate multiple types of tumors in deep tissue *in vivo* (Fig. 4d, illumination wavelength, 780 nm).<sup>65,74</sup> In addition, a split version of DrBphP-PCM has been engineered, providing the first bimolecular PA complementation reporter to detect protein-protein interactions in deep tissue *in vivo* (Fig. 4e).<sup>74</sup>

Substantial progress in the development of micro/nanorobots has been accomplished for biomedical applications in recent years. However, present micro/nanorobot platforms are ineffective for imaging and motion control in deep tissue *in vivo*. Recently, a PACT-navigated microrobotic system that achieved controlled propulsion and prolonged cargo retention *in vivo* was reported, where a full-ring transducer array (512 elements, 5-MHz central frequency, 90% one-way bandwidth) was used for acoustic detection.<sup>76,77</sup> Thanks to the molecular contrast and high spatiotemporal resolution at depths, PACT can locate and navigate the microrobots *in vivo*. As shown in Fig. 3f (illumination wavelength, 750 nm), PACT visualizes the migration of microrobots in the intestines in real time *in vivo*.<sup>76</sup> The integration of PACT and the newly engineered microrobotic system allows deep tissue imaging and accurate control of the microrobots *in vivo* and paves a new path for precision medicine.

## PAT of the brain

Studying how the brain works will directly benefit basic science and help us to better understand and treat neurological disorders, such as Alzheimer's and Parkinson's disease.<sup>78</sup> Thanks to optical contrast and deep penetration, PAT provides a powerful tool for multi-scale functional brain imaging. Based on the endogenous molecules, including hemoglobin, lipid, and cytochrome, label-free PAT has resolved whole-brain vasculature and structures. Taking advantage of the high resolution and high sensitivity, PAM has mapped cortical vasculature and quantified oxygen saturation of hemoglobin (sO<sub>2</sub>) at the single capillary level with the skull intact (Fig. 5a, illumination wavelength, 532 nm).<sup>79</sup> After mitigating the blood signals from the brain, label-free PACT has imaged detailed internal brain structures with MRI image quality. Based on the lipid and cytochrome contrast, a full-ring transducer array based PACT (512 elements, 5-MHz central frequency, 90% one-way bandwidth) identified brain structures, including neocortex, corpus callosum, hippocampus, inferior colliculus, and cerebellum (Fig. 5b, illumination wavelength, 600 nm).<sup>80</sup> Employing NIR light illumination, a hemispherical transducer array based PACT (512 elements, 140° angular tomographic coverage, 5-MHz central frequency, 100% one-way bandwidth) has imaged a mouse brain with structural details revealed in 3D *ex vivo* (Fig. 5c, illumination wavelength, 740 nm).<sup>81</sup> A full-ring transducer array based PACT (512 elements, 5-MHz central frequency, 90% one-way bandwidth) has also monitored the whole rat brain resting-state hemodynamics and mapped the whole-brain functional connectivity (Fig. 5d, illumination wavelength, 1064 nm).<sup>58</sup>

The deep penetration and high spatiotemporal resolution enable PACT to monitor large-scale neural activities in detail. As a demonstration, a hemispherical transducer array based PACT (512 elements, 5-MHz central frequency, 100% one-way bandwidth) has imaged rapid calcium responses to electrical stimulations in a GCaMP6s expressing mouse brain *in vivo* (Fig. 5e, illumination wavelength, 488 nm).<sup>82</sup> In addition, a linear transducer array based PACT (256 elements, 21-MHz central frequency, 52% one-way bandwidth) has also visualized the epileptic wave propagation across the whole brain during a seizure (Fig. 5f, illumination wavelength, 1064 nm).<sup>83</sup> With the capability to monitor neuronal activities and hemodynamics across the whole brain, PAT has displayed encouraging potentials in

studying various brain disorders and diseases, such as traumatic disorders, brain cancer, stroke, Alzheimer's disease and seizures of various etiologies.

## Broad spectrum PAM

Any molecule has a unique absorption spectrum and has a fluorescence quantum yield lower than 100%, providing contrast for PAT.<sup>27</sup> By tuning the wavelength from ultraviolet (UV) to mid-infrared (MIR), spectral PAT has detected endogenous biological molecules, including cytochromes, DNA/RNA, hemoglobin, myoglobin, melanin, protein, lipids, and water.<sup>19,20,84–89</sup>

Breast-conserving surgery aims to excise all cancer cells. However, no intraoperative device that can quickly examine the lumpectomy specimen at a microscopic scale is available. Thus, up to 60% of patients have to undertake second surgeries to reach clear margins. PAM employing UV illumination (UV-PAM) can specifically highlight cell nuclei, yielding microscopic images with a similar contrast as hematoxylin labeled histological images.<sup>20</sup> A fast UV-PAM system was recently developed to provide high-resolution histology-like images of unprocessed and unlabeled breast tissues.<sup>89</sup> Figure 6a shows UV-PAM images of a fixed, unprocessed breast tumor specimen with a field of view (FOV) of 10 mm × 4.2 mm (illumination wavelength, 266 nm). The close-up images (Fig. 6b and c), corresponding to the red and yellow dash boxed regions in Fig. 6a, reveal detailed structures of the carcinoma. As shown in Fig. 6d (a zoomed-in image of the magenta dash boxed region in Fig. 6a), it is clear UV-PAM can resolve individual cell nuclei. Label-free UV-PAM with histology-like imaging capability has demonstrated its potential as an intraoperative margin assessment tool for surgeons and pathologists to identify tumor margins.<sup>89</sup>

Conventional optical microscopy is fundamentally limited by a lack of chemical specificity or by cellular phototoxicity, thus label-free optical imaging cannot directly reveal biomolecule dynamics in living cells.<sup>90,91</sup> Recently, MIR-PAM, based on chemically specific vibrational excitation by MIR absorption, has been developed to provide label-free bond-selective metabolic imaging in live cells.<sup>92</sup> MIR-PAM enables spatiotemporal profiling of lipids, proteins and carbohydrates in cells and tissues. As a demonstration, MIR-PAM monitored lipid and protein dynamics during lipolysis in live cells (Fig. 6e, illumination wavelength, 3500 nm), proving its capability of imaging of biomolecular dynamics in living cells without labeling.<sup>92</sup>

However, the long MIR wavelength fundamentally limits the spatial resolution of MIR-PAM due to optical diffraction.<sup>92</sup> In addition, the high water content in fresh samples severely reduces the imaging contrast due to its strong MIR absorption. A recent innovation that employs MIR PA imaging localized with a pulsed UV illumination (266 nm) successfully overcomes the above limitations. This technology, termed ultraviolet-localized MIR PAM (ULM-PAM), has achieved high-resolution MIR imaging of fresh samples without water background.<sup>86</sup> ULM-PAM employed a focused mid-MIR laser pulse to excite the sample. A confocally aligned UV laser pulse photoacoustically detected the MIR-induced temperature rise, thereby revealing MIR absorption contrast. ULM-PAM's lateral resolution, defined by the UV wavelength, is over 10-fold higher than conventional MIR microscopy. Moreover,



most biomolecules in living cells, including lipids, proteins, and nucleic acids, have strong absorption of UV light (200–230 nm). Meanwhile, UV light is transmissive in water, which significantly suppresses the water background in ULM-MIR. Illuminating formalin-fixed 3T3 mouse fibroblast cells at wavelengths of 3420 nm and 6050 nm, respectively, ULM-PAM mapped detailed distribution of lipids and proteins in living cells (Fig. 6f and g).<sup>86</sup> As a comparison, the MIR-PAM images of lipids (Fig. 6h) and proteins (Fig. 6i) display a high water background and low spatial resolution.<sup>86</sup> As a demonstration, ULM-PAM has imaged neonatal (Fig. 6j) and mature (Fig. 6k) cells, showing high-resolution, high-contrast imaging of lipids (blue), proteins (green), and nucleic acids (red).<sup>86</sup> Therefore, ULM-PAM permits label-free imaging of biological samples at high-resolution with high-contrast.

## PACT of human breasts

Breast cancer is the number two cause of cancer death globally (11.6%), with a worrying mortality rate of 6.6%.<sup>93,94</sup> Recently, substantial progress has been made for noninvasive breast cancer diagnosis utilizing PAT. A technical advance, termed single-breath-hold PACT (SBH-PACT), can obtain a 3D image of a whole breast within a single breath hold (~15 s).<sup>95</sup> SBH-PACT employed a full-ring transducer array (512 elements, 2.25-MHz central frequency, 95% one-way bandwidth) for acoustic detection and 1064-nm light for excitation. The high detection sensitivity of SBH-PACT allows detecting breast tumors in detail, promising wide applications in clinical breast care. By detecting local angiogenesis, SBH-PACT can differentiate lesions from normal tissues. Highly correlated with the tumor sites indicated in X-ray mammograms (Fig. 7a), SBH-PACT can identify a tumor by revealing its higher blood vessel densities (Fig. 7b).<sup>95</sup> By further examining a tumor-containing slice (marked by white dashed lines in Fig. 7b), the same tumor, displaying higher PA signals, can be visualized at the corresponding location (Fig. 7c).<sup>95</sup> Furthermore, a vessel density map of the breast was computed, where the breast tumor was highlighted due to its high vessel density (Fig. 7d).<sup>95</sup>

A PACT system with spiral scanning of a hemispherical transducer array (512 elements, 2-MHz central frequency, 90% one-way bandwidth) was recently reported for human breast imaging.<sup>96</sup> The dense spatial sampling and the isotropic 3D resolutions yielded high-quality imaging of a healthy breast (Fig. 7e, illumination wavelength, 795 nm). By illuminating the breast tissue at two wavelengths of 756 nm and 797 nm, respectively, hemoglobin oxygen saturation ( $sO_2$ ) was also evaluated using an approximate parameter, termed S-factor, computed from measurements obtained at the two wavelengths.<sup>97</sup> The S-factors can be evaluated on neighboring vessels, such as adjacent arteries and veins, assuming that the optical fluence in the neighboring region is the same. As shown in Fig. 7f, arteries and veins in a healthy breast can be clearly distinguished in the S-factor image.<sup>97</sup> Figure 7g, a fusion of the ultrasound (red color) and S-factor images, shows an example of the results obtained from a breast cancer lesion. Notably, the detailed vasculature surrounding the tumor is clearly visible. In addition, the arterioles and venules show clustering.<sup>97</sup>

Recent advances in PACT for breast imaging hold the potential to complement X-ray mammography for breast cancer diagnosis and treatment monitoring. Unlike X-ray mammography, PACT uses light for excitation, which is non-ionizing and safe. In the

meantime, it provides sufficient penetration in the breast. Moreover, the optical absorption provides a much higher soft-tissue contrast than the X-ray mammography. Further, PACT can noninvasively monitor breast tumor responses to chemotherapy as treatment proceeds.

## PACT of human extremities

Vascular disease is the leading cause of death in the United States (~30%).<sup>97</sup> And vascular disease most commonly presents in appendicular regions. In addition, peripheral blood vessel examination can identify the visceral disease and estimate an individual's lifestyle.<sup>98</sup> Therefore, angiographic imaging of extremities can offer significant insights into the health conditions of patients, especially for hypertensive, diabetic, and hyperlipidemic patients.<sup>97</sup> Thanks to its high sensitivity of blood, deep tissue penetration, and noninvasiveness, PACT becomes a promising modality for imaging the vasculature of extremities.

A newly designed PACT system equipped with a hemispherical transducer array (1024 elements, 3.34-MHz central frequency, 85% one-way bandwidth) was developed to image the vasculature of human limbs.<sup>99</sup> After scanning the detector array, it can image a field of view up to 180 mm × 270 mm within 10 minutes, providing high-quality images of human limbs. As shown in Fig. 8a–d (illumination wavelength, 797 nm), various extremities, including palm (Fig. 8a), back of the hand (Fig. 8b), forearm (Fig. 8c), and lower thigh (Fig. 8d), have been imaged, revealing detailed vasculature.<sup>99</sup>

Recently, a PACT system with a Fabry-Perot interferometric ultrasound sensor (30-MHz bandwidth, –3 dB points) was built.<sup>100</sup> A volumetric image (14 × 14 × 14 mm<sup>3</sup>) can be acquired within 90 s at a 30-Hz laser repetition rate (illumination wavelength, 750 nm). Fingertips were imaged before and after a thermal stimulus. As shown in Fig. 8e, fewer vessels were depicted after cold water immersion, showing thermally induced peripheral vasoconstriction.<sup>100</sup>

These experiments demonstrated the capability of PACT to image peripheral vessels and their responses to vasomotor changes, promising the diagnosis of peripheral vascular diseases.

## Photoacoustic topography through an ergodic relay

In previous sections, we review the advanced technical innovations and their widespread applications of both PAM and PACT. However, till now, PAM still requires serial detection using a focused detector for data acquisition, which limits the throughput. While PACT employs parallel detection using multiple detection elements and multi-channel amplifiers and digitizers, which is complex and expensive. Recently, a high-throughput PA imaging technique based on an ergodic relay (ER) has been developed, which employs a single-element detector to obtain snapshot wide-field images. This technology is termed as photoacoustic topography through an ER (PATER).<sup>101</sup> The ER, a critical component in PATER, is a waveguide that permits acoustic waves originated from any input point to reach any other output point with distinctive reverberant characteristics.<sup>102</sup> Here, the ER effectively encodes each one-dimensional depth image into a unique temporal sequence. Because of the uniqueness of each temporal signal, PA waves from the whole volume



through the ER can be recorded in parallel. Finally, we can decode them mathematically to reconstruct 2D projection images.

The mechanism of PATER has been illustrated in Fig. 9a–f. PATER has two imaging modes: the calibration mode (Fig. 9a) and the wide-field mode (Fig. 9b). In the calibration mode, a focused laser pulse illuminates the object. The PA wave excited by the focused laser pulse can be treated as a spatiotemporal delta function of the PATER system. An unfocused needle transducer (20-MHz central frequency, 56% one-way bandwidth) was used for acoustic detection. Each calibration acquisition measures the impulse response of the system at each pixel. After raster scanning the focus across the entire FOV, the whole system's impulse responses are obtained (Fig. 9c). By computing the root-mean-squared amplitude of the PA signal received at each pixel, PATER forms a projection image of the object (Fig. 9d). In the wide-field mode, a broad laser beam excites the whole FOV. PA waves from the whole volume are detected in parallel, allowing high-speed snapshot wide-field imaging (Fig. 9e). By solving an inverse problem, PATER reconstructs wide-field images (Fig. 9f).

Because of the high propagation speed of blood pulse waves in major arteries, visualization of pulse wave propagation is still challenging. Thanks to PATER's high imaging speed, real-time imaging of pulse wave propagation *in vivo* has been achieved. After calibration, wide-field measurements the mouse brain were taken. A laser diode generated two focal spots to heat the middle cerebral arteries (MCAs) during the wide-field recording. The wide-field images (Fig. 9g, illumination wavelength, 532 nm) reveal thermal wave propagation with the blood flow in the MCAs. By analyzing the time-lapse images, we can compute the pulse wave velocity (PWV) in major arteries, a key physiological parameter indicating cardiovascular disease progression. PATER facilitates imaging of fast dynamics *in vivo* and holds great potentials for wide-ranging biomedical applications—such as high-speed imaging of neural action potentials and high-throughput analysis of histological tissues.

## Machine learning based image reconstruction

Machine learning and deep learning approaches have prompted a paradigm shift in biomedical imaging.<sup>103,104</sup> In PAT, various image reconstruction algorithms based on physics have been studied;<sup>105–109</sup> the recent development of machine learning (ML) offers new possibilities for advanced reconstruction and processing methods. For example, ML methods have been used to identify cancer lesions by feature learning of texture patches<sup>110</sup>. Convolutional neural networks (CNNs) have been explored to remove streak-type artifacts from the PACT images<sup>111–113</sup> and applied to quantitative spectral-unmixing of multiple chromophores in deep tissue<sup>114</sup>. Here we mainly focus on two major applications—reconstruction with sparse data and limited-view PACT.

In PACT, strategies for fast data acquisition usually involve sub-optimal spatial or temporal sampling of the raw channel data, leading to a trade-off between the imaging quality and speed. Recently, a deep CNN based framework for high-performance image reconstruction based on sparsely-sampled PA data was proposed.<sup>115</sup> The U-Net architecture (Fig. 10a) was trained to recognize the artifacts from sparse data reconstruction.<sup>115</sup> Then, the trained network was applied to *in vivo* imaging. Mice were imaged by a PACT scanner equipped

with a full-ring transducer array consisting of 512 elements (5-MHz central frequency, 80% one-way bandwidth). The reconstructed images with all 512-channel data were used as the ground truth for training. Results of the artifact removal from cross-sectional images reconstructed with 128-channel data are shown in Fig. 10 b–e (illumination wavelength, 1064 nm), respectively.<sup>115</sup> The internal structures of the mouse, which can barely be resolved in the original reconstruction (Fig. 10b) due to the overwhelming artifacts, have been revealed after the network correction (Fig. 10d).<sup>115</sup> The reconstruction algorithm can differentiate the small blood vessels from the streak artifacts (labeled by red arrows in the close-up panels in Fig. 10c and e) and selectively suppress the artifacts.<sup>115</sup>

Linear transducer array based PACT often suffers from limited-view imaging.<sup>116,117</sup> Recently, a hybrid neural network was proposed to address this issue. The hybrid neural network, Y-Net, is a CNN framework to reconstruct PA images. Y-Net combines the raw channel data and beamformed images as input.<sup>118</sup> As illustrated in Fig. 10f, the Y-Net connects two encoder paths with one decoder path, utilizing the information from channel data and beamformed images. Encoder I encodes the texture features, while Encoder II encodes the physical features. The decoder concatenates the two encoder outputs and yields the final images. Y-Net's performance was tested by numerical simulations (Fig. 10g). Comparing with the ground truth image (Fig. 10g, top left panel), the images reconstructed using conventional methods, including universal back-projection (Fig. 10g, top right panel) and time reversal reconstruction (Fig. 10g, bottom left panel), reveal missing vertical features due to the limited view issue. The Y-Net output successfully mitigates the problem and reconstructs an image much closer to the ground truth.<sup>118</sup>

## Outlook

The rapidly growing applications of PAT in basic life science study and clinical translation provide strong momentum for the continuous development of PAT. The recent advances in small-animal imaging promise whole-brain monitoring of neural activities. For visualizing neural responses in the deep brain, new NIR calcium- or voltage-sensitive indicators with performance optimized for PAT need to be engineered. The tremendous progress in genetically encoded PA probes enables noninvasive tracking of tumor growth and metastasizing, providing a powerful tool to better understand the tumor and find an effective treatment. The elegant marriage between PAT and the microrobotic system permits real-time navigation of the microrobots and promises precision medicine.

The high-impact applications of human breast and extremity imaging boost a fast expansion of PAT in clinical translations. PAT provides a noninvasive and powerful approach for breast cancer diagnosis, chemotherapy monitoring, and peripheral vascular imaging for diabetic patients, complementing current clinical methods in contrast mechanism, spatiotemporal resolution, and penetration. Currently, photon dissipation restricts the ultimate light penetration to ~10 cm in mammalian tissue, which prevents wholebody imaging of human adults using PAT. A potential solution is to employ microwaves for excitation. Biological tissues are more transparent to microwaves, which holds the promise to penetrate beyond 10 cm and permit visualization of internal organs in adults or whole neonatal bodies.

The innovation of PATER offers a solution for low-cost high-throughput PA imaging, enabling the miniaturization of the PAT system toward portable and wearable applications. Particularly, PATER is envisaged to be used as a wearable device to monitor human vital signs. ML-based PA image reconstruction and image processing methods enjoy the latest development and achievements from CNN, providing a novel perspective to address the artifacts in image reconstruction and enhance the performance of PAT. With the recent advances in PAT, we expect more widespread and impactful applications in fundamental science, preclinical research, and clinical translation.

## Acknowledgment

This work was supported in part by National Institutes of Health grants R01 NS102213, U01 NS099717 (BRAIN Initiative), R35 CA220436 (Outstanding Investigator Award), R01 EB028277, and U01 EB029823 (BRAIN Initiative).

## Competing interests

L.V.W. has financial interests in Microphotoacoustics, Inc., CalPACT, LLC, and Union Photoacoustic Technologies, Ltd., which did not support this work.

## References

1. Ntziachristos V Going deeper than microscopy: the optical imaging frontier in biology. *Nat. Methods* 7, 603–614, doi:10.1038/nmeth.1483 (2010). [PubMed: 20676081]
2. Ntziachristos V, Ripoll J, Wang LV & Weissleder R Looking and listening to light: the evolution of whole-body photonic imaging. *Nat. Biotechnol* 23, 313–320, doi:10.1038/nbt1074 (2005). [PubMed: 15765087]
3. Beard P Biomedical photoacoustic imaging. *Interface Focus* 1, 602–631, doi:10.1098/rsfs.2011.0028 (2011). [PubMed: 22866233]
4. Bell AG ART. XXXIV.--On the Production and Reproduction of Sound by Light. *American Journal of Science* (1880–1910) 20, 305 (1880).
5. Li L, Lin L & V. Wang L Multiscale Photoacoustic Tomography. *Opt. Photon. News* 29, 32–39, doi:10.1364/OPN.29.4.000032 (2018).
6. Wang LV & Yao J A practical guide to photoacoustic tomography in the life sciences. *Nat. Methods* 13, 627–638 (2016). [PubMed: 27467726]
7. Kruger RA et al. Thermoacoustic CT with radio waves: a medical imaging paradigm. *Radiology* 211, 275–278, doi:10.1148/radiology.211.1.r99ap05275 (1999). [PubMed: 10189483]
8. Kruger RA et al. Breast cancer in vivo: contrast enhancement with thermoacoustic CT at 434 MHz-feasibility study. *Radiology* 216, 279–283, doi:10.1148/radiology.216.1.r00jl30279 (2000). [PubMed: 10887262]
9. Xu M & Wang LV Photoacoustic imaging in biomedicine. *Review of Scientific Instruments* 77, 041101, doi:10.1063/1.2195024 (2006).
10. Wang LV & Wu H *Biomedical Optics: Principles and Imaging* 2007. Hoboken: John Wiley & Sons.
11. Wang X et al. Noninvasive laser-induced photoacoustic tomography for structural and functional in vivo imaging of the brain. *Nat. Biotechnol* 21, 803 (2003). [PubMed: 12808463]
12. Zhang HF, Maslov K, Stoica G & Wang LV Functional photoacoustic microscopy for high-resolution and noninvasive in vivo imaging. *Nat. Biotechnol* 24, 848 (2006). [PubMed: 16823374]
13. Razansky D, Buehler A & Ntziachristos V Volumetric real-time multispectral optoacoustic tomography of biomarkers. *Nat. Protocols* 6, 1121–1129, doi:<http://www.nature.com/nprot/journal/v6/n8/abs/nprot.2011.351.html#supplementary-information> (2011). [PubMed: 21738125]

14. Razansky D et al. Multispectral opto-acoustic tomography of deep-seated fluorescent proteins in vivo. *Nature Photonics* 3, 412, doi:10.1038/nphoton.2009.98 <https://www.nature.com/articles/nphoton.2009.98#supplementary-information> (2009).
15. Weber J, Beard PC & Bohndiek SE Contrast agents for molecular photoacoustic imaging. *Nat. Methods* 13, 639–650 (2016). [PubMed: 27467727]
16. Li L et al. Fully motorized optical-resolution photoacoustic microscopy. *Optics letters* 39, 2117–2120 (2014). [PubMed: 24686689]
17. Lin L, Yao J, Li L & Wang LV In vivo photoacoustic tomography of myoglobin oxygen saturation. *Journal of biomedical optics* 21, 061002 (2015).
18. Krumholz A, Yao J, Wang LV, Fleming T & Gillanders WE Photoacoustic microscopy of tyrosinase reporter gene in vivo. *Journal of biomedical optics* 16, 080503 (2011). [PubMed: 21895303]
19. Zhang C, Zhang YS, Yao D-K, Xia Y & Wang L Label-free photoacoustic microscopy of cytochromes. *Journal of biomedical optics* 18, 020504 (2013). [PubMed: 23370407]
20. Yao D-K, Maslov K, Shung KK, Zhou Q & Wang LV In vivo label-free photoacoustic microscopy of cell nuclei by excitation of DNA and RNA. *Optics letters* 35, 4139–4141 (2010). [PubMed: 21165116]
21. Zhou Y, Zhang C, Yao D-K & Wang L Photoacoustic microscopy of bilirubin in tissue phantoms. *Journal of biomedical optics* 17, 126019 (2012). [PubMed: 23235894]
22. Wang H-W et al. Label-free bond-selective imaging by listening to vibrationally excited molecules. *Physical review letters* 106, 238106 (2011). [PubMed: 21770549]
23. Xu Z, Li C & Wang LV Photoacoustic tomography of water in phantoms and tissue. *Journal of biomedical optics* 15, 036019 (2010). [PubMed: 20615021]
24. Omar M, Aguirre J & Ntziachristos V Optoacoustic mesoscopy for biomedicine. *Nature biomedical engineering*, 1 (2019).
25. Aguirre J et al. Precision assessment of label-free psoriasis biomarkers with ultra-broadband optoacoustic mesoscopy. *Nature Biomedical Engineering* 1, 0068 (2017).
26. Yao J, Maslov K, Zhang Y, Xia Y & Wang L Label-free oxygen-metabolic photoacoustic microscopy in vivo. *Journal of biomedical optics* 16, 076003 (2011). [PubMed: 21806264]
27. Yao J & Wang LV Sensitivity of photoacoustic microscopy. *Photoacoustics* 2, 87–101, doi:10.1016/j.pacs.2014.04.002 (2014). [PubMed: 25302158]
28. Li L, Yao J & Wang LV in *Wiley Encyclopedia of Electrical and Electronics Engineering* 1–14 (John Wiley & Sons, Inc., 2016).
29. Zhou M et al. in *Design and Applications of Nanoparticles in Biomedical Imaging* 337–356 (Springer, Cham, 2017).
30. Li L Multi-Contrast Photoacoustic Computed Tomography, California Institute of Technology, (2019).
31. Li L et al. in *Photons Plus Ultrasound: Imaging and Sensing 2019*. 1087818 (International Society for Optics and Photonics).
32. Yao J et al. in *Photons Plus Ultrasound: Imaging and Sensing 2016*. 97082U (International Society for Optics and Photonics).
33. Ntziachristos V & Razansky D Molecular imaging by means of multispectral optoacoustic tomography (MSOT). *Chemical reviews* 110, 2783–2794 (2010). [PubMed: 20387910]
34. Wang LHV & Hu S Photoacoustic tomography: in vivo imaging from organelles to organs. *Science* 335, 1458–1462, doi:10.1126/science.1216210 (2012). [PubMed: 22442475]
35. Nyayapathi N et al. Dual Scan Mammoscope (DSM)—A New Portable Photoacoustic Breast Imaging System With Scanning in Craniocaudal Plane. *IEEE Transactions on Biomedical Engineering* 67, 1321–1327, doi:10.1109/TBME.2019.2936088 (2020). [PubMed: 31425013]
36. Wang LV Multiscale photoacoustic microscopy and computed tomography. *Nature Photonics* 3, 503–509, doi:10.1038/nphoton.2009.157 (2009). [PubMed: 20161535]
37. Danielli A et al. Label-free photoacoustic nanoscopy. *Journal of biomedical optics* 19, 086006 (2014). [PubMed: 25104412]


38. Yao J, Wang L, Li C, Zhang C & Wang LV Photoimprint Photoacoustic Microscopy for Three-Dimensional Label-Free Subdiffraction Imaging. *Physical Review Letters* 112, 014302, doi:10.1103/PhysRevLett.112.014302 (2014). [PubMed: 24483902]
39. Deán-Ben XL, Fehm TF, Ford SJ, Gottschalk S & Razansky D Spiral volumetric optoacoustic tomography visualizes multi-scale dynamics in mice. *Light: Science & Applications* 6, e16247–e16247 (2017).
40. Xia J & Wang LV Small-animal whole-body photoacoustic tomography: a review. *IEEE Transactions on Biomedical Engineering* 61, 1380–1389 (2013). [PubMed: 24108456]
41. Wu D & Zhang J In vivo mapping of macroscopic neuronal projections in the mouse hippocampus using high-resolution diffusion MRI. *NeuroImage* 125, 84–93 (2016). [PubMed: 26499812]
42. Alomair OI, Brereton IM, Smith MT, Galloway GJ & Kurniawan ND In vivo high angular resolution diffusion-weighted imaging of mouse brain at 16.4 Tesla. *PloS one* 10, e0130133 (2015). [PubMed: 26110770]
43. Brenner DJ & Hall EJ Computed tomography—an increasing source of radiation exposure. *New England Journal of Medicine* 357, 2277–2284 (2007). [PubMed: 18046031]
44. Greco A et al. Ultrasound biomicroscopy in small animal research: applications in molecular and preclinical imaging. *Journal of Biomedicine and Biotechnology* 2012 (2012).
45. Yao J, Maslov K, Hu S & Wang L Evans blue dye-enhanced capillary-resolution photoacoustic microscopy in vivo. *Journal of biomedical optics* 14, 054049 (2009). [PubMed: 19895150]
46. Razansky D, Baeten J & Ntziachristos V Sensitivity of molecular target detection by multispectral optoacoustic tomography (MSOT). *Medical Physics* 36, 939–945, doi:10.1118/1.3077120 (2009). [PubMed: 19378754]
47. Pan D et al. Molecular Photoacoustic Tomography with Colloidal Nanobeacons. *Angewandte Chemie International Edition* 48, 4170–4173, doi:10.1002/anie.200805947 (2009). [PubMed: 19418503]
48. Song KH, Kim C, Cogley CM, Xia Y & Wang LV Near-Infrared Gold Nanocages as a New Class of Tracers for Photoacoustic Sentinel Lymph Node Mapping on a Rat Model. *Nano Lett.* 9, 183–188, doi:10.1021/nl802746w (2009). [PubMed: 19072058]
49. Li P-C et al. In vivo photoacoustic molecular imaging with simultaneous multiple selective targeting using antibody-conjugated gold nanorods. *Optics Express* 16, 18605–18615 (2008). [PubMed: 19581946]
50. Ku G & Wang LV Deeply penetrating photoacoustic tomography in biological tissues enhanced with an optical contrast agent. *Optics letters* 30, 507–509 (2005). [PubMed: 15789718]
51. Kim C, Favazza C & Wang LV In Vivo Photoacoustic Tomography of Chemicals: High-Resolution Functional and Molecular Optical Imaging at New Depths. *Chemical reviews* 110, 2756–2782, doi:10.1021/cr900266s (2010). [PubMed: 20210338]
52. Filonov GS et al. Deep-Tissue Photoacoustic Tomography of a Genetically Encoded Near-Infrared Fluorescent Probe. *Angewandte Chemie International Edition* 51, 1448–1451, doi:10.1002/anie.201107026 (2012). [PubMed: 22213541]
53. Song KH, Stein E, Margenthaler J & Wang L Noninvasive photoacoustic identification of sentinel lymph nodes containing methylene blue in vivo in a rat model. *Journal of biomedical optics* 13, 054033 (2008). [PubMed: 19021413]
54. Kim C, Qin R, Xu J, Wang L & Xu R Multifunctional microbubbles and nanobubbles for photoacoustic and ultrasound imaging. *Journal of biomedical optics* 15, 010510 (2010). [PubMed: 20210423]
55. De La Zerda A et al. Carbon nanotubes as photoacoustic molecular imaging agents in living mice. *Nat. Nanotechnol* 3, 557–562, doi:10.1038/nnano.2008.231 (2008). [PubMed: 18772918]
56. Baker M The whole picture. *Nature* 463, 977–979 (2010). [PubMed: 20164931]
57. Zanzonico P, Kiessling F & Pichler J *Small Animal Imaging: Basics and Practical Guide*. Noninvasive imaging for supporting basic research. Heidelberg: Springer, 3–16 (2011).
58. Li L et al. Single-impulse panoramic photoacoustic computed tomography of small-animal whole-body dynamics at high spatiotemporal resolution. *Nature biomedical engineering* 1, 1–11 (2017).
59. Fehm TF, Deán-Ben XL, Ford SJ & Razansky D In vivo whole-body optoacoustic scanner with real-time volumetric imaging capacity. *Optica* 3, 1153–1159 (2016).

60. Li M-L et al. Simultaneous molecular and hypoxia imaging of brain tumors in vivo using spectroscopic photoacoustic tomography. *Proc. IEEE* 96, 481–489 (2008).
61. Zhang P, Li L, Lin L, Shi J & Wang LV In vivo superresolution photoacoustic computed tomography by localization of single dyed droplets. *Light: Science & Applications* 8, 1–9 (2019).
62. Zlitni A, Gowrishankar G, Steinberg I, Haywood T & Sam Gambhir S Maltotriose-based probes for fluorescence and photoacoustic imaging of bacterial infections. *Nature communications* 11, 1250, doi:10.1038/s41467-020-14985-8 (2020).
63. Yao J et al. Multiscale photoacoustic tomography using reversibly switchable bacterial phytochrome as a near-infrared photochromic probe. *Nat. Methods* 13, 67–73 (2016). [PubMed: 26550774]
64. Jathoul AP et al. Deep in vivo photoacoustic imaging of mammalian tissues using a tyrosinase-based genetic reporter. *Nature Photonics* 9, 239–246 (2015).
65. Mishra K et al. Multiplexed whole-animal imaging with reversibly switchable photoacoustic proteins. *Science Advances* 6, eaaz6293, doi:10.1126/sciadv.aaz6293 (2020). [PubMed: 32582850]
66. Märk J et al. Dual-wavelength 3D photoacoustic imaging of mammalian cells using a photoswitchable phytochrome reporter protein. *Communications Physics* 1, 3, doi:10.1038/s42005-017-0003-2 (2018).
67. Kim J-W, Galanzha EI, Shashkov EV, Moon H-M & Zharov VP Golden carbon nanotubes as multimodal photoacoustic and photothermal high-contrast molecular agents. *Nat. Nanotechnol* 4, 688–694 (2009). [PubMed: 19809462]
68. Chen Y-S, Zhao Y, Yoon SJ, Gambhir SS & Emelianov S Miniature gold nanorods for photoacoustic molecular imaging in the second near-infrared optical window. *Nat. Nanotechnol* 14, 465–472, doi:10.1038/s41565-019-0392-3 (2019). [PubMed: 30833692]
69. Pohling C et al. Smart-Dust-Nanorice for Enhancement of Endogenous Raman Signal, Contrast in Photoacoustic Imaging, and T2-Shortening in Magnetic Resonance Imaging. *Small* 14, 1703683, doi:10.1002/sml.201703683 (2018).
70. Lovell JF et al. Porphysome nanovesicles generated by porphyrin bilayers for use as multimodal biophotonic contrast agents. *Nat. Mater* 10, 324–332, doi:10.1038/nmat2986 (2011). [PubMed: 21423187]
71. Li L et al. Integration of Multitargeted Polymer-Based Contrast Agents with Photoacoustic Computed Tomography: An Imaging Technique to Visualize Breast Cancer Intratumor Heterogeneity. *ACS Nano* 15, 2413–2427 (2021). [PubMed: 33464827]
72. Jo J, Lee CH, Kopelman R & Wang X In vivo quantitative imaging of tumor pH by nanosonophore assisted multispectral photoacoustic imaging. *Nature communications* 8, 1–10 (2017).
73. Gujrati V et al. Bioengineered bacterial vesicles as biological nano-heaters for photoacoustic imaging. *Nature communications* 10, 1–10 (2019).
74. Li L et al. Small near-infrared photochromic protein for photoacoustic multi-contrast imaging and detection of protein interactions in vivo. *Nature communications* 9, 1–14 (2018).
75. Wang L, Verkhusha VV, Yao J, Kaberniuk AA & Li L (US Patent App. 15/261,116, 2017).
76. Wu Z et al. A microrobotic system guided by photoacoustic computed tomography for targeted navigation in intestines in vivo. *Science Robotics* 4, eaax0613 (2019). [PubMed: 32632399]
77. Li L et al. in *Photons Plus Ultrasound: Imaging and Sensing 2020*. 112402R (International Society for Optics and Photonics).
78. Pollock J Deep imaging technology needed for NIH BRAIN initiative. *Journal of biomedical optics* 19, 030601 (2014).
79. Yao J et al. High-speed label-free functional photoacoustic microscopy of mouse brain in action. *Nat. Methods* 12, 407 (2015). [PubMed: 25822799]
80. Li L et al. Label-free photoacoustic tomography of whole mouse brain structures ex vivo. *NEURO* 3, 035001 (2016).
81. Ovsepian SV, Olefir I, Westmeyer G, Razansky D & Ntziachristos V Pushing the boundaries of neuroimaging with photoacoustics. *Neuron* 96, 966–988 (2017). [PubMed: 29216459]




82. Gottschalk S et al. Rapid volumetric optoacoustic imaging of neural dynamics across the mouse brain. *Nature Biomedical Engineering* 3, 392–401, doi:10.1038/s41551-019-0372-9 (2019).
83. Zhang P et al. High-resolution deep functional imaging of the whole mouse brain by photoacoustic computed tomography in vivo. *Journal of biophotonics* 11, e201700024 (2018).
84. Zhang HF, Maslov K, Stoica G & Wang LV Functional photoacoustic microscopy for high-resolution and noninvasive in vivo imaging. *Nat. Biotechnol* 24, 848–851, doi:10.1038/nbt1220 (2006). [PubMed: 16823374]
85. Xu Z, Zhu Q & Wang LV In vivo photoacoustic tomography of mouse cerebral edema induced by cold injury. *Journal of biomedical optics* 16, doi:10.1117/1.3584847 (2011).
86. Shi J et al. High-resolution, high-contrast mid-infrared imaging of fresh biological samples with ultraviolet-localized photoacoustic microscopy. *Nature photonics* 13, 609–615 (2019). [PubMed: 31440304]
87. Galanzha EI et al. In vivo liquid biopsy using Cytophone platform for photoacoustic detection of circulating tumor cells in patients with melanoma. *Science translational medicine* 11 (2019).
88. He Y et al. In vivo label-free photoacoustic flow cytography and on-the-spot laser killing of single circulating melanoma cells. *Scientific reports* 6, 39616 (2016). [PubMed: 28000788]
89. Wong TT et al. Fast label-free multilayered histology-like imaging of human breast cancer by photoacoustic microscopy. *Science advances* 3, e1602168 (2017). [PubMed: 28560329]
90. Baker MJ et al. Using Fourier transform IR spectroscopy to analyze biological materials. *Nat. Protoc* 9, 1771 (2014). [PubMed: 24992094]
91. Cheng J-X & Xie XS Vibrational spectroscopic imaging of living systems: An emerging platform for biology and medicine. *Science* 350 (2015).
92. Pleitez MA et al. Label-free metabolic imaging by mid-infrared optoacoustic microscopy in living cells. *Nat. Biotechnol* 38, 293–296, doi:10.1038/s41587-019-0359-9 (2020). [PubMed: 31873214]
93. Bray F, F. J, Soerjomataram I, Siegel R, Torre L, Jemal A. Global cancer statistics 2018: GLOBOCAN estimates of incidence and mortality worldwide for 36 cancers in 185 countries. *C A Cancer J Clin.* 68, doi:10.3322/caac.21492 (2018).
94. Cronin K, L. A, Scott S, Sherman R, Noone A, Howlander N, et al. Annual report to the nation on the status of cancer, part I: National cancer statistics. *Cancer* 124, 2785–2800, doi:10.1002/cncr.31551 (2018). [PubMed: 29786848]
95. Lin L et al. Single-breath-hold photoacoustic computed tomography of the breast. *Nature communications* 9, 1–9 (2018).
96. Toi M et al. Visualization of tumor-related blood vessels in human breast by photoacoustic imaging system with a hemispherical detector array. *Scientific reports* 7, 41970 (2017). [PubMed: 28169313]
97. Matsumoto Y et al. Visualising peripheral arterioles and venules through high-resolution and large-area photoacoustic imaging. *Scientific reports* 8, 1–11 (2018). [PubMed: 29311619]
98. Matsumoto Y et al. Label-free photoacoustic imaging of human palmar vessels: a structural morphological analysis. *Scientific reports* 8, 1–8 (2018). [PubMed: 29311619]
99. Nagae K et al. Real-time 3D Photoacoustic Visualization System with a Wide Field of View for Imaging Human Limbs [version 2; peer review: 2 approved]. *F1000Research* 7, doi:10.12688/f1000research.16743.2 (2019).
100. Plumb AA, Huynh NT, Guggenheim J, Zhang E & Beard P Rapid volumetric photoacoustic tomographic imaging with a Fabry-Perot ultrasound sensor depicts peripheral arteries and microvascular vasomotor responses to thermal stimuli. *European radiology* 28, 1037–1045 (2018). [PubMed: 29018924]
101. Li Y et al. Snapshot photoacoustic topography through an ergodic relay for high-throughput imaging of optical absorption. *Nature Photonics* 14, 164–170 (2020). [PubMed: 34178097]
102. Draeger C & Fink M One-channel time reversal of elastic waves in a chaotic 2D-silicon cavity. *Physical Review Letters* 79, 407 (1997).
103. Zhu B, Liu JZ, Cauley SF, Rosen BR & Rosen MS Image reconstruction by domain-transform manifold learning. *Nature* 555, 487–492 (2018). [PubMed: 29565357]

104. Hyun CM, Kim HP, Lee SM, Lee S & Seo JK Deep learning for undersampled MRI reconstruction. *Physics in Medicine & Biology* 63, 135007 (2018). [PubMed: 29787383]
105. Poudel J, Matthews TP, Li L, Anastasio MA & Wang LV Mitigation of artifacts due to isolated acoustic heterogeneities in photoacoustic computed tomography using a variable data truncation-based reconstruction method. *Journal of biomedical optics* 22, 041018 (2017). [PubMed: 28267192]
106. Li L, Zhu L, Shen Y & Wang LV Multiview Hilbert transformation in full-ring transducer array-based photoacoustic computed tomography. *Journal of biomedical optics* 22, 076017 (2017). [PubMed: 28745385]
107. Matthews TP, Poudel J, Li L, Wang LV & Anastasio MA Parameterized joint reconstruction of the initial pressure and sound speed distributions for photoacoustic computed tomography. *SIAM journal on imaging sciences* 11, 1560–1588 (2018). [PubMed: 30956749]
108. Wang K, Su R, Oraevsky AA & Anastasio MA Investigation of iterative image reconstruction in three-dimensional photoacoustic tomography. *Physics in Medicine & Biology* 57, 5399 (2012). [PubMed: 22864062]
109. Treeby BE, Zhang EZ & Cox BT Photoacoustic tomography in absorbing acoustic media using time reversal. *Inverse Probl.* 26, 115003 (2010).
110. Bungart BL et al. Photoacoustic tomography of intact human prostates and vascular texture analysis identify prostate cancer biopsy targets. *Photoacoustics* 11, 46–55 (2018). [PubMed: 30109195]
111. Antholzer S, Haltmeier M & Schwab J Deep learning for photoacoustic tomography from sparse data. *Inverse problems in science and engineering* 27, 987–1005 (2019). [PubMed: 31057659]
112. Guan S, Khan AA, Sikdar S & Chitnis PV Fully Dense UNet for 2-D Sparse Photoacoustic Tomography Artifact Removal. *IEEE journal of biomedical and health informatics* 24, 568–576 (2019). [PubMed: 31021809]
113. Hauptmann A et al. Model-based learning for accelerated, limited-view 3-d photoacoustic tomography. *IEEE Trans. Med. Imaging* 37, 1382–1393 (2018). [PubMed: 29870367]
114. Cai C, Deng K, Ma C & Luo J End-to-end deep neural network for optical inversion in quantitative photoacoustic imaging. *Optics letters* 43, 2752–2755 (2018). [PubMed: 29905680]
115. Davoudi N, Deán-Ben XL & Razansky D Deep learning photoacoustic tomography with sparse data. *Nature Machine Intelligence* 1, 453–460 (2019).
116. Li G, Li L, Zhu L, Xia J & Wang LV Multiview Hilbert transformation for full-view photoacoustic computed tomography using a linear array. *Journal of biomedical optics* 20, 066010 (2015). [PubMed: 26112369]
117. Liu W et al. Correcting the limited view in optical-resolution photoacoustic microscopy. *Journal of biophotonics* 11, e201700196 (2018).
118. Lan H et al. in 2019 41st Annual International Conference of the IEEE Engineering in Medicine and Biology Society (EMBC). 6367–6370 (IEEE).




**Figure 1.**  
The principle of PAT.



**Figure 2.**

Multicontrast and multiscale PAT. (a) Absorption spectra of endogenous molecules at normal concentrations *in vivo*.<sup>27</sup> Bilirubin, 12 mg L<sup>-1</sup> in blood; DNA/RNA, 1 g L<sup>-1</sup> in cell nuclei; HbO<sub>2</sub>, Oxy-hemoglobin, HbR, deoxy-hemoglobin, 2.3 mM in blood; MbO<sub>2</sub>, oxy-myoglobin, MbR, reduced myoglobin, 0.5% mass concentration in skeletal muscle; melanin, 14.3 g L<sup>-1</sup> in skin; lipid, 20% volume concentration in tissue; water, 80% volume concentration in tissue. (b) Noise equivalent molar concentrations of some widely used exogenous contrast agents, based on reported values from literature.<sup>27</sup> Illumination fluence is not compensated. EB, evens blue<sup>45</sup>; EGFP, enhanced green fluorescent protein<sup>46</sup>; GNB, gold nanobeacon<sup>47</sup>; GNC, gold nanocage<sup>48</sup>; GNR, gold nanorod<sup>49</sup>; ICG, indocyanine green<sup>50</sup>; IRDye800, near-infrared Dye800<sup>51</sup>; iRFP, near-infrared red fluorescent protein<sup>52</sup>; MB, methylene blue<sup>53</sup>; mCherry, monomeric cherry protein<sup>46</sup>; Microbubble<sup>54</sup>; RFP, red fluorescent protein<sup>52</sup>; SWNT, single-walled nanotube<sup>55</sup>. The dashed curve is power function


fitting  $y = 0.1x^{-1}$ , where  $y$  is the noise equivalent concentration in molar,  $x$  the molar extinction coefficient in  $\text{cm}^{-1} \text{M}^{-1}$ . (c) Multiscale PAT and representative images. Organelles, PA nanoscopy of a single mitochondrion, scale bar, 500 nm.<sup>37</sup> Single cells, optical-resolution PAM of red blood cells, scale bar, 20  $\mu\text{m}$ .<sup>38</sup> Tissues, acoustic-resolution PAM of human skin, scale bar, 500  $\mu\text{m}$ .<sup>25</sup> Wholebody small animals, wholebody PACT of a nude mouse *in vivo*, scale bar, 4 mm.<sup>39</sup>



**Figure 3.**


Wholebody PACT of small animals.<sup>58,59</sup> (a) Schematic of the SIP-PACT system for trunk and brain (blue dashed boxed inset) imaging.<sup>58</sup> Dual-wavelength illumination is used. BC, beam combiner; CL, conical lens; DAQ, data acquisition system; MBS, magnetic base scanner; OC, optical condenser; USTA, (full-ring) ultrasonic transducer array; WT, water tank. (b) Close up of the green dashed in (a), showing the confocal design of light illumination and acoustic detection. (c–f) Representative cross-sectional images of the brain (c), the liver (d), the upper abdominal cavity (e), and the lower abdominal cavity (f) in a live mouse, acquired by SIP-PACT.<sup>58</sup> Scale bar, 5 mm. (g) Layout of the spiral scanning PACT system for small-animal wholebody imaging.<sup>59</sup> DAQ, data acquisition unit. (h) Representative 3D wholebody images of a live mouse. Each image overlays the beating mouse heart (color) onto a wholebody anatomical image (gray) of the same mouse.<sup>59</sup> Scale bar, 5 mm.






**Figure 4.**

Representative images of molecular PACT. (a) *In vivo* PA images of Tyr-expressing K562 cells after subcutaneous injection into the flank of a nude mouse (vasculature is color-coded for depth; K562 cells are false-colored yellow). Scale bar, 1 mm.<sup>64</sup> (b) Quantitative PACT of pH *in vivo*. Functional PA image in pseudo-color is superimposed on the gray-scale ultrasound image. Scale bar, 2 mm.<sup>72</sup> (c) *In vivo* PA image of a 4T1 tumor-bearing mouse, given a single injection of 150  $\mu\text{g}$  of OMV<sup>Me1</sup> via the tail vein. The image was acquired at 3 h post-injection, showing the accumulation of OMV<sup>Me1</sup> in tumor tissue, where the OMV<sup>Me1</sup> is in color, and the background tissue is in gray.<sup>73</sup> (d) *In vivo* multicontrast PACT of two types of tumor cells in the liver. Two types of tumors expressing different photoswitchable proteins are separated by their decay characteristics. The tumors are shown in color, and the background tissues are shown in gray. Norm., normalized.<sup>74</sup> (e) PA image of a hydrodynamic-transfected liver. The photoswitching signals are shown in color, confirming the existence of reconstituted DrSplit induced by protein-protein interactions. The background tissues are shown in gray.<sup>74</sup> (f) PA image of the microrobots in the intestines *in vivo*. The migrating microrobots are shown in color, and the mouse tissues are shown in gray. The yellow arrows indicate the direction of migration.<sup>76</sup>




**Figure 5.** Multiscale PAT of the brain. (a) PAM of oxygen saturation of hemoglobin in a mouse brain.<sup>79</sup> (b) A cross-sectional PACT image of a saline-perfused mouse brain (horizontal plane) at 2.8-mm depth, showing internal structures of the brain clearly. Nc, neocortex; CC, corpus callosum; Hp, hippocampus; Cb, cerebellum; IC, inferior colliculus.<sup>80</sup> (c) 3D PACT image of a mouse brain *ex vivo*. Illumination wavelength, 740 nm; V2MM, secondary visual cortex, medio-medial; CA1, hippocampal CA1 area; DG, dentate gyrus; D3V, dorsal third ventricle; ZID, zona incerta dorsal; SNr, substantia nigra reticulata; VTA, ventral tegmental area; IFN, inter-fascicular nucleus.<sup>81</sup> (d) Functional mapping of the resting-state connectivity in a rat whole brain (coronal plane), showing a clear correlation between corresponding regions across the left and right hemispheres.<sup>58</sup> (e) PACT of GCaMP6s responses to electrical stimulation of the right or left hind paw. First from the left, maximum amplitude projection along the depth direction of the 3D images of a GCaMP6s-expressing mouse; second to last, relative increases in PA signal with respect to the baseline for a slice at ~1 mm depth at different time points following the stimulation pulse for the GCaMP6s-expressing mouse.<sup>82</sup> (f) PACT images of epileptic activities during a seizure at different times. The fractional changes (color) in the PA amplitude are overlaid on the anatomical image (gray, Bregma -1.0 mm). The arrow indicates the injection site, and the dashed green arrows indicate the epileptic wave propagation direction.<sup>83</sup>



**Figure 6.**


Broad-spectrum PAM of tissues and cells. (a) PAM image of a fixed, unprocessed breast tumor. Illumination wavelength, 266 nm.<sup>89</sup> (b–d) Zoomed-in PAM images of the red, yellow, and magenta dashed regions in (a), respectively. IDC, invasive ductal carcinoma; DCIS, ductal carcinoma in situ; CN, cell nuclei.<sup>89</sup> (e) Monitoring intrinsic lipid contrast during lipolysis in differentiated 3T3-L1 adipocytes at 2,857 cm<sup>-1</sup>. Two regions of interest (ROIs) enclosing individual adipocytes are marked; green dashed circle for ROI 1 and red dashed circle for ROI 2. The white arrow follows the process of lipid droplet remodeling in a single adipocyte enclosed in ROI 1.<sup>92</sup> (f, g) ULM-PAM images of lipids (f) and proteins (g).<sup>86</sup> (h, i) MIR-PAM images of lipids (h) and proteins (i), imaged at 3,420 nm and 6,050 nm, respectively.<sup>86</sup> (j, k) Composite images of cells formed by overlaying the images of lipids (blue), proteins (green), and nucleic acids (red) in different color channels at neonatal (j) and mature (k) stages<sup>86</sup>.






**Figure 7.**

PACT of human breasts. (a) X-ray mammograms of an affected breast.<sup>95</sup> (b) Depth-encoded angiogram of the affected breast acquired by SBH-PACT. The breast tumor is identified by a white circle, the nipple is marked by a magenta circle.<sup>95</sup> (c) Maximum amplitude projection (MAP) images of the thick slice in sagittal planes marked by white dashed lines in (b).<sup>95</sup> (d) Automatic tumor detection on vessel density maps. Tumors are identified by green circles. Background images in gray scale are the MAP of vessels deeper than the nipple.<sup>95</sup> (e) Depth-encoded 3D PACT image of a healthy breast.<sup>96</sup> (f) Evaluation of the S-factor in a healthy breast, where the color represents the measured S-factor.<sup>97</sup> (g) A fusion image of the S-factor and 3D-US images (red color).<sup>97</sup>




**Figure 8.** PACT images of human extremities. (a–d) PACT images of various extremities, including palm (a), back of the hand (b), forearm (c), and lower thigh.<sup>99</sup> (e) MAP images of human fingertips after cold (left-hand panels) and warm (right-hand panels) water immersion, color-coded for depth. While arrows show the same vessels in each imaging condition.<sup>100</sup>



**Figure 9.**

Principles of PATER and representative images.<sup>101</sup> (a–f) Principles of PATER. (a) Simulation of acoustic propagation in the ER in calibration mode. Norm., normalized. (b) Simulation of acoustic propagation in the ER in wide-field mode. (c) In calibration mode, light is focused on each pixel to acquire the impulse response encoded by the ER,  $k_i$ , and then raster scanned over the FOV. (d) Calibration image formed by computing the root-mean-squared amplitude of each received PA signal at every calibration position. (e) Snapshot wide-field imaging. A broad laser beam illuminates the entire FOV to acquire encoded signals,  $s_i$ , which can be repeated for high-speed imaging. (f) Reconstructed wide-field images. The reconstruction algorithm uses calibrated impulse responses to decode the wide-field signals and then display wide-field images. (g) Monitoring of blood pulse wave propagation. Wide-field images at different time points illustrate the thermal wave propagation in the middle cerebral arteries. The yellow circles, labeled  $L_1$ ,  $L_2$ , indicate the locations of the laser heating spots during wide-field measurement. The thermal wave signals are shown in color, and the background vessels are shown in gray.





**Figure 10.**

Image reconstructions assisted by machine learning. (a–e) Deep learning PAT with sparse data.<sup>115</sup> (a) The U-Net network architecture, consisting of contracting (downsampling) and expansive (upsampling) paths, which is used for the image reconstruction with sparse data. (b) Artefactual reconstructed image with undersampled (128 projections) data, showing the reconstruction artifacts due to the sparse data. (c) Zoom-in images of the yellow and green boxed regions in (b). (d) Artifact-free counterpart of (b), obtained with the trained network. (e) Zoom-in images of the yellow and green boxed regions in (d). (f, g) Hybrid neural network for limited-view PACT.<sup>118</sup> (f) The global architecture of Y-Net. Two encoders extract different input features, which concatenate into the decoder. Both encoders have skip connections with the decoder. (g) Comparison of reconstructed images. Top left, ground truth; top right, image reconstructed using the universal back-projection method; bottom left, image reconstructed using the time-reversal method; bottom right, image reconstructed using the trained Y-net.